

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: **8081B** Datafile : PL094834.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167272BS	alpha-BHC	16.65	17.4	ug/kg	105				84	123	
	beta-BHC	16.65	17.1	ug/kg	103				82	123	
	delta-BHC	16.65	17.1	ug/kg	103				83	126	
	gamma-BHC (Lindane)	16.65	16.9	ug/kg	102				83	125	
	Heptachlor	16.65	16.5	ug/kg	99				83	122	
	Aldrin	16.65	16.7	ug/kg	100				82	124	
	Heptachlor epoxide	16.65	16.9	ug/kg	102				83	120	
	Endosulfan I	16.65	16.8	ug/kg	101				81	124	
	Dieldrin	16.65	17.1	ug/kg	103				85	121	
	4,4'-DDE	16.65	17.8	ug/kg	107				81	123	
	Endrin	16.65	17.4	ug/kg	105				76	130	
	Endosulfan II	16.65	17.3	ug/kg	104				80	125	
	4,4'-DDD	16.65	18.7	ug/kg	112				80	131	
	Endosulfan sulfate	16.65	17.9	ug/kg	108				81	122	
	4,4'-DDT	16.65	18.7	ug/kg	112				70	129	
	Methoxychlor	16.65	19.3	ug/kg	116				60	119	
	Endrin ketone	16.65	18.0	ug/kg	108				77	132	
	Endrin aldehyde	16.65	17.4	ug/kg	105				79	124	
	alpha-Chlordane	16.65	16.8	ug/kg	101				84	120	
	gamma-Chlordane	16.65	16.9	ug/kg	102				83	122	