

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 03/27/2025

Lab Sample No.: HSTDCCC750 Data File : PS029588.D Time Analyzed: 22:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.080	8.982	9.182	718.400	712.500	0.8
2,4-D	8.220	8.122	8.322	741.930	705.000	5.2
2,4-DCAA	7.114	7.015	7.215	743.510	750.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: WALS01

Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL02 Date Analyzed: 03/27/2025

Lab Sample No.: HSTDCCC750 Data File : PS029588.D Time Analyzed: 22:51

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.630	9.534	9.734	728.840	712.500	2.3
2,4-D	8.745	8.648	8.848	714.600	705.000	1.4
2,4-DCAA	7.534	7.437	7.637	753.070	750.000	0.4