

Surrogate Summary

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094566.D	PIBLK-PL094566.D	Decachlorobiphenyl	1	20	22.7	114		43	140
		Tetrachloro-m-xylene	1	20	20.6	103		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	20.2	101		77	126
I.BLK-PL094849.D	PIBLK-PL094849.D	Decachlorobiphenyl	1	20	19.9	99		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	14.2	71		43	140
		Tetrachloro-m-xylene	2	20	19.1	95		77	126
PB167311BL	PB167311BL	Decachlorobiphenyl	1	20	20.4	102		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	16.6	83		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
PB167275TB	PB167275TB	Decachlorobiphenyl	1	20	21.2	106		43	140
		Tetrachloro-m-xylene	1	20	18.2	91		77	126
		Decachlorobiphenyl	2	20	18.0	90		43	140
		Tetrachloro-m-xylene	2	20	19.7	99		77	126
Q1609-03MS	WC-SCRN-01-CMS	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	18.9	94		77	126
		Decachlorobiphenyl	2	20	18.2	91		43	140
		Tetrachloro-m-xylene	2	20	19.8	99		77	126
Q1609-03MSD	WC-SCRN-01-CMSD	Decachlorobiphenyl	1	20	21.6	108		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	18.3	91		43	140
		Tetrachloro-m-xylene	2	20	19.9	100		77	126
Q1626-03	CO-32-1	Decachlorobiphenyl	1	20	27.8	139		43	140
		Tetrachloro-m-xylene	1	20	22.2	111		77	126
		Decachlorobiphenyl	2	20	24.3	121		43	140
		Tetrachloro-m-xylene	2	20	23.1	115		77	126
I.BLK-PL094862.D	PIBLK-PL094862.D	Decachlorobiphenyl	1	20	19.9	99		43	140
		Tetrachloro-m-xylene	1	20	18.1	91		77	126
		Decachlorobiphenyl	2	20	18.4	92		43	140
		Tetrachloro-m-xylene	2	20	18.6	93		77	126
I.BLK-PL094893.D	PIBLK-PL094893.D	Decachlorobiphenyl	1	20	21.5	107		43	140
		Tetrachloro-m-xylene	1	20	19.2	96		77	126
		Decachlorobiphenyl	2	20	19.9	100		43	140
		Tetrachloro-m-xylene	2	20	19.7	99		77	126
PB167311BS	PB167311BS	Decachlorobiphenyl	1	20	24.6	123		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	21.0	105		43	140
		Tetrachloro-m-xylene	2	20	21.4	107		77	126
I.BLK-PL094901.D	PIBLK-PL094901.D	Decachlorobiphenyl	1	20	21.2	106		43	140

Surrogate Summary

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL094901.D	PIBLK-PL094901.D	Tetrachloro-m-xylene	1	20	17.9	90		77	126
		Decachlorobiphenyl	2	20	20.2	101		43	140
		Tetrachloro-m-xylene	2	20	18.9	95		77	126