

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: 8081B

DataFile : PL094857.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1609-03MS	WC-SCRN-01-CMS											
	gamma-BHC (Lindane)	5	0	5.00	ug/L	100				60	152	
	Heptachlor	5	0	4.70	ug/L	94				56	147	
	Heptachlor epoxide	5	0	4.80	ug/L	96				77	143	
	Endrin	5	0	4.90	ug/L	98				76	144	
	Methoxychlor	5	0	5.10	ug/L	102				70	142	