

TPH GC INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: WALS01
 ProjectID: Walsh CO-032 Sampling
 Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG No.: Q1626

Calibration Sequence : FG030325		Test : TPH GC	
Concentration (PPM)	Area Count	Reference Factor	File ID
1700	180093239	105937	FG015426.D
850	88070199	103612	FG015427.D
340	36924679	108602	FG015428.D
170	19919720	117175	FG015429.D
85	10372340	122028	FG015430.D
AVG RF : 111471		% RSD : 7.015	AVG RT : 15.0338