

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF032425

Review By	anahy	Review On	3/25/2025 9:38:18 AM		
Supervise By	Jagrut	Supervise On	3/25/2025 9:39:38 AM		
SubDirectory	BF032425	HP Acquire Method	BNA_F	HP Processing Method	bf031025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729
CCC	SP6725
Internal Standard/PEM	S12655,10ul/1000ul sample
ICV/I.BLK	SP6686
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142044.D	24 Mar 2025 09:43		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF142045.D	24 Mar 2025 10:17		RC/JU	Ok
3	PB167228BL	PB167228BL	BF142046.D	24 Mar 2025 10:47		RC/JU	Ok
4	Q1597-02	1-CONCRETE-SLAB	BF142047.D	24 Mar 2025 11:27		RC/JU	Ok
5	Q1597-04	2-CONCRETE-SLAB	BF142048.D	24 Mar 2025 11:56		RC/JU	Ok
6	Q1592-02MS	OILY-DEBRIS-COMP	BF142049.D	24 Mar 2025 12:27		RC/JU	Ok,M
7	Q1592-02MSD	OILY-DEBRIS-COMP	BF142050.D	24 Mar 2025 12:56		RC/JU	Ok,M
8	Q1619-13DL	TP-5DL	BF142051.D	24 Mar 2025 13:26	Need 2X Further Dilution	RC/JU	Dilution
9	Q1618-01	TP20250320-01	BF142052.D	24 Mar 2025 13:55		RC/JU	Ok
10	Q1618-02	TP20250320-02	BF142053.D	24 Mar 2025 14:25		RC/JU	Ok
11	Q1619-13DL2	TP-5DL2	BF142054.D	24 Mar 2025 14:55		RC/JU	Ok,M
12	Q1606-11	N48965	BF142055.D	24 Mar 2025 15:25	Internal Standard Fail	RC/JU	ReRun
13	Q1606-02	CHRT24743	BF142056.D	24 Mar 2025 15:55		RC/JU	Ok
14	Q1606-04	RBR251346	BF142057.D	24 Mar 2025 16:24		RC/JU	Ok
15	Q1606-06	RT4534	BF142058.D	24 Mar 2025 16:54		RC/JU	Ok
16	Q1606-08	RT3025	BF142059.D	24 Mar 2025 17:24		RC/JU	Ok
17	Q1606-10	CHRT28607	BF142060.D	24 Mar 2025 17:54		RC/JU	Ok
18	Q1610-03	SOIL-PILE	BF142061.D	24 Mar 2025 18:23		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF032425

Review By	anahy	Review On	3/25/2025 9:38:18 AM		
Supervise By	Jagrut	Supervise On	3/25/2025 9:39:38 AM		
SubDirectory	BF032425	HP Acquire Method	BNA_F	HP Processing Method	bf031025
STD. NAME	STD REF.#				
Tune/Reschk	SP6717				
Initial Calibration Stds	SP6722,SP6723,SP6724,SP6725,SP6726,SP6727,SP6728,SP6729				
CCC	SP6725				
Internal Standard/PEM	S12655,10ul/1000ul sample				
ICV/I.BLK	SP6686				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q1619-12	TP-4	BF142062.D	24 Mar 2025 18:53		RC/JU	Ok
20	Q1624-01	OK-01-03212025	BF142063.D	24 Mar 2025 19:23		RC/JU	Ok,M
21	Q1626-01	CO-32-1	BF142064.D	24 Mar 2025 19:53		RC/JU	Ok,M
22	Q1626-01MS	CO-32-1MS	BF142065.D	24 Mar 2025 20:22		RC/JU	Ok
23	Q1626-01MSD	CO-32-1MSD	BF142066.D	24 Mar 2025 20:51	Internal standard fail.	RC/JU	Ok

M : Manual Integration