

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1626-01MSD	Client Sample ID:	CO-32-1MSD					DataFile:	BF142066.D		
Benzaldehyde	1800	0	1900	ug/Kg	106	*	0		10	86	20
Phenol	1800	0	1700	ug/Kg	94		0		67	126	20
bis(2-Chloroethyl)ether	1800	0	1800	ug/Kg	100		0		54	125	20
2-Chlorophenol	1800	0	1800	ug/Kg	100		0		79	107	20
2-Methylphenol	1800	0	1700	ug/Kg	94		0		66	122	20
2,2-oxybis(1-Chloropropane)	1800	0	1600	ug/Kg	89		0		65	110	20
Acetophenone	1800	0	1800	ug/Kg	100		10		75	111	20
3+4-Methylphenols	1800	0	1600	ug/Kg	89		0		66	104	20
N-Nitroso-di-n-propylamine	1800	0	1500	ug/Kg	83		7		59	119	20
Hexachloroethane	1800	0	1600	ug/Kg	89		5		65	117	20
Nitrobenzene	1800	0	1800	ug/Kg	100		6		70	119	20
Isophorone	1800	0	1800	ug/Kg	100		6		76	122	20
2-Nitrophenol	1800	0	1600	ug/Kg	89		17		54	145	20
2,4-Dimethylphenol	1800	0	2400	ug/Kg	133		4		44	135	20
bis(2-Chloroethoxy)methane	1800	0	1800	ug/Kg	100		6		68	112	20
2,4-Dichlorophenol	1800	0	1700	ug/Kg	94		6		72	118	20
Naphthalene	1800	0	1800	ug/Kg	100		6		72	110	20
4-Chloroaniline	1800	0	700	ug/Kg	39		23	*	10	91	20
Hexachlorobutadiene	1800	0	2000	ug/Kg	111		0		66	114	20
Caprolactam	1800	0	1600	ug/Kg	89		0		51	134	20
4-Chloro-3-methylphenol	1800	0	1600	ug/Kg	89		0		57	132	20
2-Methylnaphthalene	1800	0	1500	ug/Kg	83		7		59	123	20
Hexachlorocyclopentadiene	3500	0	980	ug/Kg	28		35	*	10	175	20
2,4,6-Trichlorophenol	1800	0	1900	ug/Kg	106		5		72	117	20
2,4,5-Trichlorophenol	1800	0	1900	ug/Kg	106		0		72	117	20
1,1-Biphenyl	1800	0	1900	ug/Kg	106		5		75	113	20
2-Chloronaphthalene	1800	0	1900	ug/Kg	106		5		67	118	20
2-Nitroaniline	1800	0	2000	ug/Kg	111		0		69	127	20
Dimethylphthalate	1800	0	1800	ug/Kg	100		10		70	113	20
Acenaphthylene	1800	0	2100	ug/Kg	117		0		79	118	20
2,6-Dinitrotoluene	1800	0	1800	ug/Kg	100		6		70	125	20
3-Nitroaniline	1800	0	1800	ug/Kg	100	*	19		30	99	20
Acenaphthene	1800	0	1700	ug/Kg	94		6		70	121	20
2,4-Dinitrophenol	3500	0	440	ug/Kg	13		47	*	10	155	20
4-Nitrophenol	3500	0	4100	ug/Kg	117		3		45	133	20
Dibenzofuran	1800	0	1800	ug/Kg	100		0		72	110	20
2,4-Dinitrotoluene	1800	0	1700	ug/Kg	94		12		55	128	20
Diethylphthalate	1800	0	1800	ug/Kg	100		6		70	112	20
4-Chlorophenyl-phenylether	1800	0	1900	ug/Kg	106		0		71	108	20
Fluorene	1800	0	2000	ug/Kg	111		5		68	116	20
4-Nitroaniline	1800	0	1900	ug/Kg	106		12		55	120	20
4,6-Dinitro-2-methylphenol	1800	0	150	ug/Kg	8	*	93	*	10	160	20
N-Nitrosodiphenylamine	1800	0	1700	ug/Kg	94		12		73	118	20
4-Bromophenyl-phenylether	1800	0	1700	ug/Kg	94		12		65	121	20
Hexachlorobenzene	1800	0	1800	ug/Kg	100		10		67	118	20
Atrazine	1800	0	2100	ug/Kg	117		9		79	127	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Pentachlorophenol	3500	0	4000	ug/Kg	114		8		47	128	20
Phenanthrene	1800	730	2400	ug/Kg	93		11		52	128	20
Anthracene	1800	200	2000	ug/Kg	100		10		62	124	20
Carbazole	1800	70.7	2000	ug/Kg	107		5		59	119	20
Di-n-butylphthalate	1800	0	1900	ug/Kg	106		5		69	118	20
Fluoranthene	1800	1400	3100	ug/Kg	94		12		44	125	20
Pyrene	1800	850	2600	ug/Kg	97		5		26	142	20
Butylbenzylphthalate	1800	0	2000	ug/Kg	111		10		64	126	20
3,3-Dichlorobenzidine	1800	0	1100	ug/Kg	61		20		33	116	20
Benzo(a)anthracene	1800	660	2500	ug/Kg	102		6		71	114	20
Chrysene	1800	580	2400	ug/Kg	101		6		57	121	20
bis(2-Ethylhexyl)phthalate	1800	240	2100	ug/Kg	103		0		42	169	20
Di-n-octyl phthalate	1800	0	2100	ug/Kg	117		0		23	175	20
Benzo(b)fluoranthene	1800	790	2700	ug/Kg	106		6		67	121	20
Benzo(k)fluoranthene	1800	340	2200	ug/Kg	103		0		57	134	20
Benzo(a)pyrene	1800	660	2600	ug/Kg	108		0		70	142	20
Indeno(1,2,3-cd)pyrene	1800	350	2100	ug/Kg	97		5		40	129	20
Dibenz(a,h)anthracene	1800	100	1800	ug/Kg	94		5		43	123	20
Benzo(g,h,i)perylene	1800	460	2100	ug/Kg	91		0		24	125	20
1,2,4,5-Tetrachlorobenzene	1800	0	2000	ug/Kg	111		5		69	124	20
1,4-Dioxane	1800	0	1700	ug/Kg	94		0		46	112	20
2,3,4,6-Tetrachlorophenol	1800	0	1900	ug/Kg	106		0		69	112	20