

## Report of Analysis

|                    |                                    |                 |          |
|--------------------|------------------------------------|-----------------|----------|
| Client:            | Walsh Construction Company II, LLC | Date Collected: | 03/20/25 |
| Project:           | Walsh CO-032 Sampling              | Date Received:  | 03/21/25 |
| Client Sample ID:  | GRID-LINE-1.2-NORTHMS              | SDG No.:        | Q1626    |
| Lab Sample ID:     | Q1627-01MS                         | Matrix:         | TCLP     |
| Analytical Method: | SW8270                             | % Solid:        | 0        |
| Sample Wt/Vol:     | 100 Units: mL                      | Final Vol:      | 1000 uL  |
| Soil Aliquot Vol:  | uL                                 | Test:           | TCLP BNA |
| Extraction Type :  | Decanted : N                       | Level :         | LOW      |
| Injection Volume : | GPC Factor : 1.0                   | GPC Cleanup :   | N PH :   |
| Prep Method :      | SW3510C                            |                 |          |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| BF142086.D        | 1         | 03/25/25 12:25 | 03/25/25 19:08 | PB167310      |

| CAS Number                | Parameter              | Conc.  | Qualifier | MDL      | LOQ / CRQL | Units    |
|---------------------------|------------------------|--------|-----------|----------|------------|----------|
| <b>TARGETS</b>            |                        |        |           |          |            |          |
| 110-86-1                  | Pyridine               | 400    |           | 12.8     | 50.0       | ug/L     |
| 106-46-7                  | 1,4-Dichlorobenzene    | 460    |           | 5.30     | 50.0       | ug/L     |
| 95-48-7                   | 2-Methylphenol         | 520    |           | 11.2     | 50.0       | ug/L     |
| 65794-96-9                | 3+4-Methylphenols      | 510    |           | 11.0     | 100        | ug/L     |
| 67-72-1                   | Hexachloroethane       | 450    |           | 6.50     | 50.0       | ug/L     |
| 98-95-3                   | Nitrobenzene           | 510    |           | 7.60     | 50.0       | ug/L     |
| 87-68-3                   | Hexachlorobutadiene    | 500    |           | 5.40     | 50.0       | ug/L     |
| 88-06-2                   | 2,4,6-Trichlorophenol  | 570    |           | 5.10     | 50.0       | ug/L     |
| 95-95-4                   | 2,4,5-Trichlorophenol  | 590    |           | 6.20     | 50.0       | ug/L     |
| 121-14-2                  | 2,4-Dinitrotoluene     | 600    |           | 12.2     | 50.0       | ug/L     |
| 118-74-1                  | Hexachlorobenzene      | 620    |           | 5.20     | 50.0       | ug/L     |
| 87-86-5                   | Pentachlorophenol      | 790    |           | 15.8     | 100        | ug/L     |
| <b>SURROGATES</b>         |                        |        |           |          |            |          |
| 367-12-4                  | 2-Fluorophenol         | 144    |           | 10 - 139 | 96%        | SPK: 150 |
| 13127-88-3                | Phenol-d6              | 135    |           | 10 - 134 | 90%        | SPK: 150 |
| 4165-60-0                 | Nitrobenzene-d5        | 108    |           | 49 - 133 | 108%       | SPK: 100 |
| 321-60-8                  | 2-Fluorobiphenyl       | 108    |           | 52 - 132 | 108%       | SPK: 100 |
| 118-79-6                  | 2,4,6-Tribromophenol   | 179    |           | 44 - 137 | 119%       | SPK: 150 |
| 1718-51-0                 | Terphenyl-d14          | 114    |           | 48 - 125 | 114%       | SPK: 100 |
| <b>INTERNAL STANDARDS</b> |                        |        |           |          |            |          |
| 3855-82-1                 | 1,4-Dichlorobenzene-d4 | 147000 | 6.875     |          |            |          |
| 1146-65-2                 | Naphthalene-d8         | 569000 | 8.157     |          |            |          |
| 15067-26-2                | Acenaphthene-d10       | 317000 | 9.91      |          |            |          |
| 1517-22-2                 | Phenanthrene-d10       | 518000 | 11.398    |          |            |          |
| 1719-03-5                 | Chrysene-d12           | 312000 | 14.033    |          |            |          |
| 1520-96-3                 | Perylene-d12           | 310000 | 15.504    |          |            |          |

## Report of Analysis

|                    |                                    |                 |              |
|--------------------|------------------------------------|-----------------|--------------|
| Client:            | Walsh Construction Company II, LLC | Date Collected: | 03/20/25     |
| Project:           | Walsh CO-032 Sampling              | Date Received:  | 03/21/25     |
| Client Sample ID:  | GRID-LINE-1.2-NORTHMS              | SDG No.:        | Q1626        |
| Lab Sample ID:     | Q1627-01MS                         | Matrix:         | TCLP         |
| Analytical Method: | SW8270                             | % Solid:        | 0            |
| Sample Wt/Vol:     | 100      Units:    mL              | Final Vol:      | 1000      uL |
| Soil Aliquot Vol:  | uL                                 | Test:           | TCLP BNA     |
| Extraction Type :  | Decanted :    N                    | Level :         | LOW          |
| Injection Volume : | GPC Factor :    1.0                | GPC Cleanup :   | N      PH :  |
| Prep Method :      | SW3510C                            |                 |              |

|                   |           |                |                |               |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date      | Date Analyzed  | Prep Batch ID |
| BF142086.D        | 1         | 03/25/25 12:25 | 03/25/25 19:08 | PB167310      |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units |
|------------|-----------|-------|-----------|-----|------------|-------|
|------------|-----------|-------|-----------|-----|------------|-------|

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products