

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN086094.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0325WBS01	Vinyl chloride	20	20.7	ug/L	104			65	117	
	1,1-Dichloroethene	20	20.6	ug/L	103			74	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	18.6	ug/L	93			77	113	
	Chloroform	20	20.4	ug/L	102			79	113	
	Benzene	20	20.8	ug/L	104			82	109	
	1,2-Dichloroethane	20	20.1	ug/L	101			80	115	
	Trichloroethene	20	18.7	ug/L	94			77	113	
	Tetrachloroethene	20	20.1	ug/L	101			67	123	
	Chlorobenzene	20	19.9	ug/L	100			82	109	