

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1626

Client: Walsh Construction Company II, LLC

Analytical Method: SW8260-Low

Datafile : VN086106.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0325WBSD01	Vinyl chloride	20	20.3	ug/L	102	2		65	117	20
	1,1-Dichloroethene	20	19.2	ug/L	96	7		74	110	20
	2-Butanone	100	87.6	ug/L	88	22	*	65	122	20
	Carbon Tetrachloride	20	18.1	ug/L	91	2		77	113	20
	Chloroform	20	19.7	ug/L	99	3		79	113	20
	Benzene	20	19.5	ug/L	98	6		82	109	20
	1,2-Dichloroethane	20	18.5	ug/L	93	8		80	115	20
	Trichloroethene	20	17.5	ug/L	88	7		77	113	20
	Tetrachloroethene	20	16.9	ug/L	85	17		67	123	20
	Chlorobenzene	20	18.7	ug/L	94	6		82	109	20