

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY030425

Review By	Maresh Dadoda	Review On	3/5/2025 12:14:39 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/5/2025 12:18:08 PM
SubDirectory	VY030425	HP Acquire Method	MSVOA_Y HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk	VP133216
Initial Calibration Stds	VP133207,VP133208,VP133209,VP133210,VP133211,VP133212
CCC	VP133214,VP133215
Internal Standard/PEM	VP131783
ICV/I.BLK	VP133213
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021395.D	04 Mar 2025 08:52		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY021396.D	04 Mar 2025 09:23	not used	SY/MD	Not Ok
3	VSTDICC010	VSTDICC010	VY021397.D	04 Mar 2025 09:46		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY021398.D	04 Mar 2025 10:09	Comp.#11 is on Linear Regression	SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY021399.D	04 Mar 2025 10:30	Comp.#95 is on Quadratic Regression	SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY021400.D	04 Mar 2025 11:06		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY021401.D	04 Mar 2025 11:29	Method fail for comp.#13	SY/MD	Ok,M
8	VIBLK	VIBLK	VY021402.D	04 Mar 2025 11:52		SY/MD	Ok
9	VSTDICC005	VSTDICC005	VY021403.D	04 Mar 2025 12:15		SY/MD	Ok,M
10	VSTDICV050	ICVVY030425	VY021404.D	04 Mar 2025 13:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY032525

Review By	Maresh Dadoda	Review On	3/26/2025 1:07:14 PM
Supervise By	Semsettin Yesilyurt	Supervise On	3/26/2025 1:11:30 PM
SubDirectory	VY032525	HP Acquire Method	HP Processing Method 82y030425s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133481
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133482,VP133483 VP131783

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY021627.D	25 Mar 2025 09:17		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY021628.D	25 Mar 2025 11:31		SY/MD	Ok,M
3	VY0325SBL01	VY0325SBL01	VY021629.D	25 Mar 2025 12:04		SY/MD	Ok
4	VY0325SBS01	VY0325SBS01	VY021630.D	25 Mar 2025 12:36		SY/MD	Ok,M
5	VY0325SBSD01	VY0325SBSD01	VY021631.D	25 Mar 2025 12:59		SY/MD	Ok,M
6	Q1636-03	WC-1-VOC	VY021632.D	25 Mar 2025 13:49	vial-A Internal Standard Fail	SY/MD	ReRun
7	Q1636-07	WC-2-VOC	VY021633.D	25 Mar 2025 14:12	vial-A	SY/MD	Ok
8	Q1636-11	WC-3-VOC	VY021634.D	25 Mar 2025 14:35	vial-A	SY/MD	Ok
9	Q1636-15	WC-4-VOC	VY021635.D	25 Mar 2025 14:59	vial-A Internal Standard Fail	SY/MD	Ok
10	Q1626-01	CO-32-1	VY021636.D	25 Mar 2025 15:22	vial-A	SY/MD	Ok
11	Q1636-19	WC-5-VOC	VY021637.D	25 Mar 2025 15:46	vial-A	SY/MD	Ok
12	Q1634-03	HR-04-032425	VY021638.D	25 Mar 2025 16:09	vial-B Internal Standard Fail	SY/MD	Ok
13	Q1634-01	HR-01-032425	VY021639.D	25 Mar 2025 16:33	vial-B Not purge	SY/MD	Not Ok
14	Q1635-03	TP-2-VOC	VY021640.D	25 Mar 2025 16:56	vial-B	SY/MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VY021641.D	25 Mar 2025 17:19		SY/MD	Ok,M

M : Manual Integration