

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626
 Lab File ID: VY021628.D Date Analyzed: 03/25/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	208523	7.71	310165	8.62	269506	11.42
UPPER LIMIT	417046	8.207	620330	9.116	539012	11.92
LOWER LIMIT	104262	7.207	155083	8.116	134753	10.92
EPA SAMPLE NO.						
CO-32-1	215826	7.71	421203	8.62	386051	11.41
VY0325SBL01	234987	7.71	440273	8.62	390435	11.41
VY0325SBS01	176531	7.71	283151	8.62	250065	11.41
VY0325SBSD01	175250	7.71	281285	8.62	248863	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.