

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WALS01
 Lab Code: CHEM Case No.: Q1626 SAS No.: Q1626 SDG NO.: Q1626
 Lab File ID: VY021628.D Date Analyzed: 03/25/2025
 Instrument ID: MSVOA_Y Time Analyzed: 11:31
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	138648	13.346				
UPPER LIMIT	277296	13.846				
LOWER LIMIT	69324	12.846				
EPA SAMPLE NO.						
CO-32-1	154220	13.35				
VY0325SBL01	148940	13.35				
VY0325SBS01	132222	13.35				
VY0325SBSD01	132727	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.