

Initial and Continuing Calibration Blank Summary

Client: Tully Construction Co., Inc.

SDG No.: Q1627

Project: DOT Harper Street Yard - North Side

RunNo.: LB135179

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/25/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/25/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	0.00096	0.0025	J	0.00096	0.005	03/25/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00096	0.005	03/25/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	03/25/2025