



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1627

Contract: TULL02

Lab Code: CHEM

Case No.: Q1627

SAS No.: Q1627

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB61	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	08:38	LB135183

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.

SDG No.: Q1627

Contract: TULL02

Lab Code: CHEM

Case No.: Q1627

SAS No.: Q1627

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB60	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	08:43	LB135183
CCB61	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	09:21	LB135183
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	09:44	LB135183
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	03/26/2025	09:55	LB135183

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Tully Construction Co., Inc.</u>		SDG No.: <u>Q1627</u>							
Contract: <u>TULL02</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1627</u>	SAS No.: <u>Q1627</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	11:18	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	11:18	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	11:18	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	11:18	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	11:18	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	11:18	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	11:18	LB135196

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc.		SDG No.: Q1627							
Contract: TULL02	Lab Code: CHEM	Case No.: Q1627	SAS No.: Q1627						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	11:53	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	11:53	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	11:53	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	11:53	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	11:53	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	11:53	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	11:53	LB135196
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	12:45	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	12:45	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	12:45	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	12:45	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	12:45	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	12:45	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	12:45	LB135196
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	13:39	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	13:39	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	13:39	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	13:39	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	13:39	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	13:39	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	13:39	LB135196
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	14:34	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	14:34	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	14:34	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	14:34	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	14:34	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	14:34	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	14:34	LB135196
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	15:07	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	15:07	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	15:07	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	15:07	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	15:07	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	15:07	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	15:07	LB135196
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	15:58	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	15:58	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	15:58	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	15:58	LB135196

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Construction Co., Inc. **SDG No.:** Q1627
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1627 **SAS No.:** Q1627

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	15:58	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	15:58	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	15:58	LB135196
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	16:48	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	16:48	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	16:48	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	16:48	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	16:48	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	16:48	LB135196
CCB08	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	16:48	LB135196
	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	17:39	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	17:39	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	17:39	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	17:39	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	17:39	LB135196
CCB09	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	17:39	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	17:39	LB135196
	Arsenic	20.0	+/-20.0	U	20.0	P	03/26/2025	18:04	LB135196
	Barium	100	+/-100	U	100	P	03/26/2025	18:04	LB135196
	Cadmium	6.00	+/-6.00	U	6.00	P	03/26/2025	18:04	LB135196
	Chromium	10.0	+/-10.0	U	10.0	P	03/26/2025	18:04	LB135196
	Lead	12.0	+/-12.0	U	12.0	P	03/26/2025	18:04	LB135196
	Selenium	20.0	+/-20.0	U	20.0	P	03/26/2025	18:04	LB135196
	Silver	10.0	+/-10.0	U	10.0	P	03/26/2025	18:04	LB135196