

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **Level:** LOW **SDG No.:** Q1627
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1627 **SAS No.:** Q1627
Matrix: Water **Sample ID:** Q1636-20 **Client ID:** WC-5DUP
Percent Solids for Sample: NA **Duplicate ID** Q1636-20DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	1850		1850		0		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	50.0	U	50.0	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tully Construction Co., Inc. **Level:** LOW **SDG No.:** Q1627
Contract: TULL02 **Lab Code:** CHEM **Case No.:** Q1627 **SAS No.:** Q1627
Matrix: Water **Sample ID:** Q1636-20MS **Client ID:** WC-5MSD
Percent Solids for Sample: NA **Duplicate ID** Q1636-20MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	3900		3950		1		P
Barium	ug/L	20	3010		2980		1		P
Cadmium	ug/L	20	933		941		1		P
Chromium	ug/L	20	1960		1980		1		P
Lead	ug/L	20	4440		4430		0		P
Mercury	ug/L	20	35.0		38.0		8		CV
Selenium	ug/L	20	9680		9870		2		P
Silver	ug/L	20	369		369		0		P