

RETENTION TIMES OF INITIAL CALIBRATION

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

Instrument ID: ECD_S Calibration Date(s): 03/27/2025 03/27/2025

Calibration Times: 10:13 11:49

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS029558.D</u>	RT 500 = <u>PS029559.D</u>
	RT 750 = <u>PS029560.D</u>	RT 1000 = <u>PS029561.D</u>
		RT 1500 = <u>PS029562.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.08	9.08	9.08	9.08	9.08	9.08	8.98	9.18
2,4-D	8.23	8.22	8.22	8.22	8.22	8.22	8.12	8.32
2,4-DCAA	7.12	7.12	7.12	7.11	7.11	7.12	7.02	7.22



Contract:	<u>TULL02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1627</u>	SAS No.:	<u>Q1627</u>	SDG NO.:	<u>Q1627</u>
Instrument ID:	<u>ECD_S</u>	Calibration Date(s):		<u>03/27/2025</u>	<u>03/27/2025</u>		
		Calibration Times:		<u>10:13</u>	<u>11:49</u>		

LAB FILE ID:	RT 200 =	<u>PS029558.D</u>	RT 500 =	<u>PS029559.D</u>
RT 750 =	<u>PS029560.D</u>	RT 1000 =	<u>PS029561.D</u>	RT 1500 =
			<u>PS029562.D</u>	

[illegible]