

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** TULL02

**Lab Code:** CHEM      **Case No.:** Q1627      **SAS No.:** Q1627      **SDG NO.:** Q1627

**Continuing Calib Date:** 03/28/2025      **Initial Calibration Date(s):** 03/27/2025      03/27/2025

**Continuing Calib Time:** 04:27      **Initial Calibration Time(s):** 10:13      11:49

**GC Column:** RTX-CLP      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.12       | 7.12      | 7.02      | 7.22 | 0.00       |
| 2,4-D            | 8.22       | 8.22      | 8.12      | 8.32 | 0.00       |
| 2,4,5-TP(Silvex) | 9.08       | 9.08      | 8.98      | 9.18 | 0.00       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** TULL02

**Lab Code:** CHEM      **Case No.:** Q1627      **SAS No.:** Q1627      **SDG NO.:** Q1627

**Continuing Calib Date:** 03/28/2025      **Initial Calibration Date(s):** 03/27/2025      03/27/2025

**Continuing Calib Time:** 04:27      **Initial Calibration Time(s):** 10:13      11:49

**GC Column:** RTX-CLP2      **ID:** 0.32      (mm)

| COMPOUND         | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|------------------|------------|-----------|-----------|------|------------|
|                  |            |           | FROM      | TO   |            |
| 2,4-DCAA         | 7.54       | 7.54      | 7.44      | 7.64 | 0.01       |
| 2,4-D            | 8.75       | 8.75      | 8.65      | 8.85 | 0.01       |
| 2,4,5-TP(Silvex) | 9.63       | 9.63      | 9.53      | 9.73 | 0.00       |