

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 03/27/2025

Lab Sample No.: HSTDCCC750 Data File : PS029576.D Time Analyzed: 18:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.082	8.982	9.182	700.560	712.500	-1.7
2,4-D	8.221	8.122	8.322	712.890	705.000	1.1
2,4-DCAA	7.115	7.015	7.215	723.930	750.000	-3.5

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL01 Date Analyzed: 03/27/2025

Lab Sample No.: HSTDCCC750 Data File : PS029576.D Time Analyzed: 18:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.631	9.534	9.734	710.800	712.500	-0.2
2,4-D	8.746	8.648	8.848	691.910	705.000	-1.9
2,4-DCAA	7.535	7.437	7.637	732.800	750.000	-2.3