

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 03/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS029612.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.081	8.982	9.182	742.110	712.500	4.2
2,4-D	8.220	8.122	8.322	764.850	705.000	8.5
2,4-DCAA	7.114	7.015	7.215	761.470	750.000	1.5

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/27/2025 03/27/2025

Client Sample No.: CCAL04 Date Analyzed: 03/28/2025

Lab Sample No.: HSTDCCC750 Data File : PS029612.D Time Analyzed: 09:15

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.628	9.534	9.734	747.960	712.500	5.0
2,4-D	8.744	8.648	8.848	727.670	705.000	3.2
2,4-DCAA	7.533	7.437	7.637	767.240	750.000	2.3