

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032825\
 Data File : PS029560.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Mar 2025 11:01
 Operator : AR\AJ
 Sample : HSTDICC750
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_S
 ClientSampleId :
 HSTDICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Mar 27 12:20:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032825.M
 Quant Title : 8080.M
 QLast Update : Thu Mar 27 12:20:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30M x 0.32mm x0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
4) S	2,4-DCAA	7.115	7.537	2806.2E6	636.0E6	750.000	750.000
Target Compounds							
1) T	Dalapon	2.593	2.610	4098.4E6	1379.6E6	682.500	682.500
2) T	3,5-DICHL...	6.304	6.525	3918.3E6	848.5E6	697.500	697.500
3) T	4-Nitroph...	6.921	7.073	773.8E6	542.6E6	682.500	682.500
5) T	DICAMBA	7.294	7.726	10932.8E6	3386.3E6	705.000	705.000
6) T	MCPD	7.474	7.832	820.3E6	148.7E6	70.500	70.500
7) T	MCPA	7.620	8.068	1020.8E6	201.6E6	69.750	69.750
8) T	DICHLORPROP	7.991	8.428	3015.6E6	851.8E6	705.000	705.000
9) T	2,4-D	8.222	8.748	2834.6E6	952.7E6	705.000	705.000
10) T	Pentachlo...	8.511	9.257	44006.2E6	15974.1E6	712.500	712.500
11) T	2,4,5-TP ...	9.082	9.634	16465.8E6	6321.2E6	712.500	712.500
12) T	2,4,5-T	9.373	10.043	14929.4E6	5713.5E6	712.500	712.500
13) T	2,4-DB	9.943	10.602	1427.0E6	624.4E6	712.500	712.500
14) T	DINOSEB	11.126	10.974	12335.2E6	4173.2E6	705.000	705.000
15) T	Picloram	10.951	12.035	11134.9E6	9119.6E6	712.500	712.500
16) T	DCPA	11.424	12.001	20585.5E6	8138.9E6	720.000	720.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_S\Data\PS032825\
Data File : PS029560.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Mar 2025 11:01
Operator : AR\AJ
Sample : HSTDICC750
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_S
ClientSampleId :
HSTDICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Mar 27 12:20:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_S\Method\PS032825.M
Quant Title : 8080.M
QLast Update : Thu Mar 27 12:20:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30M x 0.32mm x 0.5 Signal #2 Info : 30M x 0.32mm x 0.25µm

