

Surrogate Summary

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS029557.D	PIBLK-PS029557.D	2,4-DCAA	1	500	455	91		39	175
		2,4-DCAA	2	500	523	105		39	175
I.BLK-PS029575.D	PIBLK-PS029575.D	2,4-DCAA	1	500	501	100		39	175
		2,4-DCAA	2	500	540	108		39	175
Q1609-03MS	WC-SCRN-01-CMS	2,4-DCAA	1	500	489	98		39	175
		2,4-DCAA	2	500	488	98		39	175
Q1609-03MSD	WC-SCRN-01-CMSD	2,4-DCAA	1	500	494	99		39	175
		2,4-DCAA	2	500	487	97		39	175
I.BLK-PS029587.D	PIBLK-PS029587.D	2,4-DCAA	1	500	518	104		39	175
		2,4-DCAA	2	500	541	108		39	175
Q1627-01	GRID-LINE-1.2-NORTH	2,4-DCAA	1	500	466	93		39	175
		2,4-DCAA	2	500	458	92		39	175
PB167312BL	PB167312BL	2,4-DCAA	1	500	495	99		39	175
		2,4-DCAA	2	500	483	97		39	175
I.BLK-PS029599.D	PIBLK-PS029599.D	2,4-DCAA	1	500	528	106		39	175
		2,4-DCAA	2	500	545	109		39	175
PB167312BS	PB167312BS	2,4-DCAA	1	500	545	109		39	175
		2,4-DCAA	2	500	531	106		39	175
PB167275TB	PB167275TB	2,4-DCAA	1	500	528	106		39	175
		2,4-DCAA	2	500	521	104		39	175
I.BLK-PS029611.D	PIBLK-PS029611.D	2,4-DCAA	1	500	529	106		39	175
		2,4-DCAA	2	500	554	111		39	175