

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: 8151A

DataFile : PS029578.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Client Sample ID: Q1609-03MSD	WC-SCRN-01-CMSD											
	2,4-D	50	0	53.1	ug/L	106		0		65	135	20
	2,4,5-TP(Silvex)	50	0	50.1	ug/L	100		0		62	139	20