

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094851.D Time Analyzed: 19:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.052	8.956	9.156	50.890	50.000	1.8
Endrin	6.570	6.474	6.674	45.150	50.000	-9.7
gamma-BHC (Lindane)	4.325	4.227	4.427	49.900	50.000	-0.2
Heptachlor	4.912	4.815	5.015	46.600	50.000	-6.8
Heptachlor epoxide	5.681	5.583	5.783	47.380	50.000	-5.2
Methoxychlor	7.498	7.400	7.600	51.020	50.000	2.0
Tetrachloro-m-xylene	3.536	3.438	3.638	51.370	50.000	2.7

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL01 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094851.D Time Analyzed: 19:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.903	7.807	8.007	42.010	50.000	-16.0
Endrin	5.630	5.534	5.734	51.350	50.000	2.7
gamma-BHC (Lindane)	3.601	3.504	3.704	57.590	50.000	15.2
Heptachlor	3.939	3.842	4.042	54.370	50.000	8.7
Heptachlor epoxide	4.721	4.625	4.825	53.160	50.000	6.3
Methoxychlor	6.603	6.507	6.707	51.470	50.000	2.9
Tetrachloro-m-xylene	2.770	2.672	2.872	56.870	50.000	13.7