

### CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094863.D Time Analyzed: 22:01

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Decachlorobiphenyl   | 9.051 | 8.956     | 9.156 | 49.060             | 50.000            | -1.9  |
| Endrin               | 6.571 | 6.474     | 6.674 | 42.730             | 50.000            | -14.5 |
| gamma-BHC (Lindane)  | 4.325 | 4.227     | 4.427 | 46.840             | 50.000            | -6.3  |
| Heptachlor           | 4.912 | 4.815     | 5.015 | 44.480             | 50.000            | -11.0 |
| Heptachlor epoxide   | 5.680 | 5.583     | 5.783 | 45.420             | 50.000            | -9.2  |
| Methoxychlor         | 7.498 | 7.400     | 7.600 | 47.920             | 50.000            | -4.2  |
| Tetrachloro-m-xylene | 3.536 | 3.438     | 3.638 | 47.390             | 50.000            | -5.2  |

### CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL02 Date Analyzed: 03/25/2025

Lab Sample No.: PSTDCCC050 Data File : PL094863.D Time Analyzed: 22:01

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Decachlorobiphenyl   | 7.903 | 7.807     | 8.007 | 49.180             | 50.000            | -1.6 |
| Endrin               | 5.630 | 5.534     | 5.734 | 49.920             | 50.000            | -0.2 |
| gamma-BHC (Lindane)  | 3.601 | 3.504     | 3.704 | 51.790             | 50.000            | 3.6  |
| Heptachlor           | 3.939 | 3.842     | 4.042 | 49.530             | 50.000            | -0.9 |
| Heptachlor epoxide   | 4.721 | 4.625     | 4.825 | 50.350             | 50.000            | 0.7  |
| Methoxychlor         | 6.603 | 6.507     | 6.707 | 49.960             | 50.000            | -0.1 |
| Tetrachloro-m-xylene | 2.769 | 2.672     | 2.872 | 52.200             | 50.000            | 4.4  |