

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/27/2025

Lab Sample No.: PSTDCCC050 Data File : PL094895.D Time Analyzed: 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.051	8.956	9.156	53.080	50.000	6.2
Endrin	6.570	6.474	6.674	47.990	50.000	-4.0
gamma-BHC (Lindane)	4.325	4.227	4.427	49.500	50.000	-1.0
Heptachlor	4.913	4.815	5.015	47.920	50.000	-4.2
Heptachlor epoxide	5.681	5.583	5.783	48.800	50.000	-2.4
Methoxychlor	7.497	7.400	7.600	54.050	50.000	8.1
Tetrachloro-m-xylene	3.537	3.438	3.638	49.950	50.000	-0.1

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL03 Date Analyzed: 03/27/2025

Lab Sample No.: PSTDCCC050 Data File : PL094895.D Time Analyzed: 08:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.902	7.807	8.007	51.450	50.000	2.9
Endrin	5.630	5.534	5.734	54.500	50.000	9.0
gamma-BHC (Lindane)	3.602	3.504	3.704	55.940	50.000	11.9
Heptachlor	3.939	3.842	4.042	53.750	50.000	7.5
Heptachlor epoxide	4.721	4.625	4.825	54.160	50.000	8.3
Methoxychlor	6.602	6.507	6.707	54.150	50.000	8.3
Tetrachloro-m-xylene	2.770	2.672	2.872	54.950	50.000	9.9