

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/27/2025

Lab Sample No.: PSTDCCC050 Data File : PL094902.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.049	8.956	9.156	51.790	50.000	3.6
Endrin	6.568	6.474	6.674	46.160	50.000	-7.7
gamma-BHC (Lindane)	4.324	4.227	4.427	48.230	50.000	-3.5
Heptachlor	4.912	4.815	5.015	46.730	50.000	-6.5
Heptachlor epoxide	5.679	5.583	5.783	47.430	50.000	-5.1
Methoxychlor	7.496	7.400	7.600	53.000	50.000	6.0
Tetrachloro-m-xylene	3.536	3.438	3.638	48.380	50.000	-3.2

CALIBRATION VERIFICATION SUMMARY

Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 03/11/2025 03/11/2025

Client Sample No.: CCAL04 Date Analyzed: 03/27/2025

Lab Sample No.: PSTDCCC050 Data File : PL094902.D Time Analyzed: 12:06

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.900	7.807	8.007	51.730	50.000	3.5
Endrin	5.628	5.534	5.734	54.210	50.000	8.4
gamma-BHC (Lindane)	3.601	3.504	3.704	55.140	50.000	10.3
Heptachlor	3.938	3.842	4.042	53.260	50.000	6.5
Heptachlor epoxide	4.719	4.625	4.825	53.580	50.000	7.2
Methoxychlor	6.601	6.507	6.707	55.630	50.000	11.3
Tetrachloro-m-xylene	2.769	2.672	2.872	54.670	50.000	9.3