

Report of Analysis

Client:	Tully Construction Co., Inc.		Date Collected:	03/19/25	
Project:	DOT Harper Street Yard - North Side		Date Received:	03/20/25	
Client Sample ID:	WC-SCRN-01-CMS		SDG No.:	Q1627	
Lab Sample ID:	Q1609-03MS		Matrix:	TCLP	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	100	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	TCLP Pesticide	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL094857.D	1	03/25/25 12:40	03/25/25 20:39	PB167311

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
58-89-9	gamma-BHC (Lindane)	5.00		0.037	0.50	ug/L
76-44-8	Heptachlor	4.70		0.027	0.50	ug/L
1024-57-3	Heptachlor epoxide	4.80		0.096	0.50	ug/L
72-20-8	Endrin	4.90		0.032	0.50	ug/L
72-43-5	Methoxychlor	5.10		0.11	0.50	ug/L
8001-35-2	Toxaphene	1.70	U	1.70	10.0	ug/L
57-74-9	Chlordane	0.88	U	0.88	5.00	ug/L
SURROGATES						
2051-24-3	Decachlorobiphenyl	21.6		43 - 140	108%	SPK: 20
877-09-8	Tetrachloro-m-xylene	19.8		77 - 126	99%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit