

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: 8081B

DataFile : PL094857.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec Qual	RPD Qual	Low	Limits High	RPD
Client Sample ID:	WC-SCRN-01-CMS									
Q1609-03MS	gamma-BHC (Lindane)	5	0	5.00	ug/L	100		60	152	
	Heptachlor	5	0	4.70	ug/L	94		56	147	
	Heptachlor epoxide	5	0	4.80	ug/L	96		77	143	
	Endrin	5	0	4.90	ug/L	98		76	144	
	Methoxychlor	5	0	5.10	ug/L	102		70	142	