

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: 8081B

DataFile : PL094858.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID:	WC-SCRN-01-CMSD											
Q1609-03MSD	gamma-BHC (Lindane)	5	0	4.90	ug/L	98		2		60	152	20
	Heptachlor	5	0	4.70	ug/L	94		0		56	147	20
	Heptachlor epoxide	5	0	4.90	ug/L	98		2		77	143	20
	Endrin	5	0	4.90	ug/L	98		0		76	144	20
	Methoxychlor	5	0	5.10	ug/L	102		0		70	142	20