

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: 8270E

DataFile: BF142096.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167310BS	Pyridine	50	46.9	ug/L	94				29	97	
	1,4-Dichlorobenzene	50	50.6	ug/L	101				76	103	
	2-Methylphenol	50	49.5	ug/L	99				69	109	
	3+4-Methylphenols	50	48.0	ug/L	96				67	106	
	Hexachloroethane	50	50.1	ug/L	100				76	118	
	Nitrobenzene	50	50.0	ug/L	100				58	106	
	Hexachlorobutadiene	50	53.3	ug/L	107		*		69	101	
	2,4,6-Trichlorophenol	50	50.3	ug/L	101				61	110	
	2,4,5-Trichlorophenol	50	53.2	ug/L	106				70	106	
	2,4-Dinitrotoluene	50	57.8	ug/L	116		*		60	115	
	Hexachlorobenzene	50	53.3	ug/L	107		*		73	106	
	Pentachlorophenol	100	100	ug/L	100				47	114	