

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG No.: Q1627

Instrument ID: MSVOA_N Calibration Date/Time: 03/25/2025 10:27

Lab File ID: VN086090.D Init. Calib. Date(s): 03/18/2025 03/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:44 14:09

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.594	0.677		13.97	20
1,1-Dichloroethene	0.512	0.557		8.79	20
2-Butanone	0.302	0.344		13.91	20
Carbon Tetrachloride	0.604	0.587		-2.82	20
Chloroform	1.107	1.133		2.35	20
Benzene	1.443	1.540		6.72	20
1,2-Dichloroethane	0.503	0.478		-4.97	20
Trichloroethene	0.374	0.361		-3.48	20
Tetrachloroethene	0.399	0.387		-3.01	20
Chlorobenzene	1.144	1.148	0.3	0.35	20
1,2-Dichloroethane-d4	0.685	0.681		-0.58	20
Dibromofluoromethane	0.349	0.358		2.58	20
Toluene-d8	1.267	1.350		6.55	20
4-Bromofluorobenzene	0.452	0.492		8.85	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.