

Surrogate Summary

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1627-01	GRID-LINE-1.2-NORTH	1,2-Dichloroethane-d4	50	56.1	112	74	125
		Dibromofluoromethane	50	52.3	105	75	124
		Toluene-d8	50	51.9	104	86	113
		4-Bromofluorobenzene	50	45.1	90	77	121

Surrogate Summary

SDG No.: Q1627

Client: Tully Construction Co., Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN0325WBL01	VN0325WBL01	1,2-Dichloroethane-d4	50	53.7	107	74	125
		Dibromofluoromethane	50	50.5	101	75	124
		Toluene-d8	50	47.6	95	86	113
		4-Bromofluorobenzene	50	43.0	86	77	121
VN0325WBS01	VN0325WBS01	1,2-Dichloroethane-d4	50	51.8	104	74	125
		Dibromofluoromethane	50	53.1	106	75	124
		Toluene-d8	50	53.9	108	86	113
		4-Bromofluorobenzene	50	52.2	104	77	121
VN0325WBSD0	VN0325WBSD01	1,2-Dichloroethane-d4	50	52.4	105	74	125
		Dibromofluoromethane	50	51.4	103	75	124
		Toluene-d8	50	52.9	106	86	113
		4-Bromofluorobenzene	50	50.5	101	77	121