

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL02
Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627
Lab File ID: VN086090.D Date Analyzed: 03/25/2025
Instrument ID: MSVOA_N Time Analyzed: 10:27
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	231359	8.22	376282	9.10	346115	11.87
UPPER LIMIT	462718	8.724	752564	9.6	692230	12.365
LOWER LIMIT	115680	7.724	188141	8.6	173058	11.365
EPA SAMPLE NO.						
GRID-LINE-1.2-NORTH	166171	8.22	310861	9.10	301467	11.87
VN0325WBL01	201384	8.22	370671	9.10	325680	11.87
VN0325WBS01	200228	8.22	324541	9.10	286017	11.87
VN0325WBSD01	194343	8.22	321428	9.10	280205	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.