284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL02

Lab Code: CHEM Case No.: Q1627 SAS No.: Q1627 SDG NO.: Q1627

Lab File ID: VN086090.D Date Analyzed: 03/25/2025

Instrument ID: MSVOA_N Time Analyzed: 10:27

GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	181838	13.788				
UPPER LIMIT	363676	14.288				
LOWER LIMIT	90919	13.288				
EPA SAMPLE NO.						
GRID-LINE-1.2-NORTH	114842	13.79				
VN0325WBL01	130926	13.79				
VN0325WBS01	139430	13.79				
VN0325WBSD01	140993	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.