

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1628
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1628 **SAS No.:** Q1628
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	13.0	2.0	650	-18	22	03/24/2025	10:58	LB135158
	Iron	97400	101000	96	85600	116500	03/24/2025	10:58	LB135158
	Lead	5.40			-12	12	03/24/2025	10:58	LB135158
ICSAB01	Copper	520	511	102	434	588	03/24/2025	11:02	LB135158
	Iron	98300	99300	99	84400	114500	03/24/2025	11:02	LB135158
	Lead	54.7	49.0	112	37	61	03/24/2025	11:02	LB135158
ICSA	Copper	514	2.0	25700	-18	22	03/24/2025	11:13	LB135158
	Iron	88000	101000	87	85600	116500	03/24/2025	11:13	LB135158
	Lead	70.9			-12	12	03/24/2025	11:13	LB135158
ICSAB	Copper	549	511	107	434	588	03/24/2025	11:17	LB135158
	Iron	95900	99300	97	84400	114500	03/24/2025	11:17	LB135158
	Lead	59.8	49.0	122	37	61	03/24/2025	11:17	LB135158