

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q1628</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1628</u> sas no.: <u>Q1628</u>
matrix: <u>Water</u>	sample id: <u>Q1595-02</u>	client id: <u>SPRING-2025MS</u>
Percent Solids for Sample: NA	Spiked ID: Q1595-02MS	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	457		308		150	99		P
Iron	ug/L	75 - 125	1540		172		1500	91		P
Lead	ug/L	75 - 125	420		6.00	U	500	84		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q1628</u>
contract: <u>TULL01</u>	lab code: <u>CHEM</u>	case no.: <u>Q1628</u> sas no.: <u>Q1628</u>
matrix: <u>Water</u>	sample id: <u>Q1595-02</u>	client id: <u>SPRING-2025MSD</u>
Percent Solids for Sample: NA	Spiked ID: Q1595-02MSD	Percent Solids for Spike Sample: NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	455		308		150	98		P
Iron	ug/L	75 - 125	1540		172		1500	91		P
Lead	ug/L	75 - 125	419		6.00	U	500	84		P