

ANALYTICAL RESULTS SUMMARY

VOLATILE ORGANICS
SEMI-VOLATILE ORGANICS

PROJECT NAME : NAVFAC NWIRP BETHPAGE, NY SITE 1 OU-2 - 32258

AECOM TECHNICAL SERVICES, INC.

13640 Briarwick Drive

Austin, TX - 78729

Phone No: 512-454-4797

ORDER ID : Q1629

ATTENTION : Eleanor Vivadou



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	9
2.1) VOCMS Group1- Case Narrative	9
2.2) SVOC-SIMGroup1- Case Narrative	12
3) Qualifier Page	14
4) QA Checklist	15
5) VOCMS Group1 Data	16
6) SVOC-SIMGroup1 Data	110
7) Shipping Document	157
7.1) CHAIN OF CUSTODY	158
7.2) ROC	159
7.3) Lab Certificate	161
7.4) Internal COC	162

1
2
3
4
5
6
7

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION
FORM S-I

SAMPLE IDENTIFICATION AND ANALYTICAL REQUIREMENT SUMMARY

NYSDEC Sample ID/Code	Laboratory Sample ID/Code	VOA GC/MS (Method #)	BNA GC/MS (Method #)	VOA GC (Method #)	Pest PCBs (Method #)	Metals (Method #)	Other (Method #)
TT174I1-20250320	Q1629-01	8260-Low	8270-Modified				
TT162S1-20250320	Q1629-02	8260-Low	8270-Modified				
TT188S1-20250320	Q1629-03	8260-Low	8270-Modified				
RW09-MW01D1-20250320	Q1629-07	8260-Low	8270-Modified				
RW09-MW01D2-20250320	Q1629-08	8260-Low	8270-Modified				
RW09-MW01D3-20250320	Q1629-09	8260-Low	8270-Modified				
TT150S1-20250320	Q1629-10	8260-Low	8270-Modified				
FB03-20250321	Q1629-11	8260-Low	8270-Modified				
TB	Q1629-12	8260-Low	8270-Modified				
RW09-MW01S-20250320	Q1629-14	8260-Low	8270-Modified				

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-IIa

SAMPLE PREPARATION AND ANALYSIS SUMMARY SEMIVOLATILE (BNA) ANALYSES

Laboratory Sample ID	Matrix	Date Collected	Date Rec'd at Lab	Date Extracted	Date Analyzed
Q1629-01	Water	03/20/25	03/21/25	03/25/25	03/29/25
Q1629-02	Water	03/20/25	03/21/25	03/25/25	03/29/25
Q1629-03	Water	03/20/25	03/21/25	03/25/25	03/28/25
Q1629-07	Water	03/20/25	03/21/25	03/25/25	03/29/25
Q1629-08	Water	03/20/25	03/21/25	03/25/25	03/29/25
Q1629-09	Water	03/20/25	03/21/25	03/25/25	03/29/25
Q1629-10	Water	03/20/25	03/21/25	03/25/25	03/29/25
Q1629-11	Water	03/21/25	03/21/25	03/25/25	03/29/25
Q1629-14	Water	03/20/25	03/21/25	03/25/25	03/31/25

* Details For Test : SVOC-SIMGroup1

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-IIb

SAMPLE PREPARATION AND ANALYSIS SUMMARY VOLATILE (VOA) ANALYSES

Laboratory Sample ID	Matrix	Date Collected	Date Rec'd at Lab	Date Extracted	Date Analyzed
Q1629-01	Water	03/20/25	03/21/25		03/25/25
Q1629-02	Water	03/20/25	03/21/25		03/25/25
Q1629-03	Water	03/20/25	03/21/25		03/26/25
Q1629-07	Water	03/20/25	03/21/25		03/25/25
Q1629-08	Water	03/20/25	03/21/25		03/25/25
Q1629-09	Water	03/20/25	03/21/25		03/25/25
Q1629-10	Water	03/20/25	03/21/25		03/25/25
Q1629-11	Water	03/21/25	03/21/25		03/25/25
Q1629-12	Water	03/21/25	03/21/25		03/25/25
Q1629-14	Water	03/20/25	03/21/25		03/26/25

* Details For Test : VOCMS Group1

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-III

SAMPLE PREPARATION AND ANALYSIS SUMMARY MISCELLANEOUS ORGANIC ANALYSES

Laboratory Sample ID	Matrix	Analytical Protocol	Extraction Method	Auxiliary Cleanup	Dil/Conc Factor
Q1629-01	Water	8260-Low	5030		
Q1629-02	Water	8260-Low	5030		
Q1629-03	Water	8260-Low	5030		
Q1629-04	Water	8260-Low	5030		
Q1629-05	Water	8260-Low	5030		
Q1629-07	Water	8260-Low	5030		
Q1629-08	Water	8260-Low	5030		
Q1629-09	Water	8260-Low	5030		
Q1629-10	Water	8260-Low	5030		
Q1629-11	Water	8260-Low	5030		
Q1629-12	Water	8260-Low	5030		
Q1629-14	Water	8260-Low	5030		
Q1629-15	Water	8260-Low	5030		
Q1629-16	Water	8260-Low	5030		

NEW YORK STATE DEPARTMENT OF ENVIRONMENTAL CONSERVATION

FORM S-III

SAMPLE PREPARATION AND ANALYSIS SUMMARY MISCELLANEOUS ORGANIC ANALYSES

Laboratory Sample ID	Matrix	Analytical Protocol	Extraction Method	Auxiliary Cleanup	Dil/Conc Factor
Q1629-01	Water	8270-Modified	3510C		
Q1629-02	Water	8270-Modified	3510C		
Q1629-03	Water	8270-Modified	3510C		
Q1629-04	Water	8270-Modified	3510C		
Q1629-05	Water	8270-Modified	3510C		
Q1629-07	Water	8270-Modified	3510C		
Q1629-08	Water	8270-Modified	3510C		
Q1629-09	Water	8270-Modified	3510C		
Q1629-10	Water	8270-Modified	3510C		
Q1629-11	Water	8270-Modified	3510C		
Q1629-14	Water	8270-Modified	3510C		
Q1629-15	Water	8270-Modified	3510C		
Q1629-16	Water	8270-Modified	3510C		

Cover Page

Order ID : Q1629

Project ID : NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Client : AECOM Technical Services, Inc.

Lab Sample Number

Q1629-01
Q1629-02
Q1629-03
Q1629-04
Q1629-05
Q1629-07
Q1629-08
Q1629-09
Q1629-10
Q1629-11
Q1629-12
Q1629-14
Q1629-15
Q1629-16

Client Sample Number

TT174I1-20250320
TT162S1-20250320
TT188S1-20250320
TT188S1-20250320MS
TT188S1-20250320MSD
RW09-MW01D1-20250320
RW09-MW01D2-20250320
RW09-MW01D3-20250320
TT150S1-20250320
FB03-20250321
TB
RW09-MW01S-20250320
RW09-MW01S-20250320MS
RW09-MW01S-20250320MSD

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 4/3/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project # N/A

Chemtech Project # Q1629

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

14 Water samples were received on 03/21/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_X were done using GC column DB-624UI 20m 0.18mm 1.0 um. Cat#121-1324UI The analysis of VOCMS Group1 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for TT188S1-20250320MSD [4-Bromofluorobenzene - 115%] and RW09-MW01S-20250320MSD [Toluene-d8 - 76%], Surrogate fail in only MSD but MS and Parent Sample is Passing For Surrogate Recoveries therefore no Corrective action was taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {Q1629-04MS} with File ID: VX045467.D recoveries met the requirements for all compounds except for Methyl Acetate[143%], due to matrix interference.

The MSD {Q1629-05MSD} with File ID: VX045468.D recoveries met the acceptable requirements except for 1,2-Dibromo-3-Chloropropane[129%], 1,2-Dichloroethane[130%], cis-1,3-Dichloropropene[126%] and t-1,3-Dichloropropene[129%], due to matrix interference.

The RPD for {Q1629-05MSD} with File ID: VX045468.D met criteria except for Bromomethane[26%], Carbon disulfide[42%], Chloroethane[23%], Chloromethane[30%], Cyclohexane[23%], Dichlorodifluoromethane[33%], Methyl

Acetate[33%], Methylcyclohexane[22%] and Vinyl chloride[25%], due to difference in results of MS and MSD.

The sample # TT188S1-20250320MS and TT188S1-20250320MSD are failing for Methyl Acetate, 1,2-Dichloroethane, t-1,3-Dichloropropene, cis-1,3-Dichloropropene, 1,2-Dibromo-3-Chloropropane and the original sample(TT188S1- 20250320) is reported with M flag for this compounds.

The Blank Spike for {VX0325WBS02} with File ID: VX045436.D met requirements for all samples except for Methyl Acetate[147%] is failing high but no positive hit in associate sample therefore no corrective action taken.

The Blank Spike for {VX0326WBS01} with File ID: VX045460.D met requirements for all samples except for Methyl Acetate[140%] failing high but no positive hit in associate sample therefore no corrective action taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration File ID VX045431.D met the requirements except for 1,1,2-Trichlorotrifluoroethane, 2-Hexanone, cis-1,3-Dichloropropene, Dichlorodifluoromethane, Methylcyclohexane and t-1,3-Dichloropropene, are failing high but no positive hit in associate samples therefore no corrective action taken.

The Continuous Calibration File ID VX045457.D met the requirements except for Dichlorodifluoromethane, Methylcyclohexane and t-1,3-Dichloropropene, are failing high but no positive hit in associate samples therefore no corrective action taken.
The Tuning criteria met requirements.

E. Additional Comments:

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

CASE NARRATIVE

AECOM Technical Services, Inc.

Project Name: NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258

Project Manager: Ernie Wu

Chemtech Project # Q1629

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

14 Water samples were received on 03/21/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements except for PB167295BL Which is not associated for required compound.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements. Except Terphenyl-d14, This compound is not associate for required compound.

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1629

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MOHAMMAD AHMED

Date: 04/03/2025

LAB CHRONICLE

OrderID:	Q1629	OrderDate:	3/21/2025 1:55:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	I41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1629-01	TT174I1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-02	TT162S1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-03	TT188S1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/26/25	03/21/25
Q1629-07	RW09-MW01D1-2025 0320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-08	RW09-MW01D2-2025 0320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-09	RW09-MW01D3-2025 0320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-10	TT150S1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-11	FB03-20250321	Water	VOCMS Group1	8260-Low	03/21/25		03/25/25	03/21/25
Q1629-12	TB	Water	VOCMS Group1	8260-Low	03/21/25		03/25/25	03/21/25
Q1629-14	RW09-MW01S-20250 320	Water	VOCMS Group1	8260-Low	03/20/25		03/26/25	03/21/25

Hit Summary Sheet SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	TT174I1-20250320								
Q1629-01	TT174I1-20250320	Water	1,1-Dichloroethene	1.20		0.23	0.75	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	Acetone	2.50	J	1.50	3.80	5.00	ug/L
Q1629-01	TT174I1-20250320	Water	Methyl tert-butyl Ether	0.39	J	0.16	0.50	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	1,1-Dichloroethane	3.40		0.23	0.50	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	Trichloroethene	2.00		0.090	0.75	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	Tetrachloroethene	0.42	J	0.23	0.50	1.00	ug/L
			Total Voc :	9.91					
			Total Concentration:	9.91					
Client ID:	TT162S1-20250320								
Q1629-02	TT162S1-20250320	Water	Acetone	1.80	J	1.50	3.80	5.00	ug/L
Q1629-02	TT162S1-20250320	Water	Trichloroethene	11.9		0.090	0.75	1.00	ug/L
Q1629-02	TT162S1-20250320	Water	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
			Total Voc :	14.8					
			Total Concentration:	14.8					
Client ID:	TT188S1-20250320								
Q1629-03	TT188S1-20250320	Water	Acetone	2.20	J	1.50	3.80	5.00	ug/L
Q1629-03	TT188S1-20250320	Water	Trichloroethene	5.30		0.090	0.75	1.00	ug/L
			Total Voc :	7.50					
			Total Concentration:	7.50					
Client ID:	RW09-MW01D1-20250320								
Q1629-07	RW09-MW01D1-20	Water	Acetone	1.80	J	1.50	3.80	5.00	ug/L
			Total Voc :	1.80					
			Total Concentration:	1.80					
Client ID:	RW09-MW01D2-20250320								
Q1629-08	RW09-MW01D2-20	Water	Acetone	2.00	J	1.50	3.80	5.00	ug/L
			Total Voc :	2.00					
			Total Concentration:	2.00					
Client ID:	TT150S1-20250320								
Q1629-10	TT150S1-20250320	Water	Acetone	2.50	J	1.50	3.80	5.00	ug/L
Q1629-10	TT150S1-20250320	Water	Trichloroethene	11.6		0.090	0.75	1.00	ug/L
Q1629-10	TT150S1-20250320	Water	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
			Total Voc :	15.2					
			Total Concentration:	15.2					
Client ID:	FB03-20250321								
Q1629-11	FB03-20250321	Water	Acetone	4.00	J	1.50	3.80	5.00	ug/L
			Total Voc :	4.00					
			Total Concentration:	4.00					
Client ID:	TB								
Q1629-12	TB	Water	Acetone	1.60	J	1.50	3.80	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				1.60					
Total Concentration:				1.60					

A

B

C

D

E

F

G



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT174I1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045445.D	1		03/25/25 17:09	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	1.20		0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.39	J	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	3.40		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	2.00		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT174I1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045445.D	1		03/25/25 17:09	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.42	J	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.0		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	47.2		89 - 112		94%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.5		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70100	5.55				
540-36-3	1,4-Difluorobenzene	138000	6.757				
3114-55-4	Chlorobenzene-d5	122000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	47200	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT174I1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045445.D	1		03/25/25 17:09	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/21/25	
Client Sample ID:	TT162S1-20250320			SDG No.:	Q1629	
Lab Sample ID:	Q1629-02			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045446.D	1		03/25/25 17:32	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	11.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT162S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045446.D	1		03/25/25 17:32	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.9		81 - 118		112%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	47.5		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	68000	5.544				
540-36-3	1,4-Difluorobenzene	134000	6.757				
3114-55-4	Chlorobenzene-d5	122000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	51000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT162S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045446.D	1		03/25/25 17:32	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320			SDG No.:	Q1629	
Lab Sample ID:	Q1629-03			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045466.D	1		03/26/25 13:43	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.20	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQM	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	UM	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	5.30		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045466.D	1		03/26/25 13:43	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	UM	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	UM	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	UM	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.5		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	46.4		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.8		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	68100	5.544				
540-36-3	1,4-Difluorobenzene	134000	6.757				
3114-55-4	Chlorobenzene-d5	120000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	47700	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045466.D	1		03/26/25 13:43	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045447.D	1		03/25/25 17:55	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045447.D	1		03/25/25 17:55	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.3		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	47.8		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.7		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	74700	5.543				
540-36-3	1,4-Difluorobenzene	147000	6.757				
3114-55-4	Chlorobenzene-d5	136000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	52500	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045447.D	1		03/25/25 17:55	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D2-20250320			SDG No.:	Q1629	
Lab Sample ID:	Q1629-08			Matrix:	Water	
Analytical Method:	SW8260			% Solid:	0	
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID :	0.18	Level :	LOW	
Prep Method :						

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045448.D	1		03/25/25 18:18	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.00	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D2-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045448.D	1		03/25/25 18:18	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.1		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	50.6		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	76200	5.55				
540-36-3	1,4-Difluorobenzene	148000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	55200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D2-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045448.D	1		03/25/25 18:18	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D3-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045449.D	1		03/25/25 18:41	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D3-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045449.D	1		03/25/25 18:41	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.2		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.0		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71700	5.544				
540-36-3	1,4-Difluorobenzene	137000	6.757				
3114-55-4	Chlorobenzene-d5	129000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	54400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D3-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045449.D	1		03/25/25 18:41	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT150S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045450.D	1		03/25/25 19:05	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	11.6		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/20/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/21/25		
Client Sample ID:	TT150S1-20250320			SDG No.:	Q1629		
Lab Sample ID:	Q1629-10			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045450.D	1		03/25/25 19:05	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.6		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	47.3		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	72000	5.55				
540-36-3	1,4-Difluorobenzene	142000	6.757				
3114-55-4	Chlorobenzene-d5	131000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	55200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT150S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045450.D	1		03/25/25 19:05	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	FB03-20250321		SDG No.:	Q1629	
Lab Sample ID:	Q1629-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045444.D	1		03/25/25 16:46	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	4.00	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	FB03-20250321		SDG No.:	Q1629	
Lab Sample ID:	Q1629-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045444.D	1		03/25/25 16:46	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.4		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	47.9		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.6		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71100	5.544				
540-36-3	1,4-Difluorobenzene	141000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	49200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	FB03-20250321		SDG No.:	Q1629	
Lab Sample ID:	Q1629-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045444.D	1		03/25/25 16:46	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TB		SDG No.:	Q1629	
Lab Sample ID:	Q1629-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045443.D	1		03/25/25 16:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.60	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.			Date Collected:	03/21/25		
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258			Date Received:	03/21/25		
Client Sample ID:	TB			SDG No.:	Q1629		
Lab Sample ID:	Q1629-12			Matrix:	Water		
Analytical Method:	SW8260			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045443.D	1		03/25/25 16:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.0		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	47.3		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70900	5.549				
540-36-3	1,4-Difluorobenzene	138000	6.756				
3114-55-4	Chlorobenzene-d5	128000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	51700	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TB		SDG No.:	Q1629	
Lab Sample ID:	Q1629-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045443.D	1		03/25/25 16:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045461.D	1		03/26/25 11:46	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045461.D	1		03/26/25 11:46	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.9		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	47.1		89 - 112		94%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	72700	5.544				
540-36-3	1,4-Difluorobenzene	140000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	52400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045461.D	1		03/26/25 11:46	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q1629**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1629-01	TT174I1-20250320	1,2-Dichloroethane-d4	50	54.0	108	81	118
		Dibromofluoromethane	50	48.8	98	80	119
		Toluene-d8	50	47.2	94	89	112
		4-Bromofluorobenzene	50	48.5	97	85	114
Q1629-02	TT162S1-20250320	1,2-Dichloroethane-d4	50	55.9	112	81	118
		Dibromofluoromethane	50	50.1	100	80	119
		Toluene-d8	50	47.5	95	89	112
		4-Bromofluorobenzene	50	51.9	104	85	114
Q1629-03	TT188S1-20250320	1,2-Dichloroethane-d4	50	53.5	107	81	118
		Dibromofluoromethane	50	49.4	99	80	119
		Toluene-d8	50	46.4	93	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
Q1629-04MS	TT188S1-20250320MS	1,2-Dichloroethane-d4	50	53.7	107	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	46.5	93	89	112
		4-Bromofluorobenzene	50	51.1	102	85	114
Q1629-05MSD	TT188S1-20250320MSD	1,2-Dichloroethane-d4	50	56.3	113	81	118
		Dibromofluoromethane	50	53.0	106	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	57.5	115 *	85	114
Q1629-07	RW09-MW01D1-20250320	1,2-Dichloroethane-d4	50	55.3	111	81	118
		Dibromofluoromethane	50	49.7	99	80	119
		Toluene-d8	50	47.9	96	89	112
		4-Bromofluorobenzene	50	50.7	101	85	114
Q1629-08	RW09-MW01D2-20250320	1,2-Dichloroethane-d4	50	55.1	110	81	118
		Dibromofluoromethane	50	50.6	101	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	50.5	101	85	114
Q1629-09	RW09-MW01D3-20250320	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	48.0	96	89	112
		4-Bromofluorobenzene	50	51.7	103	85	114
Q1629-10	TT150S1-20250320	1,2-Dichloroethane-d4	50	54.6	109	81	118
		Dibromofluoromethane	50	49.7	99	80	119
		Toluene-d8	50	47.3	95	89	112
		4-Bromofluorobenzene	50	50.9	102	85	114
Q1629-11	FB03-20250321	1,2-Dichloroethane-d4	50	55.4	111	81	118
		Dibromofluoromethane	50	49.5	99	80	119
		Toluene-d8	50	47.9	96	89	112
		4-Bromofluorobenzene	50	49.6	99	85	114
Q1629-12	TB	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	47.3	95	89	112
		4-Bromofluorobenzene	50	50.6	101	85	114
Q1629-14	RW09-MW01S-20250320	1,2-Dichloroethane-d4	50	50.9	102	81	118
		Dibromofluoromethane	50	48.4	97	80	119
		Toluene-d8	50	47.1	94	89	112
		4-Bromofluorobenzene	50	49.3	99	85	114
Q1629-15MS	RW09-MW01S-20250320MS	1,2-Dichloroethane-d4	50	54.0	108	81	118
		Dibromofluoromethane	50	51.2	102	80	119

Surrogate Summary

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1629-15MS	RW09-MW01S-20250320MS	Toluene-d8	50	47.4	95	89	112
		4-Bromofluorobenzene	50	53.5	107	85	114
Q1629-16MSD	RW09-MW01S-20250320MSD	1,2-Dichloroethane-d4	50	41.9	84	81	118
		Dibromofluoromethane	50	40.1	80	80	119
		Toluene-d8	50	38.1	76 *	89	112
		4-Bromofluorobenzene	50	42.5	85	85	114
VX0325WBL01	VX0325WBL01	1,2-Dichloroethane-d4	50	58.3	117	81	118
		Dibromofluoromethane	50	54.2	108	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	54.8	110	85	114
VX0325WBS02	VX0325WBS02	1,2-Dichloroethane-d4	50	56.6	113	81	118
		Dibromofluoromethane	50	53.4	107	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	54.3	109	85	114
VX0326WBL01	VX0326WBL01	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	48.1	96	80	119
		Toluene-d8	50	46.4	93	89	112
		4-Bromofluorobenzene	50	47.1	94	85	114
VX0326WBS01	VX0326WBS01	1,2-Dichloroethane-d4	50	53.8	108	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	48.3	97	89	112
		4-Bromofluorobenzene	50	52.8	106	85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

										Limits	
Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD	Low	High	RPD
		Result			Qual	Qual					
Lab Sample ID :	Q1629-15MS	Client Sample ID :	RW09-MW01S-20250320MS					Datafile :	VX045462.D		
Dichlorodifluoromethane	50	0	61.2	ug/L	122				32	152	
Chloromethane	50	0	50.8	ug/L	102				50	139	
Vinyl chloride	50	0	48.8	ug/L	98				58	137	
Bromomethane	50	0	54.3	ug/L	109				53	141	
Chloroethane	50	0	54.0	ug/L	108				60	138	
Trichlorofluoromethane	50	0	53.8	ug/L	108				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	59.0	ug/L	118				70	136	
1,1-Dichloroethene	50	0	52.8	ug/L	106				71	131	
Acetone	250	0	260	ug/L	104				39	160	
Carbon disulfide	50	0	48.6	ug/L	97				64	133	
Methyl tert-butyl Ether	50	0	51.9	ug/L	104				71	124	
Methyl Acetate	50	0	48.3	ug/L	97				56	136	
Methylene Chloride	50	0	50.4	ug/L	101				74	124	
trans-1,2-Dichloroethene	50	0	52.9	ug/L	106				75	124	
1,1-Dichloroethane	50	0	52.3	ug/L	105				77	125	
Cyclohexane	50	0	53.6	ug/L	107				71	130	
2-Butanone	250	0	270	ug/L	108				56	143	
Carbon Tetrachloride	50	0	54.9	ug/L	110				72	136	
cis-1,2-Dichloroethene	50	0	51.1	ug/L	102				78	123	
Bromochloromethane	50	0	49.7	ug/L	99				78	123	
Chloroform	50	0	53.9	ug/L	108				79	124	
1,1,1-Trichloroethane	50	0	56.0	ug/L	112				74	131	
Methylcyclohexane	50	0	57.5	ug/L	115				72	132	
Benzene	50	0	51.3	ug/L	103				79	120	
1,2-Dichloroethane	50	0	57.6	ug/L	115				73	128	
Trichloroethene	50	0	49.6	ug/L	99				79	123	
1,2-Dichloropropane	50	0	50.6	ug/L	101				78	122	
Bromodichloromethane	50	0	53.9	ug/L	108				79	125	
4-Methyl-2-Pentanone	250	0	280	ug/L	112				67	130	
Toluene	50	0	52.9	ug/L	106				80	121	
t-1,3-Dichloropropene	50	0	57.5	ug/L	115				73	127	
cis-1,3-Dichloropropene	50	0	55.7	ug/L	111				75	124	
1,1,2-Trichloroethane	50	0	51.9	ug/L	104				80	119	
2-Hexanone	250	0	280	ug/L	112				57	139	
Dibromochloromethane	50	0	54.4	ug/L	109				74	126	
1,2-Dibromoethane	50	0	53.2	ug/L	106				77	121	
Tetrachloroethene	50	0	53.2	ug/L	106				74	129	
Chlorobenzene	50	0	52.9	ug/L	106				82	118	
Ethyl Benzene	50	0	54.9	ug/L	110				79	121	
m/p-Xylenes	100	0	110	ug/L	110				80	121	
o-Xylene	50	0	54.1	ug/L	108				78	122	
Styrene	50	0	56.1	ug/L	112				78	123	
Bromoform	50	0	54.6	ug/L	109				66	130	
Isopropylbenzene	50	0	55.1	ug/L	110				72	131	
1,1,2,2-Tetrachloroethane	50	0	50.8	ug/L	102				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec			RPD	Limits		
					Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	52.0	ug/L	104				80	119	
1,4-Dichlorobenzene	50	0	51.2	ug/L	102				79	118	
1,2-Dichlorobenzene	50	0	52.2	ug/L	104				80	119	
1,2-Dibromo-3-Chloropropane	50	0	58.9	ug/L	118				62	128	
1,2,4-Trichlorobenzene	50	0	51.7	ug/L	103				69	130	
1,2,3-Trichlorobenzene	50	0	51.3	ug/L	103				69	129	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

											Limits	
Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD	Low	High	RPD	
		Result			Qual	Qual						
Lab Sample ID :	Q1629-16MSD	Client Sample ID :	RW09-MW01S-20250320MSD					Datafile :	VX045463.D			
Dichlorodifluoromethane	50	0	61.4	ug/L	123		0		32	152	20	
Chloromethane	50	0	48.8	ug/L	98		4		50	139	20	
Vinyl chloride	50	0	48.1	ug/L	96		1		58	137	20	
Bromomethane	50	0	55.4	ug/L	111		2		53	141	20	
Chloroethane	50	0	56.6	ug/L	113		5		60	138	20	
Trichlorofluoromethane	50	0	53.1	ug/L	106		1		65	141	20	
1,1,2-Trichlorotrifluoroethane	50	0	58.3	ug/L	117		1		70	136	20	
1,1-Dichloroethene	50	0	52.9	ug/L	106		0		71	131	20	
Acetone	250	0	260	ug/L	104		0		39	160	20	
Carbon disulfide	50	0	50.0	ug/L	100		3		64	133	20	
Methyl tert-butyl Ether	50	0	53.7	ug/L	107		3		71	124	20	
Methyl Acetate	50	0	49.5	ug/L	99		2		56	136	20	
Methylene Chloride	50	0	50.6	ug/L	101		0		74	124	20	
trans-1,2-Dichloroethene	50	0	53.1	ug/L	106		0		75	124	20	
1,1-Dichloroethane	50	0	52.7	ug/L	105		1		77	125	20	
Cyclohexane	50	0	54.9	ug/L	110		2		71	130	20	
2-Butanone	250	0	270	ug/L	108		0		56	143	20	
Carbon Tetrachloride	50	0	58.7	ug/L	117		7		72	136	20	
cis-1,2-Dichloroethene	50	0	52.9	ug/L	106		3		78	123	20	
Bromochloromethane	50	0	48.4	ug/L	97		3		78	123	20	
Chloroform	50	0	55.1	ug/L	110		2		79	124	20	
1,1,1-Trichloroethane	50	0	57.6	ug/L	115		3		74	131	20	
Methylcyclohexane	50	0	60.0	ug/L	120		4		72	132	20	
Benzene	50	0	53.4	ug/L	107		4		79	120	20	
1,2-Dichloroethane	50	0	60.7	ug/L	121		5		73	128	20	
Trichloroethene	50	0	54.0	ug/L	108		8		79	123	20	
1,2-Dichloropropane	50	0	54.1	ug/L	108		7		78	122	20	
Bromodichloromethane	50	0	57.6	ug/L	115		7		79	125	20	
4-Methyl-2-Pentanone	250	0	290	ug/L	116		4		67	130	20	
Toluene	50	0	55.7	ug/L	111		5		80	121	20	
t-1,3-Dichloropropene	50	0	61.9	ug/L	124		7		73	127	20	
cis-1,3-Dichloropropene	50	0	59.3	ug/L	119		6		75	124	20	
1,1,2-Trichloroethane	50	0	54.5	ug/L	109		5		80	119	20	
2-Hexanone	250	0	300	ug/L	120		7		57	139	20	
Dibromochloromethane	50	0	58.5	ug/L	117		7		74	126	20	
1,2-Dibromoethane	50	0	56.7	ug/L	113		6		77	121	20	
Tetrachloroethene	50	0	53.5	ug/L	107		1		74	129	20	
Chlorobenzene	50	0	54.3	ug/L	109		3		82	118	20	
Ethyl Benzene	50	0	56.4	ug/L	113		3		79	121	20	
m/p-Xylenes	100	0	110	ug/L	110		0		80	121	20	
o-Xylene	50	0	54.6	ug/L	109		1		78	122	20	
Styrene	50	0	57.2	ug/L	114		2		78	123	20	
Bromoform	50	0	56.6	ug/L	113		4		66	130	20	
Isopropylbenzene	50	0	56.7	ug/L	113		3		72	131	20	
1,1,2,2-Tetrachloroethane	50	0	53.3	ug/L	107		5		71	121	20	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec			RPD	Limits		
					Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	53.6	ug/L	107		3		80	119	20
1,4-Dichlorobenzene	50	0	53.5	ug/L	107		4		79	118	20
1,2-Dichlorobenzene	50	0	53.2	ug/L	106		2		80	119	20
1,2-Dibromo-3-Chloropropane	50	0	60.7	ug/L	121		3		62	128	20
1,2,4-Trichlorobenzene	50	0	53.9	ug/L	108		4		69	130	20
1,2,3-Trichlorobenzene	50	0	52.6	ug/L	105		3		69	129	20

A

B

C

D

E

F

G

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
Lab Sample ID :	Q1629-04MS	Client Sample ID :	TT188S1-20250320MS						Datafile :	VX045467.D	
Dichlorodifluoromethane	50	0	46.4	ug/L	93				32	152	
Chloromethane	50	0	40.7	ug/L	81				50	139	
Vinyl chloride	50	0	40.6	ug/L	81				58	137	
Bromomethane	50	0	46.0	ug/L	92				53	141	
Chloroethane	50	0	43.2	ug/L	86				60	138	
Trichlorofluoromethane	50	0	46.8	ug/L	94				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	54.8	ug/L	110				70	136	
1,1-Dichloroethene	50	0	47.9	ug/L	96				71	131	
Acetone	250	2.20	280	ug/L	111				39	160	
Carbon disulfide	50	0	34.3	ug/L	69				64	133	
Methyl tert-butyl Ether	50	0	54.1	ug/L	108				71	124	
Methyl Acetate	50	0	71.5	ug/L	143	*			56	136	
Methylene Chloride	50	0	49.1	ug/L	98				74	124	
trans-1,2-Dichloroethene	50	0	47.9	ug/L	96				75	124	
1,1-Dichloroethane	50	0	52.7	ug/L	105				77	125	
Cyclohexane	50	0	46.2	ug/L	92				71	130	
2-Butanone	250	0	290	ug/L	116				56	143	
Carbon Tetrachloride	50	0	53.1	ug/L	106				72	136	
cis-1,2-Dichloroethene	50	0	50.5	ug/L	101				78	123	
Bromochloromethane	50	0	53.5	ug/L	107				78	123	
Chloroform	50	0	55.2	ug/L	110				79	124	
1,1,1-Trichloroethane	50	0	55.3	ug/L	111				74	131	
Methylcyclohexane	50	0	48.9	ug/L	98				72	132	
Benzene	50	0	49.7	ug/L	99				79	120	
1,2-Dichloroethane	50	0	58.4	ug/L	117				73	128	
Trichloroethene	50	5.30	53.5	ug/L	96				79	123	
1,2-Dichloropropane	50	0	51.4	ug/L	103				78	122	
Bromodichloromethane	50	0	56.3	ug/L	113				79	125	
4-Methyl-2-Pentanone	250	0	290	ug/L	116				67	130	
Toluene	50	0	51.6	ug/L	103				80	121	
t-1,3-Dichloropropene	50	0	56.5	ug/L	113				73	127	
cis-1,3-Dichloropropene	50	0	55.2	ug/L	110				75	124	
1,1,2-Trichloroethane	50	0	53.5	ug/L	107				80	119	
2-Hexanone	250	0	300	ug/L	120				57	139	
Dibromochloromethane	50	0	56.2	ug/L	112				74	126	
1,2-Dibromoethane	50	0	53.5	ug/L	107				77	121	
Tetrachloroethene	50	0	49.3	ug/L	99				74	129	
Chlorobenzene	50	0	51.8	ug/L	104				82	118	
Ethyl Benzene	50	0	53.2	ug/L	106				79	121	
m/p-Xylenes	100	0	100	ug/L	100				80	121	
o-Xylene	50	0	51.7	ug/L	103				78	122	
Styrene	50	0	54.0	ug/L	108				78	123	
Bromoform	50	0	54.9	ug/L	110				66	130	
Isopropylbenzene	50	0	51.8	ug/L	104				72	131	
1,1,2,2-Tetrachloroethane	50	0	52.0	ug/L	104				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	50.4	ug/L	101				80	119	
1,4-Dichlorobenzene	50	0	50.0	ug/L	100				79	118	
1,2-Dichlorobenzene	50	0	52.3	ug/L	105				80	119	
1,2-Dibromo-3-Chloropropane	50	0	60.7	ug/L	121				62	128	
1,2,4-Trichlorobenzene	50	0	51.8	ug/L	104				69	130	
1,2,3-Trichlorobenzene	50	0	51.1	ug/L	102				69	129	

A

B

C

D

E

F

G

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
Lab Sample ID :	Q1629-05MSD	Client Sample ID :	TT188S1-20250320MSD					Datafile :	VX045468.D		
Dichlorodifluoromethane	50	0	65.0	ug/L	130		33	*	32	152	20
Chloromethane	50	0	55.3	ug/L	111		30	*	50	139	20
Vinyl chloride	50	0	52.1	ug/L	104		25	*	58	137	20
Bromomethane	50	0	59.9	ug/L	120		26	*	53	141	20
Chloroethane	50	0	54.7	ug/L	109		23	*	60	138	20
Trichlorofluoromethane	50	0	57.1	ug/L	114		20		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	63.4	ug/L	127		15		70	136	20
1,1-Dichloroethene	50	0	57.1	ug/L	114		18		71	131	20
Acetone	250	2.20	300	ug/L	119		7		39	160	20
Carbon disulfide	50	0	52.6	ug/L	105		42	*	64	133	20
Methyl tert-butyl Ether	50	0	58.3	ug/L	117		7		71	124	20
Methyl Acetate	50	0	51.2	ug/L	102		33	*	56	136	20
Methylene Chloride	50	0	55.0	ug/L	110		11		74	124	20
trans-1,2-Dichloroethene	50	0	57.5	ug/L	115		18		75	124	20
1,1-Dichloroethane	50	0	57.7	ug/L	115		9		77	125	20
Cyclohexane	50	0	58.0	ug/L	116		23	*	71	130	20
2-Butanone	250	0	310	ug/L	124		7		56	143	20
Carbon Tetrachloride	50	0	61.2	ug/L	122		14		72	136	20
cis-1,2-Dichloroethene	50	0	56.9	ug/L	114		12		78	123	20
Bromochloromethane	50	0	49.6	ug/L	99		8		78	123	20
Chloroform	50	0	59.6	ug/L	119		8		79	124	20
1,1,1-Trichloroethane	50	0	62.0	ug/L	124		11		74	131	20
Methylcyclohexane	50	0	61.0	ug/L	122		22	*	72	132	20
Benzene	50	0	56.5	ug/L	113		13		79	120	20
1,2-Dichloroethane	50	0	64.8	ug/L	130	*	10		73	128	20
Trichloroethene	50	5.30	60.4	ug/L	110		14		79	123	20
1,2-Dichloropropane	50	0	56.8	ug/L	114		10		78	122	20
Bromodichloromethane	50	0	61.0	ug/L	122		8		79	125	20
4-Methyl-2-Pentanone	250	0	320	ug/L	128		10		67	130	20
Toluene	50	0	59.1	ug/L	118		14		80	121	20
t-1,3-Dichloropropene	50	0	64.7	ug/L	129	*	14		73	127	20
cis-1,3-Dichloropropene	50	0	63.1	ug/L	126	*	13		75	124	20
1,1,2-Trichloroethane	50	0	57.8	ug/L	116		8		80	119	20
2-Hexanone	250	0	330	ug/L	132		10		57	139	20
Dibromochloromethane	50	0	61.7	ug/L	123		9		74	126	20
1,2-Dibromoethane	50	0	59.5	ug/L	119		11		77	121	20
Tetrachloroethene	50	0	56.7	ug/L	113		14		74	129	20
Chlorobenzene	50	0	58.3	ug/L	117		12		82	118	20
Ethyl Benzene	50	0	59.8	ug/L	120		12		79	121	20
m/p-Xylenes	100	0	120	ug/L	120		18		80	121	20
o-Xylene	50	0	58.7	ug/L	117		13		78	122	20
Styrene	50	0	60.9	ug/L	122		12		78	123	20
Bromoform	50	0	60.5	ug/L	121		10		66	130	20
Isopropylbenzene	50	0	58.0	ug/L	116		11		72	131	20
1,1,2,2-Tetrachloroethane	50	0	55.2	ug/L	110		6		71	121	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD	RPD Qual	Limits		
					Rec	Qual			Low	High	RPD
1,3-Dichlorobenzene	50	0	56.0	ug/L	112			11	80	119	20
1,4-Dichlorobenzene	50	0	54.8	ug/L	110			9	79	118	20
1,2-Dichlorobenzene	50	0	56.8	ug/L	114			8	80	119	20
1,2-Dibromo-3-Chloropropane	50	0	64.5	ug/L	129	*		6	62	128	20
1,2,4-Trichlorobenzene	50	0	55.8	ug/L	112			7	69	130	20
1,2,3-Trichlorobenzene	50	0	54.6	ug/L	109			7	69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045436.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0325WBS02	Dichlorodifluoromethane	20	19.5	ug/L	98			32	152	
	Chloromethane	20	16.0	ug/L	80			50	139	
	Vinyl chloride	20	16.0	ug/L	80			58	137	
	Bromomethane	20	19.6	ug/L	98			53	141	
	Chloroethane	20	19.7	ug/L	99			60	138	
	Trichlorofluoromethane	20	20.0	ug/L	100			65	141	
	1,1,2-Trichlorotrifluoroethane	20	22.6	ug/L	113			70	136	
	1,1-Dichloroethene	20	19.6	ug/L	98			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	14.0	ug/L	70			64	133	
	Methyl tert-butyl Ether	20	21.5	ug/L	108			71	124	
	Methyl Acetate	20	29.4	ug/L	147		*	56	136	
	Methylene Chloride	20	20.2	ug/L	101			74	124	
	trans-1,2-Dichloroethene	20	20.1	ug/L	101			75	124	
	1,1-Dichloroethane	20	21.5	ug/L	108			77	125	
	Cyclohexane	20	19.7	ug/L	99			71	130	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	21.8	ug/L	109			72	136	
	cis-1,2-Dichloroethene	20	20.8	ug/L	104			78	123	
	Bromochloromethane	20	24.3	ug/L	121			78	123	
	Chloroform	20	22.3	ug/L	112			79	124	
	1,1,1-Trichloroethane	20	22.3	ug/L	112			74	131	
	Methylcyclohexane	20	20.5	ug/L	103			72	132	
	Benzene	20	20.5	ug/L	103			79	120	
	1,2-Dichloroethane	20	23.9	ug/L	119			73	128	
	Trichloroethene	20	20.5	ug/L	103			79	123	
	1,2-Dichloropropane	20	21.1	ug/L	106			78	122	
	Bromodichloromethane	20	22.7	ug/L	114			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	21.4	ug/L	107			80	121	
	t-1,3-Dichloropropene	20	22.1	ug/L	111			73	127	
	cis-1,3-Dichloropropene	20	22.4	ug/L	112			75	124	
	1,1,2-Trichloroethane	20	22.3	ug/L	112			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	22.0	ug/L	110			74	126	
	1,2-Dibromoethane	20	21.7	ug/L	109			77	121	
	Tetrachloroethene	20	21.6	ug/L	108			74	129	
	Chlorobenzene	20	21.4	ug/L	107			82	118	
	Ethyl Benzene	20	21.6	ug/L	108			79	121	
	m/p-Xylenes	40	43.1	ug/L	108			80	121	
	o-Xylene	20	21.4	ug/L	107			78	122	
	Styrene	20	21.7	ug/L	109			78	123	
	Bromoform	20	21.4	ug/L	107			66	130	
	Isopropylbenzene	20	22.5	ug/L	113			72	131	
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107			71	121	
	1,3-Dichlorobenzene	20	21.0	ug/L	105			80	119	
	1,4-Dichlorobenzene	20	20.5	ug/L	103			79	118	
	1,2-Dichlorobenzene	20	21.1	ug/L	106			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045436.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0325WBS02	1,2-Dibromo-3-Chloropropane	20	22.1	ug/L	111			62	128	
	1,2,4-Trichlorobenzene	20	20.1	ug/L	101			69	130	
	1,2,3-Trichlorobenzene	20	20.3	ug/L	102			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045460.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0326WBS01	Dichlorodifluoromethane	20	18.0	ug/L	90			32	152	
	Chloromethane	20	15.2	ug/L	76			50	139	
	Vinyl chloride	20	15.3	ug/L	77			58	137	
	Bromomethane	20	18.6	ug/L	93			53	141	
	Chloroethane	20	20.0	ug/L	100			60	138	
	Trichlorofluoromethane	20	18.3	ug/L	92			65	141	
	1,1,2-Trichlorotrifluoroethane	20	21.6	ug/L	108			70	136	
	1,1-Dichloroethene	20	17.8	ug/L	89			71	131	
	Acetone	100	100	ug/L	100			39	160	
	Carbon disulfide	20	12.8	ug/L	64			64	133	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			71	124	
	Methyl Acetate	20	27.9	ug/L	140		*	56	136	
	Methylene Chloride	20	19.0	ug/L	95			74	124	
	trans-1,2-Dichloroethene	20	18.7	ug/L	94			75	124	
	1,1-Dichloroethane	20	20.1	ug/L	101			77	125	
	Cyclohexane	20	18.2	ug/L	91			71	130	
	2-Butanone	100	100	ug/L	100			56	143	
	Carbon Tetrachloride	20	20.3	ug/L	102			72	136	
	cis-1,2-Dichloroethene	20	19.4	ug/L	97			78	123	
	Bromochloromethane	20	21.0	ug/L	105			78	123	
	Chloroform	20	21.1	ug/L	106			79	124	
	1,1,1-Trichloroethane	20	21.0	ug/L	105			74	131	
	Methylcyclohexane	20	18.5	ug/L	93			72	132	
	Benzene	20	19.1	ug/L	96			79	120	
	1,2-Dichloroethane	20	22.3	ug/L	112			73	128	
	Trichloroethene	20	18.9	ug/L	95			79	123	
	1,2-Dichloropropane	20	19.8	ug/L	99			78	122	
	Bromodichloromethane	20	21.5	ug/L	108			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	19.7	ug/L	99			80	121	
	t-1,3-Dichloropropene	20	20.4	ug/L	102			73	127	
	cis-1,3-Dichloropropene	20	20.3	ug/L	102			75	124	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	21.1	ug/L	106			74	126	
	1,2-Dibromoethane	20	20.4	ug/L	102			77	121	
	Tetrachloroethene	20	20.4	ug/L	102			74	129	
	Chlorobenzene	20	19.8	ug/L	99			82	118	
	Ethyl Benzene	20	20.4	ug/L	102			79	121	
	m/p-Xylenes	40	40.8	ug/L	102			80	121	
	o-Xylene	20	19.8	ug/L	99			78	122	
	Styrene	20	20.5	ug/L	103			78	123	
	Bromoform	20	20.6	ug/L	103			66	130	
	Isopropylbenzene	20	20.9	ug/L	104			72	131	
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102			71	121	
	1,3-Dichlorobenzene	20	19.4	ug/L	97			80	119	
	1,4-Dichlorobenzene	20	19.2	ug/L	96			79	118	
	1,2-Dichlorobenzene	20	19.9	ug/L	100			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629
Client: AECOM Technical Services, Inc.
Analytical Method: SW8260-Low **Datafile :** VX045460.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0326WBS01	1,2-Dibromo-3-Chloropropane	20	21.2	ug/L	106			62	128	
	1,2,4-Trichlorobenzene	20	18.9	ug/L	95			69	130	
	1,2,3-Trichlorobenzene	20	18.7	ug/L	94			69	129	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0325WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1629

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: VX045433.D

Lab Sample ID: VX0325WBL01

Date Analyzed: 03/25/2025

Time Analyzed: 12:22

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0325WBS02	VX0325WBS02	VX045436.D	03/25/2025
TB	Q1629-12	VX045443.D	03/25/2025
FB03-20250321	Q1629-11	VX045444.D	03/25/2025
TT174I1-20250320	Q1629-01	VX045445.D	03/25/2025
TT162S1-20250320	Q1629-02	VX045446.D	03/25/2025
RW09-MW01D1-20250320	Q1629-07	VX045447.D	03/25/2025
RW09-MW01D2-20250320	Q1629-08	VX045448.D	03/25/2025
RW09-MW01D3-20250320	Q1629-09	VX045449.D	03/25/2025
TT150S1-20250320	Q1629-10	VX045450.D	03/25/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0326WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1629

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: VX045459.D

Lab Sample ID: VX0326WBL01

Date Analyzed: 03/26/2025

Time Analyzed: 10:54

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0326WBS01	VX0326WBS01	VX045460.D	03/26/2025
RW09-MW01S-20250320	Q1629-14	VX045461.D	03/26/2025
RW09-MW01S-20250320MS	Q1629-15MS	VX045462.D	03/26/2025
RW09-MW01S-20250320MSD	Q1629-16MSD	VX045463.D	03/26/2025
TT188S1-20250320	Q1629-03	VX045466.D	03/26/2025
TT188S1-20250320MS	Q1629-04MS	VX045467.D	03/26/2025
TT188S1-20250320MSD	Q1629-05MSD	VX045468.D	03/26/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045067.D BFB Injection Date: 02/28/2025
Instrument ID: MSVOA_X BFB Injection Time: 01:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.7
75	30.0 - 60.0% of mass 95	53.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	73.8
175	5.0 - 9.0% of mass 174	5.8 (7.9) 1
176	95.0 - 101.0% of mass 174	70.6 (95.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045068.D	02/28/2025	01:27
VSTDIC005	VSTDIC005	VX045069.D	02/28/2025	02:13
VSTDIC020	VSTDIC020	VX045070.D	02/28/2025	02:37
VSTDIC050	VSTDIC050	VX045071.D	02/28/2025	03:00
VSTDIC100	VSTDIC100	VX045072.D	02/28/2025	03:23
VSTDIC150	VSTDIC150	VX045073.D	02/28/2025	03:47

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045430.D BFB Injection Date: 03/25/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.2
75	30.0 - 60.0% of mass 95	55.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	72.4
175	5.0 - 9.0% of mass 174	5.4 (7.5) 1
176	95.0 - 101.0% of mass 174	69.5 (96.1) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045431.D	03/25/2025	11:31
VX0325WBL01	VX0325WBL01	VX045433.D	03/25/2025	12:22
VX0325WBS02	VX0325WBS02	VX045436.D	03/25/2025	13:35
TB	Q1629-12	VX045443.D	03/25/2025	16:22
FB03-20250321	Q1629-11	VX045444.D	03/25/2025	16:46
TT174I1-20250320	Q1629-01	VX045445.D	03/25/2025	17:09
TT162S1-20250320	Q1629-02	VX045446.D	03/25/2025	17:32
RW09-MW01D1-20250320	Q1629-07	VX045447.D	03/25/2025	17:55
RW09-MW01D2-20250320	Q1629-08	VX045448.D	03/25/2025	18:18
RW09-MW01D3-20250320	Q1629-09	VX045449.D	03/25/2025	18:41
TT150S1-20250320	Q1629-10	VX045450.D	03/25/2025	19:05
VSTDCCC050EC	VSTDCCC050	VX045455.D	03/25/2025	21:00

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045456.D BFB Injection Date: 03/26/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:33
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.2
173	Less than 2.0% of mass 174	0.8 (1.1) 1
174	50.0 - 100.0% of mass 95	70.8
175	5.0 - 9.0% of mass 174	5.4 (7.6) 1
176	95.0 - 101.0% of mass 174	68.7 (97.1) 1
177	5.0 - 9.0% of mass 176	4.9 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045457.D	03/26/2025	10:03
VX0326WBL01	VX0326WBL01	VX045459.D	03/26/2025	10:54
VX0326WBS01	VX0326WBS01	VX045460.D	03/26/2025	11:18
RW09-MW01S-20250320	Q1629-14	VX045461.D	03/26/2025	11:46
RW09-MW01S-20250320MS	Q1629-15MS	VX045462.D	03/26/2025	12:09
RW09-MW01S-20250320MSD	Q1629-16MSD	VX045463.D	03/26/2025	12:33
TT188S1-20250320	Q1629-03	VX045466.D	03/26/2025	13:43
TT188S1-20250320MS	Q1629-04MS	VX045467.D	03/26/2025	14:06
TT188S1-20250320MSD	Q1629-05MSD	VX045468.D	03/26/2025	14:29
VSTDCCC050EC	VSTDCCC050	VX045470.D	03/26/2025	15:35

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045431.D Date Analyzed: 03/25/2025
Instrument ID: MSVOA_X Time Analyzed: 11:31
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	83260	5.54	144945	6.76	129576	10.05
UPPER LIMIT	166520	6.044	289890	7.257	259152	10.549
LOWER LIMIT	41630	5.044	72472.5	6.257	64788	9.549
EPA SAMPLE NO.						
TT174I1-20250320	70075	5.55	137552	6.76	122364	10.05
TT162S1-20250320	68003	5.54	133718	6.76	122480	10.05
RW09-MW01D1-20250320	74720	5.54	146822	6.76	135722	10.05
RW09-MW01D2-20250320	76218	5.55	148282	6.76	136799	10.05
RW09-MW01D3-20250320	71670	5.54	137481	6.76	129073	10.05
TT150S1-20250320	71974	5.55	142479	6.76	130670	10.05
FB03-20250321	71073	5.54	141090	6.76	129742	10.05
TB	70921	5.55	138131	6.76	128366	10.06
VX0325WBL01	71237	5.55	137150	6.76	123936	10.05
VX0325WBS02	88260	5.54	159282	6.76	137098	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045431.D Date Analyzed: 03/25/2025
Instrument ID: MSVOA_X Time Analyzed: 11:31
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63016	12.018				
UPPER LIMIT	126032	12.518				
LOWER LIMIT	31508	11.518				
EPA SAMPLE NO.						
TT174I1-20250320	47236	12.02				
TT162S1-20250320	50952	12.02				
RW09-MW01D1-20250320	52479	12.02				
RW09-MW01D2-20250320	55234	12.02				
RW09-MW01D3-20250320	54362	12.02				
TT150S1-20250320	55180	12.02				
FB03-20250321	49151	12.02				
TB	51682	12.02				
VX0325WBL01	50683	12.02				
VX0325WBS02	60910	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045457.D Date Analyzed: 03/26/2025
Instrument ID: MSVOA_X Time Analyzed: 10:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91964	5.54	157025	6.75	138010	10.05
UPPER LIMIT	183928	6.038	314050	7.251	276020	10.549
LOWER LIMIT	45982	5.038	78512.5	6.251	69005	9.549
EPA SAMPLE NO.						
TT188S1-20250320	68093	5.54	134396	6.76	120469	10.06
TT188S1-20250320MS	79755	5.54	146350	6.76	128351	10.06
TT188S1-20250320MSD	78228	5.55	141777	6.76	126420	10.05
RW09-MW01S-20250320	72703	5.54	139868	6.76	129727	10.05
RW09-MW01S-20250320MS	81846	5.54	149200	6.76	128995	10.05
RW09-MW01S-20250320MSD	81226	5.55	143287	6.76	127464	10.06
VX0326WBL01	65792	5.54	128574	6.76	116243	10.06
VX0326WBS01	86104	5.55	155857	6.76	136930	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045457.D Date Analyzed: 03/26/2025
Instrument ID: MSVOA_X Time Analyzed: 10:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63053	12.018				
UPPER LIMIT	126106	12.518				
LOWER LIMIT	31526.5	11.518				
EPA SAMPLE NO.						
TT188S1-20250320	47673	12.02				
TT188S1-20250320MS	58769	12.02				
TT188S1-20250320MSD	58644	12.02				
RW09-MW01S-20250320	52360	12.02				
RW09-MW01S-20250320MS	58223	12.02				
RW09-MW01S-20250320MSD	56894	12.02				
VX0326WBL01	46967	12.02				
VX0326WBS01	61262	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0325WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045433.D	1		03/25/25 12:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0325WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045433.D	1		03/25/25 12:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	58.3		81 - 118		117%	SPK: 50
1868-53-7	Dibromofluoromethane	54.2		80 - 119		108%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.8		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71200	5.55				
540-36-3	1,4-Difluorobenzene	137000	6.757				
3114-55-4	Chlorobenzene-d5	124000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50700	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0325WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045433.D	1		03/25/25 12:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045459.D	1		03/26/25 10:54	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045459.D	1		03/26/25 10:54	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.2		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	48.1		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	46.4		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	65800	5.544				
540-36-3	1,4-Difluorobenzene	129000	6.757				
3114-55-4	Chlorobenzene-d5	116000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	47000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045459.D	1		03/26/25 10:54	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBS02		SDG No.:	Q1629
Lab Sample ID:	VX0325WBS02		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045436.D	1		03/25/25 13:35	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.5		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	16.0		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.0		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	19.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.7		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.0		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	22.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	110		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	14.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.5		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	29.4		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	20.2		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.1		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.5		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.8		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.8		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	24.3		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	22.3		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	22.3		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	20.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	23.9		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.5		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	22.7		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	21.4		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBS02		SDG No.:	Q1629
Lab Sample ID:	VX0325WBS02		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045436.D	1		03/25/25 13:35	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	22.1		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	22.4		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	22.3		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	22.0		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	21.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	21.6		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	21.4		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	21.6		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	43.1		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	21.4		0.12	0.50	1.00	ug/L
100-42-5	Styrene	21.7		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	21.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	22.5		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	21.0		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.1		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	22.1		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.1		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.3		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.6		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	53.4		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	88300	5.544				
540-36-3	1,4-Difluorobenzene	159000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	60900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBS02		SDG No.:	Q1629
Lab Sample ID:	VX0325WBS02		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045436.D	1		03/25/25 13:35	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBS01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045460.D	1		03/26/25 11:18	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	18.0		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	15.2		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	15.3		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.3		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	100		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	12.8		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	27.9		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.7		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.1		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	100		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	20.3		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.4		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	21.0		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	21.1		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.1		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	22.3		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.8		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	21.5		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.7		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBS01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045460.D	1		03/26/25 11:18	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.4		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.1		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.4		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.4		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.4		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	40.8		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	20.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.9		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.4		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.2		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.9		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	21.2		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.9		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.7		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	48.3		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	86100	5.55				
540-36-3	1,4-Difluorobenzene	156000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	61300	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBS01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045460.D	1		03/26/25 11:18	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045467.D	1		03/26/25 14:06	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	46.4		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	40.7		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	40.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	46.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	43.2		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	46.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	54.8		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	47.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	280		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	34.3		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	54.1		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	71.5		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	49.1		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	47.9		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	52.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	46.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	290		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	53.1		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	50.5		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	53.5		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	55.2		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	55.3		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	48.9		0.16	0.50	1.00	ug/L
71-43-2	Benzene	49.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	58.4		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	53.5		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	51.4		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	56.3		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	290		0.68	2.50	5.00	ug/L
108-88-3	Toluene	51.6		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045467.D	1		03/26/25 14:06	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	56.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	55.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	53.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	300		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	56.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	53.5		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	49.3		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	51.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	53.2		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	100		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	51.7		0.12	0.50	1.00	ug/L
100-42-5	Styrene	54.0		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	54.9		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	51.8		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	52.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	50.4		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	50.0		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	52.3		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	60.7		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	51.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	51.1		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.7		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	46.5		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	79800	5.544				
540-36-3	1,4-Difluorobenzene	146000	6.757				
3114-55-4	Chlorobenzene-d5	128000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	58800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045467.D	1		03/26/25 14:06	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-15MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045462.D	1		03/26/25 12:09	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	61.2		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	50.8		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	48.8		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	54.3		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	54.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	53.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	59.0		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	52.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	260		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	48.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	51.9		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	48.3		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	50.4		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	52.9		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	52.3		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	53.6		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	270		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	54.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	51.1		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	49.7		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	53.9		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	56.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	57.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	51.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	57.6		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	49.6		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	50.6		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	53.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	280		0.68	2.50	5.00	ug/L
108-88-3	Toluene	52.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-15MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045462.D	1		03/26/25 12:09	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	57.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	55.7		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	51.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	280		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	54.4		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	53.2		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	53.2		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	52.9		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	54.9		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	110		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	54.1		0.12	0.50	1.00	ug/L
100-42-5	Styrene	56.1		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	54.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	55.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	50.8		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	52.0		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	51.2		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	52.2		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	58.9		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	51.7		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	51.3		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.0		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	47.4		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.5		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	81800	5.544				
540-36-3	1,4-Difluorobenzene	149000	6.757				
3114-55-4	Chlorobenzene-d5	129000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	58200	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-15MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045462.D	1		03/26/25 12:09	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045468.D	1		03/26/25 14:29	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	65.0		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	55.3		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	52.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	59.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	54.7		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	57.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	63.4		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	57.1		0.23	0.75	1.00	ug/L
67-64-1	Acetone	300		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	52.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	58.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	51.2		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	55.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	57.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	57.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	58.0		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	310		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	61.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	56.9		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	49.6		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	59.6		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	62.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	61.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	56.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	64.8		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	60.4		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	56.8		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	61.0		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	320		0.68	2.50	5.00	ug/L
108-88-3	Toluene	59.1		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045468.D	1		03/26/25 14:29	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	64.7		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	63.1		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	57.8		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	330		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	61.7		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	59.5		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	56.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	58.3		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	59.8		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	120		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	58.7		0.12	0.50	1.00	ug/L
100-42-5	Styrene	60.9		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	60.5		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	58.0		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	55.2		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	56.0		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	54.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	56.8		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	64.5		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	55.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	54.6		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.3		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	53.0		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	57.5	*	85 - 114		115%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	78200	5.55				
540-36-3	1,4-Difluorobenzene	142000	6.757				
3114-55-4	Chlorobenzene-d5	126000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	58600	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045468.D	1		03/26/25 14:29	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-16MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045463.D	1		03/26/25 12:33	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	61.4		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	48.8		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	48.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	55.4		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	56.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	53.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	58.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	52.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	260		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	50.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	53.7		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	49.5		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	50.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	53.1		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	52.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	54.9		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	270		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	58.7		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	52.9		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	48.4		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	55.1		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	57.6		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	60.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	53.4		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	60.7		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	54.0		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	54.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	57.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	290		0.68	2.50	5.00	ug/L
108-88-3	Toluene	55.7		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-16MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045463.D	1		03/26/25 12:33	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	61.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	59.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	54.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	300		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	58.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	56.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	53.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	54.3		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	56.4		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	110		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	54.6		0.12	0.50	1.00	ug/L
100-42-5	Styrene	57.2		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	56.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	56.7		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	53.3		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	53.6		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	53.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	53.2		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	60.7		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	53.9		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	52.6		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	41.9		81 - 118		84%	SPK: 50
1868-53-7	Dibromofluoromethane	40.1		80 - 119		80%	SPK: 50
2037-26-5	Toluene-d8	38.1	*	89 - 112		76%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.5		85 - 114		85%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	81200	5.55				
540-36-3	1,4-Difluorobenzene	143000	6.757				
3114-55-4	Chlorobenzene-d5	127000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	56900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-16MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045463.D	1		03/26/25 12:33	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045068.D		RRF005 = VX045069.D		RRF020 = VX045070.D		RRF050 = VX045071.D		RRF100 = VX045072.D		RRF150 = VX045073.D	
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD					
Dichlorodifluoromethane	0.624	0.646	0.646	0.627	0.626	0.607	0.629	2.4					
Chloromethane	0.755	0.821	0.828	0.753	0.721	0.744	0.770	5.7					
Vinyl Chloride	0.773	0.755	0.774	0.761	0.765	0.758	0.764	1					
Bromomethane		0.337	0.298	0.292	0.284	0.291	0.300	7					
Chloroethane	0.373	0.421	0.366	0.373	0.297	0.286	0.352	14.5					
Trichlorofluoromethane	0.978	1.061	1.050	1.050	0.982	0.948	1.012	4.7					
1,1,2-Trichlorotrifluoroethane	0.526	0.613	0.595	0.594	0.596	0.563	0.581	5.4					
1,1-Dichloroethene	0.609	0.620	0.612	0.623	0.620	0.603	0.614	1.3					
Acetone	0.414	0.363	0.384	0.351	0.345	0.356	0.369	7					
Carbon Disulfide	1.584	1.582	1.587	1.660	1.708	1.698	1.636	3.6					
Methyl tert-butyl Ether	1.955	1.913	2.127	2.083	2.132	2.158	2.061	5					
Methyl Acetate	0.954	0.835	0.903	0.867	0.869	0.899	0.888	4.6					
Methylene Chloride	0.806	0.730	0.752	0.698	0.694	0.706	0.731	5.8					
trans-1,2-Dichloroethene	0.540	0.619	0.603	0.631	0.634	0.616	0.607	5.7					
1,1-Dichloroethane	1.200	1.223	1.280	1.242	1.270	1.264	1.247	2.5					
Cyclohexane		1.021	1.087	1.108	1.142	1.052	1.082	4.3					
2-Butanone	0.476	0.545	0.610	0.579	0.553	0.570	0.555	8.1					
Carbon Tetrachloride	0.463	0.463	0.447	0.468	0.489	0.463	0.465	3					
cis-1,2-Dichloroethene	0.687	0.746	0.765	0.762	0.767	0.769	0.749	4.2					
Bromochloromethane	0.557	0.633	0.613	0.603	0.563	0.596	0.594	5					
Chloroform	1.206	1.247	1.278	1.246	1.230	1.225	1.239	2					
1,1,1-Trichloroethane	0.908	0.992	1.009	1.024	1.044	1.025	1.000	4.8					
Methylcyclohexane	0.464	0.530	0.560	0.585	0.607	0.550	0.549	9.1					
Benzene	1.321	1.459	1.491	1.496	1.497	1.424	1.448	4.7					
1,2-Dichloroethane	0.487	0.525	0.545	0.528	0.524	0.520	0.521	3.6					
Trichloroethene	0.319	0.351	0.339	0.341	0.354	0.336	0.340	3.7					
1,2-Dichloropropane	0.354	0.378	0.382	0.371	0.376	0.373	0.372	2.7					
Bromodichloromethane	0.478	0.503	0.536	0.524	0.528	0.528	0.516	4.2					
4-Methyl-2-Pentanone	0.535	0.570	0.647	0.610	0.579	0.579	0.587	6.5					
Toluene	0.716	0.872	0.892	0.898	0.874	0.845	0.849	8					

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629
Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025
Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47
GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:								
RRF001 = VX045068.D			RRF005 = VX045069.D			RRF020 = VX045070.D		
RRF050 = VX045071.D			RRF100 = VX045072.D			RRF150 = VX045073.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.304	0.389	0.436	0.469	0.490	0.502	0.431	17.3
cis-1,3-Dichloropropene	0.404	0.463	0.509	0.535	0.555	0.553	0.503	11.8
1,1,2-Trichloroethane	0.346	0.348	0.371	0.356	0.341	0.336	0.350	3.6
2-Hexanone	0.349	0.412	0.476	0.448	0.431	0.436	0.425	10.1
Dibromochloromethane	0.305	0.349	0.390	0.384	0.385	0.380	0.366	9
1,2-Dibromoethane	0.311	0.352	0.371	0.356	0.355	0.350	0.349	5.7
Tetrachloroethene	0.315	0.326	0.319	0.324	0.329	0.309	0.320	2.3
Chlorobenzene	0.968	1.054	1.090	1.092	1.100	1.045	1.058	4.7
Ethyl Benzene	1.566	1.794	1.889	1.952	1.972	1.888	1.843	8.1
m/p-Xylenes	0.555	0.672	0.711	0.724	0.715	0.673	0.675	9.3
o-Xylene	0.609	0.689	0.702	0.706	0.707	0.670	0.681	5.5
Styrene	0.879	1.060	1.170	1.181	1.183	1.134	1.101	10.7
Bromoform	0.209	0.234	0.276	0.276	0.300	0.300	0.266	13.9
Isopropylbenzene	3.397	4.034	3.999	4.135	4.006	3.845	3.903	6.8
1,1,2,2-Tetrachloroethane	1.395	1.479	1.513	1.419	1.391	1.396	1.432	3.6
1,3-Dichlorobenzene	1.502	1.652	1.710	1.668	1.675	1.649	1.643	4.4
1,4-Dichlorobenzene	1.605	1.702	1.665	1.687	1.669	1.643	1.662	2.1
1,2-Dichlorobenzene	1.479	1.695	1.735	1.668	1.687	1.622	1.648	5.5
1,2-Dibromo-3-Chloropropane	0.237	0.243	0.285	0.289	0.300	0.320	0.279	11.6
1,2,4-Trichlorobenzene	0.668	0.821	0.978	1.009	1.074	1.080	0.938	17.3
1,2,3-Trichlorobenzene	0.747	0.897	1.032	1.059	1.107	1.096	0.990	14.2
1,2-Dichloroethane-d4		0.836	0.784	0.757	0.783	0.817	0.795	3.9
Dibromofluoromethane		0.329	0.335	0.329	0.340	0.338	0.334	1.5
Toluene-d8		1.237	1.191	1.210	1.219	1.203	1.212	1.4
4-Bromofluorobenzene		0.383	0.393	0.402	0.410	0.421	0.402	3.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 11:31

Lab File ID: VX045431.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.798		26.87	20
Chloromethane	0.770	0.804	0.1	4.42	20
Vinyl Chloride	0.764	0.720		-5.76	20
Bromomethane	0.300	0.346		15.33	20
Chloroethane	0.352	0.394		11.93	20
Trichlorofluoromethane	1.012	1.093		8	20
1,1,2-Trichlorotrifluoroethane	0.581	0.699		20.31	20
1,1-Dichloroethene	0.614	0.651		6.03	20
Acetone	0.369	0.394		6.78	20
Carbon Disulfide	1.636	1.643		0.43	20
Methyl tert-butyl Ether	2.061	2.180		5.77	20
Methyl Acetate	0.888	0.856		-3.6	20
Methylene Chloride	0.731	0.729		-0.27	20
trans-1,2-Dichloroethene	0.607	0.641		5.6	20
1,1-Dichloroethane	1.247	1.311	0.1	5.13	20
Cyclohexane	1.082	1.188		9.8	20
2-Butanone	0.555	0.596		7.39	20
Carbon Tetrachloride	0.465	0.558		20	20
cis-1,2-Dichloroethene	0.749	0.773		3.2	20
Bromochloromethane	0.594	0.584		-1.68	20
Chloroform	1.239	1.317		6.3	20
1,1,1-Trichloroethane	1.000	1.132		13.2	20
Methylcyclohexane	0.549	0.674		22.77	20
Benzene	1.448	1.547		6.84	20
1,2-Dichloroethane	0.521	0.624		19.77	20
Trichloroethene	0.340	0.365		7.35	20
1,2-Dichloropropane	0.372	0.390		4.84	20
Bromodichloromethane	0.516	0.591		14.53	20
4-Methyl-2-Pentanone	0.587	0.681		16.01	20
Toluene	0.849	0.944		11.19	20
t-1,3-Dichloropropene	0.431	0.547		26.91	20
cis-1,3-Dichloropropene	0.503	0.610		21.27	20
1,1,2-Trichloroethane	0.350	0.370		5.71	20
2-Hexanone	0.425	0.519		22.12	20
Dibromochloromethane	0.366	0.425		16.12	20
1,2-Dibromoethane	0.349	0.390		11.75	20
Tetrachloroethene	0.320	0.353		10.31	20
Chlorobenzene	1.058	1.130	0.3	6.8	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 11:31

Lab File ID: VX045431.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.068		12.21	20
m/p-Xylenes	0.675	0.760		12.59	20
o-Xylene	0.681	0.753		10.57	20
Styrene	1.101	1.269		15.26	20
Bromoform	0.266	0.306	0.1	15.04	20
Isopropylbenzene	3.903	4.076		4.43	20
1,1,2,2-Tetrachloroethane	1.432	1.422	0.3	-0.7	20
1,3-Dichlorobenzene	1.643	1.755		6.82	20
1,4-Dichlorobenzene	1.662	1.732		4.15	20
1,2-Dichlorobenzene	1.648	1.733		5.16	20
1,2-Dibromo-3-Chloropropane	0.279	0.323		15.77	20
1,2,4-Trichlorobenzene	0.938	1.047		11.62	20
1,2,3-Trichlorobenzene	0.990	1.048		5.86	20
1,2-Dichloroethane-d4	0.795	0.876		10.19	20
Dibromofluoromethane	0.334	0.368		10.18	20
Toluene-d8	1.212	1.260		3.96	20
4-Bromofluorobenzene	0.402	0.481		19.65	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 21:00

Lab File ID: VX045455.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.819		30.21	50
Chloromethane	0.770	0.884	0.1	14.81	50
Vinyl Chloride	0.764	0.774		1.31	50
Bromomethane	0.300	0.362		20.67	50
Chloroethane	0.352	0.433		23.01	50
Trichlorofluoromethane	1.012	1.154		14.03	50
1,1,2-Trichlorotrifluoroethane	0.581	0.697		19.97	50
1,1-Dichloroethene	0.614	0.684		11.4	50
Acetone	0.369	0.395		7.05	50
Carbon Disulfide	1.636	1.714		4.77	50
Methyl tert-butyl Ether	2.061	2.334		13.25	50
Methyl Acetate	0.888	0.929		4.62	50
Methylene Chloride	0.731	0.785		7.39	50
trans-1,2-Dichloroethene	0.607	0.696		14.66	50
1,1-Dichloroethane	1.247	1.404	0.1	12.59	50
Cyclohexane	1.082	1.254		15.9	50
2-Butanone	0.555	0.613		10.45	50
Carbon Tetrachloride	0.465	0.566		21.72	50
cis-1,2-Dichloroethene	0.749	0.825		10.15	50
Bromochloromethane	0.594	0.654		10.1	50
Chloroform	1.239	1.426		15.09	50
1,1,1-Trichloroethane	1.000	1.219		21.9	50
Methylcyclohexane	0.549	0.662		20.58	50
Benzene	1.448	1.606		10.91	50
1,2-Dichloroethane	0.521	0.655		25.72	50
Trichloroethene	0.340	0.376		10.59	50
1,2-Dichloropropane	0.372	0.410		10.22	50
Bromodichloromethane	0.516	0.612		18.6	50
4-Methyl-2-Pentanone	0.587	0.701		19.42	50
Toluene	0.849	0.978		15.19	50
t-1,3-Dichloropropene	0.431	0.530		22.97	50
cis-1,3-Dichloropropene	0.503	0.608		20.88	50
1,1,2-Trichloroethane	0.350	0.391		11.71	50
2-Hexanone	0.425	0.525		23.53	50
Dibromochloromethane	0.366	0.438		19.67	50
1,2-Dibromoethane	0.349	0.404		15.76	50
Tetrachloroethene	0.320	0.359		12.19	50
Chlorobenzene	1.058	1.173	0.3	10.87	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 21:00

Lab File ID: VX045455.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.140		16.11	50
m/p-Xylenes	0.675	0.775		14.81	50
o-Xylene	0.681	0.774		13.66	50
Styrene	1.101	1.315		19.44	50
Bromoform	0.266	0.310	0.1	16.54	50
Isopropylbenzene	3.903	4.278		9.61	50
1,1,2,2-Tetrachloroethane	1.432	1.479	0.3	3.28	50
1,3-Dichlorobenzene	1.643	1.781		8.4	50
1,4-Dichlorobenzene	1.662	1.787		7.52	50
1,2-Dichlorobenzene	1.648	1.805		9.53	50
1,2-Dibromo-3-Chloropropane	0.279	0.328		17.56	50
1,2,4-Trichlorobenzene	0.938	1.039		10.77	50
1,2,3-Trichlorobenzene	0.990	1.055		6.57	50
1,2-Dichloroethane-d4	0.795	0.954		20	50
Dibromofluoromethane	0.334	0.375		12.27	50
Toluene-d8	1.212	1.279		5.53	50
4-Bromofluorobenzene	0.402	0.501		24.63	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 10:03

Lab File ID: VX045457.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.797		26.71	20
Chloromethane	0.770	0.767	0.1	-0.39	20
Vinyl Chloride	0.764	0.728		-4.71	20
Bromomethane	0.300	0.338		12.67	20
Chloroethane	0.352	0.394		11.93	20
Trichlorofluoromethane	1.012	1.079		6.62	20
1,1,2-Trichlorotrifluoroethane	0.581	0.681		17.21	20
1,1-Dichloroethene	0.614	0.637		3.75	20
Acetone	0.369	0.430		16.53	20
Carbon Disulfide	1.636	1.630		-0.37	20
Methyl tert-butyl Ether	2.061	2.084		1.12	20
Methyl Acetate	0.888	0.831		-6.42	20
Methylene Chloride	0.731	0.719		-1.64	20
trans-1,2-Dichloroethene	0.607	0.640		5.44	20
1,1-Dichloroethane	1.247	1.286	0.1	3.13	20
Cyclohexane	1.082	1.149		6.19	20
2-Butanone	0.555	0.589		6.13	20
Carbon Tetrachloride	0.465	0.551		18.5	20
cis-1,2-Dichloroethene	0.749	0.757		1.07	20
Bromochloromethane	0.594	0.619		4.21	20
Chloroform	1.239	1.300		4.92	20
1,1,1-Trichloroethane	1.000	1.097		9.7	20
Methylcyclohexane	0.549	0.665		21.13	20
Benzene	1.448	1.526		5.39	20
1,2-Dichloroethane	0.521	0.613		17.66	20
Trichloroethene	0.340	0.363		6.76	20
1,2-Dichloropropane	0.372	0.391		5.11	20
Bromodichloromethane	0.516	0.597		15.7	20
4-Methyl-2-Pentanone	0.587	0.648		10.39	20
Toluene	0.849	0.930		9.54	20
t-1,3-Dichloropropene	0.431	0.522		21.11	20
cis-1,3-Dichloropropene	0.503	0.600		19.28	20
1,1,2-Trichloroethane	0.350	0.371		6	20
2-Hexanone	0.425	0.477		12.23	20
Dibromochloromethane	0.366	0.419		14.48	20
1,2-Dibromoethane	0.349	0.386		10.6	20
Tetrachloroethene	0.320	0.355		10.94	20
Chlorobenzene	1.058	1.132	0.3	6.99	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 10:03

Lab File ID: VX045457.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.063		11.94	20
m/p-Xylenes	0.675	0.754		11.7	20
o-Xylene	0.681	0.732		7.49	20
Styrene	1.101	1.239		12.53	20
Bromoform	0.266	0.303	0.1	13.91	20
Isopropylbenzene	3.903	4.277		9.58	20
1,1,2,2-Tetrachloroethane	1.432	1.413	0.3	-1.33	20
1,3-Dichlorobenzene	1.643	1.715		4.38	20
1,4-Dichlorobenzene	1.662	1.707		2.71	20
1,2-Dichlorobenzene	1.648	1.684		2.18	20
1,2-Dibromo-3-Chloropropane	0.279	0.313		12.19	20
1,2,4-Trichlorobenzene	0.938	0.998		6.4	20
1,2,3-Trichlorobenzene	0.990	1.018		2.83	20
1,2-Dichloroethane-d4	0.795	0.847		6.54	20
Dibromofluoromethane	0.334	0.361		8.08	20
Toluene-d8	1.212	1.218		0.5	20
4-Bromofluorobenzene	0.402	0.444		10.45	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 15:35

Lab File ID: VX045470.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.846		34.5	50
Chloromethane	0.770	0.790	0.1	2.6	50
Vinyl Chloride	0.764	0.765		0.13	50
Bromomethane	0.300	0.359		19.67	50
Chloroethane	0.352	0.417		18.47	50
Trichlorofluoromethane	1.012	1.170		15.61	50
1,1,2-Trichlorotrifluoroethane	0.581	0.755		29.95	50
1,1-Dichloroethene	0.614	0.697		13.52	50
Acetone	0.369	0.438		18.7	50
Carbon Disulfide	1.636	1.741		6.42	50
Methyl tert-butyl Ether	2.061	2.308		11.98	50
Methyl Acetate	0.888	0.898		1.13	50
Methylene Chloride	0.731	0.780		6.7	50
trans-1,2-Dichloroethene	0.607	0.694		14.33	50
1,1-Dichloroethane	1.247	1.395	0.1	11.87	50
Cyclohexane	1.082	1.261		16.54	50
2-Butanone	0.555	0.621		11.89	50
Carbon Tetrachloride	0.465	0.601		29.25	50
cis-1,2-Dichloroethene	0.749	0.830		10.81	50
Bromochloromethane	0.594	0.601		1.18	50
Chloroform	1.239	1.418		14.45	50
1,1,1-Trichloroethane	1.000	1.213		21.3	50
Methylcyclohexane	0.549	0.736		34.06	50
Benzene	1.448	1.694		16.99	50
1,2-Dichloroethane	0.521	0.678		30.13	50
Trichloroethene	0.340	0.392		15.29	50
1,2-Dichloropropane	0.372	0.431		15.86	50
Bromodichloromethane	0.516	0.638		23.64	50
4-Methyl-2-Pentanone	0.587	0.694		18.23	50
Toluene	0.849	1.027		20.97	50
t-1,3-Dichloropropene	0.431	0.582		35.03	50
cis-1,3-Dichloropropene	0.503	0.665		32.21	50
1,1,2-Trichloroethane	0.350	0.399		14	50
2-Hexanone	0.425	0.510		20	50
Dibromochloromethane	0.366	0.453		23.77	50
1,2-Dibromoethane	0.349	0.413		18.34	50
Tetrachloroethene	0.320	0.397		24.06	50
Chlorobenzene	1.058	1.242	0.3	17.39	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 15:35

Lab File ID: VX045470.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.270		23.17	50
m/p-Xylenes	0.675	0.834		23.56	50
o-Xylene	0.681	0.807		18.5	50
Styrene	1.101	1.377		25.07	50
Bromoform	0.266	0.322	0.1	21.05	50
Isopropylbenzene	3.903	4.772		22.26	50
1,1,2,2-Tetrachloroethane	1.432	1.521	0.3	6.22	50
1,3-Dichlorobenzene	1.643	1.873		14	50
1,4-Dichlorobenzene	1.662	1.881		13.18	50
1,2-Dichlorobenzene	1.648	1.816		10.19	50
1,2-Dibromo-3-Chloropropane	0.279	0.329		17.92	50
1,2,4-Trichlorobenzene	0.938	1.092		16.42	50
1,2,3-Trichlorobenzene	0.990	1.095		10.61	50
1,2-Dichloroethane-d4	0.795	0.782		-1.63	50
Dibromofluoromethane	0.334	0.335		0.3	50
Toluene-d8	1.212	1.120		-7.59	50
4-Bromofluorobenzene	0.402	0.415		3.23	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q1629	OrderDate:	3/21/2025 1:55:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	I41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1629-01	TT174I1-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/29/25	03/21/25
Q1629-02	TT162S1-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/29/25	03/21/25
Q1629-03	TT188S1-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/28/25	03/21/25
Q1629-07	RW09-MW01D1-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/29/25	03/21/25
Q1629-08	RW09-MW01D2-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/29/25	03/21/25
Q1629-09	RW09-MW01D3-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/29/25	03/21/25
Q1629-10	TT150S1-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/29/25	03/21/25
Q1629-11	FB03-20250321	Water	SVOC-SIMGroup1	8270-Modified	03/21/25	03/25/25	03/29/25	03/21/25
Q1629-14	RW09-MW01S-20250320	Water	SVOC-SIMGroup1	8270-Modified	03/20/25	03/25/25	03/31/25	03/21/25

Hit Summary Sheet SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Sample ID	Client ID		Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : TT174H1-20250320									
Q1629-01	TT174H1-20250320	WATER	1,4-Dioxane	0.700		0.07	0.2	0.2	ug/L
			Total Svoc :			0.70			
			Total Concentration:			0.70			
Client ID : TT162S1-20250320									
Q1629-02	TT162S1-20250320	WATER	1,4-Dioxane	0.240		0.07	0.2	0.2	ug/L
			Total Svoc :			0.24			
			Total Concentration:			0.24			
Client ID : TT188S1-20250320									
Q1629-03	TT188S1-20250320	WATER	1,4-Dioxane	0.150	J	0.07	0.2	0.2	ug/L
			Total Svoc :			0.15			
			Total Concentration:			0.15			
Client ID : RW09-MW01D1-20250320									
Q1629-07	RW09-MW01D1-202503	WATER	1,4-Dioxane	0.350		0.07	0.2	0.2	ug/L
			Total Svoc :			0.35			
			Total Concentration:			0.35			
Client ID : TT150S1-20250320									
Q1629-10	TT150S1-20250320	WATER	1,4-Dioxane	0.240		0.07	0.2	0.2	ug/L
			Total Svoc :			0.24			
			Total Concentration:			0.24			
Client ID : RW09-MW01S-20250320									
Q1629-14	RW09-MW01S-20250320	WATER	1,4-Dioxane	0.400		0.07	0.2	0.2	ug/L
			Total Svoc :			0.40			
			Total Concentration:			0.40			



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	TT174I1-20250320	SDG No.:	Q1629
Lab Sample ID:	Q1629-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036797.D	1	03/25/25 08:41	03/29/25 06:19	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.70		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		79%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		74%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.31		53 - 106		79%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.41		58 - 132		102%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1890	7.696				
1146-65-2	Naphthalene-d8	4300	10.487				
15067-26-2	Acenaphthene-d10	2640	14.334				
1517-22-2	Phenanthrene-d10	5200	17.086				
1719-03-5	Chrysene-d12	3940	21.277				
1520-96-3	Perylene-d12	3340	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	TT162S1-20250320	SDG No.:	Q1629
Lab Sample ID:	Q1629-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036798.D	1	03/25/25 08:41	03/29/25 06:56	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.24		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		77%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		106%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.26		55 - 111		66%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.29		53 - 106		72%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		106%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1620	7.695				
1146-65-2	Naphthalene-d8	3710	10.487				
15067-26-2	Acenaphthene-d10	2380	14.334				
1517-22-2	Phenanthrene-d10	5160	17.086				
1719-03-5	Chrysene-d12	4060	21.277				
1520-96-3	Perylene-d12	3120	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	TT188S1-20250320	SDG No.:	Q1629
Lab Sample ID:	Q1629-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036768.D	1	03/25/25 08:41	03/28/25 11:34	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.15	J	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.44		30 - 150		110%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		74%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.32		53 - 106		81%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		94%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1850		7.696			
1146-65-2	Naphthalene-d8	4440		10.488			
15067-26-2	Acenaphthene-d10	2730		14.334			
1517-22-2	Phenanthrene-d10	6130		17.087			
1719-03-5	Chrysene-d12	5470		21.277			
1520-96-3	Perylene-d12	4750		23.519			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	RW09-MW01D1-20250320	SDG No.:	Q1629
Lab Sample ID:	Q1629-07	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036803.D	1	03/25/25 08:41	03/29/25 09:57	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		86%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		107%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.29		55 - 111		72%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		111%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1720	7.696				
1146-65-2	Naphthalene-d8	4160	10.487				
15067-26-2	Acenaphthene-d10	2580	14.334				
1517-22-2	Phenanthrene-d10	5280	17.086				
1719-03-5	Chrysene-d12	3900	21.277				
1520-96-3	Perylene-d12	3270	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	RW09-MW01D2-20250320	SDG No.:	Q1629
Lab Sample ID:	Q1629-08	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036799.D	1	03/25/25 08:41	03/29/25 07:32	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.39		30 - 150		97%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.26		55 - 111		64%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.25		53 - 106		63%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		94%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1590	7.696				
1146-65-2	Naphthalene-d8	3760	10.488				
15067-26-2	Acenaphthene-d10	2610	14.334				
1517-22-2	Phenanthrene-d10	5010	17.087				
1719-03-5	Chrysene-d12	3750	21.277				
1520-96-3	Perylene-d12	3120	23.519				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	RW09-MW01D3-20250320	SDG No.:	Q1629
Lab Sample ID:	Q1629-09	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036800.D	1	03/25/25 08:41	03/29/25 08:08	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.37		30 - 150		92%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		102%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		89%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1640	7.695				
1146-65-2	Naphthalene-d8	3880	10.477				
15067-26-2	Acenaphthene-d10	2530	14.334				
1517-22-2	Phenanthrene-d10	5190	17.086				
1719-03-5	Chrysene-d12	3680	21.277				
1520-96-3	Perylene-d12	3000	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25
Client Sample ID:	TT150S1-20250320		SDG No.:	Q1629
Lab Sample ID:	Q1629-10		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036801.D	1	03/25/25 08:41	03/29/25 08:44	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.24		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.34		30 - 150		84%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		102%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.29		55 - 111		72%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.34		53 - 106		84%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1640	7.696				
1146-65-2	Naphthalene-d8	3930	10.477				
15067-26-2	Acenaphthene-d10	2500	14.334				
1517-22-2	Phenanthrene-d10	5310	17.087				
1719-03-5	Chrysene-d12	3780	21.277				
1520-96-3	Perylene-d12	3110	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25
Client Sample ID:	FB03-20250321		SDG No.:	Q1629
Lab Sample ID:	Q1629-11		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036802.D	1	03/25/25 08:41	03/29/25 09:20	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		86%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		78%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		82%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1680	7.695				
1146-65-2	Naphthalene-d8	4130	10.477				
15067-26-2	Acenaphthene-d10	2700	14.334				
1517-22-2	Phenanthrene-d10	5610	17.086				
1719-03-5	Chrysene-d12	4080	21.277				
1520-96-3	Perylene-d12	3140	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629
Lab Sample ID:	Q1629-14		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036811.D	1	03/25/25 08:41	03/31/25 11:51	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.40		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.44		30 - 150		109%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.29		55 - 111		72%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		82%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		105%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1750	7.696				
1146-65-2	Naphthalene-d8	4430	10.477				
15067-26-2	Acenaphthene-d10	2710	14.334				
1517-22-2	Phenanthrene-d10	5740	17.087				
1719-03-5	Chrysene-d12	4630	21.277				
1520-96-3	Perylene-d12	4170	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167295BL	PB167295BL	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.37	94		30	150
		Nitrobenzene-d5	0.4	0.28	70		55	111
		2-Fluorobiphenyl	0.4	0.27	67		53	106
		Terphenyl-d14	0.4	0.36	90		58	132
PB167295BS	PB167295BS	2-Methylnaphthalene-d10	0.4	0.37	93		30	150
		Fluoranthene-d10	0.4	0.33	83		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.33	82		53	106
		Terphenyl-d14	0.4	0.37	91		58	132
Q1629-01	TT174I1-20250320	2-Methylnaphthalene-d10	0.4	0.32	79		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.30	74		55	111
		2-Fluorobiphenyl	0.4	0.31	79		53	106
		Terphenyl-d14	0.4	0.41	102		58	132
Q1629-02	TT162S1-20250320	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.42	106		30	150
		Nitrobenzene-d5	0.4	0.26	66		55	111
		2-Fluorobiphenyl	0.4	0.29	72		53	106
		Terphenyl-d14	0.4	0.42	106		58	132
Q1629-03	TT188S1-20250320	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.44	110		30	150
		Nitrobenzene-d5	0.4	0.30	74		55	111
		2-Fluorobiphenyl	0.4	0.32	81		53	106
		Terphenyl-d14	0.4	0.38	94		58	132
Q1629-04MS	TT188S1-20250320MS	2-Methylnaphthalene-d10	0.4	0.38	95		30	150
		Fluoranthene-d10	0.4	0.42	105		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.35	87		53	106
		Terphenyl-d14	0.4	0.40	99		58	132
Q1629-05MSD	TT188S1-20250320MSD	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.43	107		30	150
		Nitrobenzene-d5	0.4	0.31	77		55	111
		2-Fluorobiphenyl	0.4	0.36	91		53	106
		Terphenyl-d14	0.4	0.37	93		58	132
Q1629-07	RW09-MW01D1-20250320	2-Methylnaphthalene-d10	0.4	0.34	86		30	150
		Fluoranthene-d10	0.4	0.43	107		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.45	111		58	132
Q1629-08	RW09-MW01D2-20250320	2-Methylnaphthalene-d10	0.4	0.30	75		30	150
		Fluoranthene-d10	0.4	0.39	97		30	150
		Nitrobenzene-d5	0.4	0.26	64		55	111
		2-Fluorobiphenyl	0.4	0.25	63		53	106
		Terphenyl-d14	0.4	0.38	94		58	132
Q1629-09	RW09-MW01D3-20250320	2-Methylnaphthalene-d10	0.4	0.37	92		30	150
		Fluoranthene-d10	0.4	0.41	102		30	150
		Nitrobenzene-d5	0.4	0.32	79		55	111
		2-Fluorobiphenyl	0.4	0.36	89		53	106
		Terphenyl-d14	0.4	0.43	108		58	132
Q1629-10	TT150S1-20250320	2-Methylnaphthalene-d10	0.4	0.34	84		30	150

Surrogate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q1629-10	TT150S1-20250320	Fluoranthene-d10	0.4	0.41	102		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
		2-Fluorobiphenyl	0.4	0.34	84		53	106
		Terphenyl-d14	0.4	0.43	108		58	132
Q1629-11	FB03-20250321	2-Methylnaphthalene-d10	0.4	0.35	86		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.31	78		55	111
		2-Fluorobiphenyl	0.4	0.33	82		53	106
Q1629-14	RW09-MW01S-20250320	Terphenyl-d14	0.4	0.43	108		58	132
		2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.44	109		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
Q1629-15MS	RW09-MW01S-20250320MS	2-Fluorobiphenyl	0.4	0.33	82		53	106
		Terphenyl-d14	0.4	0.42	105		58	132
		2-Methylnaphthalene-d10	0.4	0.30	75		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
Q1629-16MSD	RW09-MW01S-20250320MSD	Nitrobenzene-d5	0.4	0.26	64		55	111
		2-Fluorobiphenyl	0.4	0.30	75		53	106
		Terphenyl-d14	0.4	0.35	86		58	132
		2-Methylnaphthalene-d10	0.4	0.33	81		30	150
		Fluoranthene-d10	0.4	0.38	95		30	150
		Nitrobenzene-d5	0.4	0.26	66		55	111
		2-Fluorobiphenyl	0.4	0.29	72		53	106
		Terphenyl-d14	0.4	0.36	89		58	132

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: Q1629
Client: AECOM Technical Services, Inc.
Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1629-04MS	Client Sample ID:	TT188S1-20250320MS					DataFile:	BN036769.D		
1,4-Dioxane	0.4	0.15	0.63	ug/L	120				70	130	

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: Q1629
Client: AECOM Technical Services, Inc.
Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1629-05MSD	Client Sample ID:	TT188S1-20250320MSD					DataFile:	BN036770.D		
1,4-Dioxane	0.4	0.15	0.56	ug/L	103		15		70	130	20

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: Q1629
Client: AECOM Technical Services, Inc.
Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1629-15MS	Client Sample ID:	RW09-MW01S-20250320MS					DataFile:	BN036812.D		
1,4-Dioxane	0.4	0.40	0.82	ug/L	105				70	130	

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: Q1629
Client: AECOM Technical Services, Inc.
Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1629-16MSD	Client Sample ID:	RW09-MW01S-20250320MSD					DataFile:	BN036813.D		
1,4-Dioxane	0.4	0.40	0.85	ug/L	113		7		70	130	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: 8270-Modified DataFile: BN036804.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167295BS	1,4-Dioxane	0.4	0.32	ug/L	80				70	130	

A

B

C

D

E

F

G

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167295BL

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1629

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: BN036792.D

Lab Sample ID: PB167295BL

Instrument ID: BNA_N

Date Extracted: 03/25/2025

Matrix: (soil/water) Water

Date Analyzed: 03/29/2025

Level: (low/med) LOW

Time Analyzed: 03:20

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
TT174I1-20250320	Q1629-01	BN036797.D	03/29/2025
PB167295BS	PB167295BS	BN036804.D	03/29/2025
TT162S1-20250320	Q1629-02	BN036798.D	03/29/2025
RW09-MW01D2-20250320	Q1629-08	BN036799.D	03/29/2025
RW09-MW01D3-20250320	Q1629-09	BN036800.D	03/29/2025
TT150S1-20250320	Q1629-10	BN036801.D	03/29/2025
FB03-20250321	Q1629-11	BN036802.D	03/29/2025
RW09-MW01D1-20250320	Q1629-07	BN036803.D	03/29/2025
RW09-MW01S-20250320	Q1629-14	BN036811.D	03/31/2025
RW09-MW01S-20250320MS	Q1629-15MS	BN036812.D	03/31/2025
RW09-MW01S-20250320MSD	Q1629-16MSD	BN036813.D	03/31/2025
TT188S1-20250320	Q1629-03	BN036768.D	03/28/2025
TT188S1-20250320MS	Q1629-04MS	BN036769.D	03/28/2025
TT188S1-20250320MSD	Q1629-05MSD	BN036770.D	03/28/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: BN036556.D

DFTPP Injection Date: 03/10/2025

Instrument ID: BNA_N

DFTPP Injection Time: 11:03

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	58.6
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	52.3
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	50.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.8
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	9.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.9 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036557.D	03/10/2025	11:42
SSTDICC0.2	SSTDICC0.2	BN036558.D	03/10/2025	12:18
SSTDICCC0.4	SSTDICCC0.4	BN036559.D	03/10/2025	12:54
SSTDICC0.8	SSTDICC0.8	BN036560.D	03/10/2025	13:31
SSTDICC1.6	SSTDICC1.6	BN036561.D	03/10/2025	14:07
SSTDICC3.2	SSTDICC3.2	BN036562.D	03/10/2025	14:43
SSTDICC5.0	SSTDICC5.0	BN036563.D	03/10/2025	15:19

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: BN036756.D

DFTPP Injection Date: 03/28/2025

Instrument ID: BNA_N

DFTPP Injection Time: 03:07

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	67
68	Less than 2.0% of mass 69	0.9 (1.5) 1
69	Mass 69 relative abundance	57.5
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	53.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.1
275	10.0 - 60.0% of mass 198	24.9
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	8.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.3 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036757.D	03/28/2025	03:46
TT188S1-20250320	Q1629-03	BN036768.D	03/28/2025	11:34
TT188S1-20250320MS	Q1629-04MS	BN036769.D	03/28/2025	12:47
TT188S1-20250320MSD	Q1629-05MSD	BN036770.D	03/28/2025	13:22
SSTDCCC0.4EC	SSTDCCC0.4	BN036771.D	03/28/2025	13:58

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: BN036790.D

DFTPP Injection Date: 03/29/2025

Instrument ID: BNA_N

DFTPP Injection Time: 01:26

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	71.7
68	Less than 2.0% of mass 69	1.1 (1.9) 1
69	Mass 69 relative abundance	58.5
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	52.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	22.7
365	Greater than 1% of mass 198	3.9
441	Present, but less than mass 443	8.2
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.1 (20.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036791.D	03/29/2025	02:44
PB167295BL	PB167295BL	BN036792.D	03/29/2025	03:20
TT174I1-20250320	Q1629-01	BN036797.D	03/29/2025	06:19
TT162S1-20250320	Q1629-02	BN036798.D	03/29/2025	06:56
RW09-MW01D2-20250320	Q1629-08	BN036799.D	03/29/2025	07:32
RW09-MW01D3-20250320	Q1629-09	BN036800.D	03/29/2025	08:08
TT150S1-20250320	Q1629-10	BN036801.D	03/29/2025	08:44
FB03-20250321	Q1629-11	BN036802.D	03/29/2025	09:20
RW09-MW01D1-20250320	Q1629-07	BN036803.D	03/29/2025	09:57
PB167295BS	PB167295BS	BN036804.D	03/29/2025	10:33
SSTDCCC0.4EC	SSTDCCC0.4	BN036807.D	03/29/2025	12:21

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: BN036808.D

DFTPP Injection Date: 03/31/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	70.2
68	Less than 2.0% of mass 69	1 (1.8) 1
69	Mass 69 relative abundance	58.4
70	Less than 2.0% of mass 69	0.4 (0.6) 1
127	10.0 - 80.0% of mass 198	54.1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.2
275	10.0 - 60.0% of mass 198	23.4
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	8.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10 (21.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036809.D	03/31/2025	10:39
RW09-MW01S-20250320	Q1629-14	BN036811.D	03/31/2025	11:51
RW09-MW01S-20250320MS	Q1629-15MS	BN036812.D	03/31/2025	13:04
RW09-MW01S-20250320MSD	Q1629-16MSD	BN036813.D	03/31/2025	13:40
SSTDCCC0.4EC	SSTDCCC0.4	BN036815.D	03/31/2025	15:08

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/28/2025

Lab File ID: BN036757.D Time Analyzed: 03:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1974	7.703	4607	10.49	2750	14.33
UPPER LIMIT	3948	8.203	9214	10.988	5500	14.834
LOWER LIMIT	987	7.203	2303.5	9.988	1375	13.834
EPA SAMPLE NO.						
01 TT188S1-20250320	1850	7.70	4438	10.49	2727	14.33
02 TT188S1-20250320MS	1547	7.70	3800	10.49	2504	14.33
03 TT188S1-20250320MSD	1415	7.70	3434	10.49	2160	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4
IS2 (NPT) = Naphthalene-d8
IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/28/2025

Lab File ID: BN036757.D Time Analyzed: 03:46

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5937	17.087	5680	21.277	5140	23.519
UPPER LIMIT	11874	17.587	11360	21.777	10280	24.019
LOWER LIMIT	2968.5	16.587	2840	20.777	2570	23.019
EPA SAMPLE NO.						
01 TT188S1-20250320	6125	17.09	5473	21.28	4747	23.52
02 TT188S1-20250320MS	5519	17.09	4772	21.28	4206	23.52
03 TT188S1-20250320MSD	4889	17.09	4383	21.28	3881	23.52

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/29/2025

Lab File ID: BN036791.D Time Analyzed: 02:44

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1826	7.696	4490	10.48	2701	14.33
UPPER LIMIT	3652	8.196	8980	10.977	5402	14.834
LOWER LIMIT	913	7.196	2245	9.977	1350.5	13.834
EPA SAMPLE NO.						
01 PB167295BL	1921	7.70	4330	10.50	2411	14.35
02 PB167295BS	1794	7.70	4263	10.48	2364	14.33
03 TT174I1-20250320	1885	7.70	4295	10.49	2635	14.33
04 TT162S1-20250320	1621	7.70	3708	10.49	2378	14.33
05 RW09-MW01D1-20250320	1719	7.70	4159	10.49	2581	14.33
06 RW09-MW01D2-20250320	1587	7.70	3759	10.49	2611	14.33
07 RW09-MW01D3-20250320	1638	7.70	3883	10.48	2532	14.33
08 TT150S1-20250320	1638	7.70	3931	10.48	2498	14.33
09 FB03-20250321	1678	7.70	4131	10.48	2695	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/29/2025

Lab File ID: BN036791.D Time Analyzed: 02:44

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	5571	17.087	4634	21.268	4032	23.516
	UPPER LIMIT	11142	17.587	9268	21.768	8064	24.016
	LOWER LIMIT	2785.5	16.587	2317	20.768	2016	23.016
	EPA SAMPLE NO.						
01	PB167295BL	4533	17.10	3414	21.28	1734 *	23.53
02	PB167295BS	4802	17.09	3163	21.28	2660	23.52
03	TT174I1-20250320	5199	17.09	3937	21.28	3340	23.52
04	TT162S1-20250320	5161	17.09	4055	21.28	3123	23.52
05	RW09-MW01D1-20250320	5278	17.09	3898	21.28	3267	23.52
06	RW09-MW01D2-20250320	5014	17.09	3745	21.28	3124	23.52
07	RW09-MW01D3-20250320	5189	17.09	3677	21.28	2995	23.52
08	TT150S1-20250320	5311	17.09	3784	21.28	3106	23.52
09	FB03-20250321	5610	17.09	4075	21.28	3144	23.52

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/31/2025

Lab File ID: BN036809.D Time Analyzed: 10:39

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1888	7.695	4656	10.48	2798	14.33
UPPER LIMIT	3776	8.195	9312	10.977	5596	14.834
LOWER LIMIT	944	7.195	2328	9.977	1399	13.834
EPA SAMPLE NO.						
01 RW09-MW01S-20250320	1749	7.70	4433	10.48	2707	14.33
02 RW09-MW01S-20250320MS	1333	7.70	3452	10.48	2071	14.33
03 RW09-MW01S-20250320MSD	1372	7.70	3583	10.48	2162	14.33

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 03/31/2025

Lab File ID: BN036809.D Time Analyzed: 10:39

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	5809	17.086	4586	21.277	3945	23.522
UPPER LIMIT	11618	17.586	9172	21.777	7890	24.022
LOWER LIMIT	2904.5	16.586	2293	20.777	1972.5	23.022
EPA SAMPLE NO.						
01 RW09-MW01S-20250320	5735	17.09	4630	21.28	4169	23.52
02 RW09-MW01S-20250320MS	4565	17.09	3721	21.28	3243	23.51
03 RW09-MW01S-20250320MSD	4671	17.09	3683	21.28	3184	23.52

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB167295BL		SDG No.:	Q1629
Lab Sample ID:	PB167295BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036792.D	1	03/25/25 08:41	03/29/25 03:20	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		77%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		94%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		70%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.27		53 - 106		67%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		90%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1920	7.703				
1146-65-2	Naphthalene-d8	4330	10.498				
15067-26-2	Acenaphthene-d10	2410	14.345				
1517-22-2	Phenanthrene-d10	4530	17.099				
1719-03-5	Chrysene-d12	3410	21.277				
1520-96-3	Perylene-d12	1730	23.525				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	PB167295BS		SDG No.:	Q1629
Lab Sample ID:	PB167295BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036804.D	1	03/25/25 08:41	03/29/25 10:33	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.32		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.37		30 - 150		93%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.33		30 - 150		83%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		77%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		82%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.37		58 - 132		91%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1790	7.696				
1146-65-2	Naphthalene-d8	4260	10.477				
15067-26-2	Acenaphthene-d10	2360	14.334				
1517-22-2	Phenanthrene-d10	4800	17.087				
1719-03-5	Chrysene-d12	3160	21.277				
1520-96-3	Perylene-d12	2660	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	TT188S1-20250320MS	SDG No.:	Q1629
Lab Sample ID:	Q1629-04MS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036769.D	1	03/25/25 08:41	03/28/25 12:47	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.63		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.38		30 - 150		95%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		105%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		77%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		87%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		99%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1550	7.696				
1146-65-2	Naphthalene-d8	3800	10.488				
15067-26-2	Acenaphthene-d10	2500	14.334				
1517-22-2	Phenanthrene-d10	5520	17.087				
1719-03-5	Chrysene-d12	4770	21.277				
1520-96-3	Perylene-d12	4210	23.519				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	RW09-MW01S-20250320MS	SDG No.:	Q1629
Lab Sample ID:	Q1629-15MS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036812.D	1	03/25/25 08:41	03/31/25 13:04	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.82		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.30		30 - 150		75%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		95%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.26		55 - 111		64%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.30		53 - 106		75%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.35		58 - 132		86%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1330	7.696				
1146-65-2	Naphthalene-d8	3450	10.477				
15067-26-2	Acenaphthene-d10	2070	14.334				
1517-22-2	Phenanthrene-d10	4570	17.087				
1719-03-5	Chrysene-d12	3720	21.277				
1520-96-3	Perylene-d12	3240	23.514				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	TT188S1-20250320MSD	SDG No.:	Q1629
Lab Sample ID:	Q1629-05MSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036770.D	1	03/25/25 08:41	03/28/25 13:22	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.56		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		107%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		77%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.36		53 - 106		91%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.37		58 - 132		93%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1420	7.696				
1146-65-2	Naphthalene-d8	3430	10.488				
15067-26-2	Acenaphthene-d10	2160	14.334				
1517-22-2	Phenanthrene-d10	4890	17.087				
1719-03-5	Chrysene-d12	4380	21.277				
1520-96-3	Perylene-d12	3880	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.	Date Collected:	03/20/25
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258	Date Received:	03/21/25
Client Sample ID:	RW09-MW01S-20250320MSD	SDG No.:	Q1629
Lab Sample ID:	Q1629-16MSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	990 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036813.D	1	03/25/25 08:41	03/31/25 13:40	PB167295

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.85		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.33		30 - 150		81%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		95%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.26		55 - 111		66%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.29		53 - 106		72%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		89%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1370	7.695				
1146-65-2	Naphthalene-d8	3580	10.477				
15067-26-2	Acenaphthene-d10	2160	14.334				
1517-22-2	Phenanthrene-d10	4670	17.086				
1719-03-5	Chrysene-d12	3680	21.277				
1520-96-3	Perylene-d12	3180	23.516				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN031025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Mar 10 16:06:28 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036557.D 0.2 =BN036558.D 0.4 =BN036559.D 0.8 =BN036560.D 1.6 =BN036561.D 3.2 =BN036562.D 5.0 =BN036563.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.434	0.439	0.498	0.451	0.440	0.445	0.399	0.444	6.60
3)	n-Nitrosodimet...	1.112	0.874	0.935	0.841	0.850	0.883	0.789	0.898	11.64
4) S	2-Fluorophenol	0.931	0.908	0.987	0.878	0.914	0.996	0.911	0.932	4.67
5) S	Phenol-d6	1.243	1.057	1.128	1.067	1.133	1.254	1.180	1.152	6.79
6)	bis(2-Chloroet...	1.426	1.150	1.183	1.129	1.132	1.210	1.104	1.190	9.22
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.572	0.396	0.415	0.401	0.402	0.450	0.411	0.435	14.47
9)	Naphthalene	1.371	1.125	1.206	1.111	1.108	1.222	1.094	1.177	8.45
10)	Hexachlorobuta...	0.296	0.283	0.294	0.267	0.261	0.286	0.251	0.277	6.24
11) SURR2	Methylnaphth...	0.656	0.549	0.606	0.562	0.577	0.633	0.581	0.595	6.55
12)	2-Methylnaphth...	0.810	0.696	0.765	0.703	0.734	0.802	0.731	0.749	6.03
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.181	0.160	0.187	0.169	0.188	0.197	0.188	0.182	7.04
15) S	2-Fluorobiphenyl	2.208	1.982	2.398	2.350	2.364	2.566	2.419	2.327	7.96
16)	Acenaphthylene	1.882	1.756	1.938	1.794	1.834	2.074	1.935	1.888	5.66
17)	Acenaphthene	1.257	1.159	1.281	1.171	1.199	1.339	1.243	1.236	5.17
18)	Fluorene	1.629	1.600	1.764	1.609	1.670	1.778	1.650	1.672	4.32
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.057	0.077	0.075	0.088	0.110	0.111	0.086		24.66
21)	4-Bromophenyl-...	0.243	0.227	0.274	0.238	0.241	0.278	0.253	0.251	7.53
22)	Hexachlorobenzene	0.306	0.288	0.336	0.295	0.283	0.322	0.289	0.303	6.58
23)	Atrazine	0.193	0.191	0.213	0.192	0.200	0.216	0.200	0.201	5.08
24)	Pentachlorophenol	0.140	0.116	0.137	0.122	0.135	0.161	0.155	0.138	11.76
25)	Phenanthrene	1.190	1.111	1.297	1.141	1.165	1.300	1.195	1.200	6.09
26)	Anthracene	1.026	0.971	1.147	1.033	1.075	1.215	1.112	1.083	7.60
27) SURR	Fluoranthene-d10	1.037	0.955	1.116	0.956	1.025	1.087	1.000	1.025	5.98
28)	Fluoranthene	1.341	1.243	1.452	1.272	1.364	1.447	1.316	1.348	5.95
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.945	2.005	2.131	1.910	1.870	1.992	1.837	1.956	5.04
31) S	Terphenyl-d14	0.962	0.965	1.028	0.924	0.915	0.987	0.926	0.958	4.23
32)	Benzo(a)anthra...	1.389	1.315	1.437	1.304	1.347	1.528	1.415	1.391	5.63
33)	Chrysene	1.486	1.509	1.610	1.507	1.462	1.616	1.448	1.520	4.44
34)	Bis(2-ethylhex...	1.196	1.100	1.044	0.865	0.946	0.912	0.870	0.990	12.74
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN031025.M

36)	Indeno(1,2,3-c...	1.160	1.316	1.546	1.404	1.417	1.693	1.571	1.444	12.27
37)	Benzo(b)fluora...	1.311	1.360	1.547	1.402	1.477	1.595	1.498	1.456	7.04
38)	Benzo(k)fluora...	1.504	1.397	1.620	1.481	1.521	1.635	1.534	1.527	5.34
39) C	Benzo(a)pyrene	1.090	1.152	1.303	1.195	1.223	1.350	1.268	1.226	7.29
40)	Dibenzo(a,h)an...	0.893	0.981	1.163	1.126	1.102	1.351	1.252	1.124	13.76
41)	Benzo(g,h,i)pe...	1.138	1.213	1.382	1.250	1.233	1.449	1.334	1.286	8.36

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: BNA_N Calibration Date/Time: 03/28/2025 03:46

Lab File ID: BN036757.D Init. Calib. Date(s): 03/10/2025 03/10/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:42 15:19

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.509		-14.5	20.0
Fluoranthene-d10	1.025	1.053		2.7	20.0
2-Fluorophenol	0.932	0.782		-16.1	20.0
Phenol-d6	1.152	0.986		-14.4	20.0
Nitrobenzene-d5	0.435	0.370		-14.9	20.0
2-Fluorobiphenyl	2.327	2.013		-13.5	20.0
2,4,6-Tribromophenol	0.182	0.184		1.1	20.0
Terphenyl-d14	0.958	0.749		-21.8	20.0
1,4-Dioxane	0.444	0.442		-0.4	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629
 Instrument ID: BNA_N Calibration Date/Time: 03/28/2025 13:58
 Lab File ID: BN036771.D Init. Calib. Date(s): 03/10/2025 03/10/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:42 15:19
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.540		-9.2	50.0
Fluoranthene-d10	1.025	1.012		-1.3	50.0
2-Fluorophenol	0.932	0.846		-9.2	50.0
Phenol-d6	1.152	1.010		-12.3	50.0
Nitrobenzene-d5	0.435	0.354		-18.6	50.0
2-Fluorobiphenyl	2.327	2.021		-13.1	50.0
2,4,6-Tribromophenol	0.182	0.177		-2.7	50.0
Terphenyl-d14	0.958	0.816		-14.8	50.0
1,4-Dioxane	0.444	0.460		3.6	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629
 Instrument ID: BNA_N Calibration Date/Time: 03/29/2025 02:44
 Lab File ID: BN036791.D Init. Calib. Date(s): 03/10/2025 03/10/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:42 15:19
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.526		-11.6	20.0
Fluoranthene-d10	1.025	1.032		0.7	20.0
2-Fluorophenol	0.932	0.833		-10.6	20.0
Phenol-d6	1.152	1.007		-12.6	20.0
Nitrobenzene-d5	0.435	0.376		-13.6	20.0
2-Fluorobiphenyl	2.327	2.041		-12.3	20.0
2,4,6-Tribromophenol	0.182	0.178		-2.2	20.0
Terphenyl-d14	0.958	0.782		-18.4	20.0
1,4-Dioxane	0.444	0.453		2.0	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: BNA_N Calibration Date/Time: 03/29/2025 12:21

Lab File ID: BN036807.D Init. Calib. Date(s): 03/10/2025 03/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:42 15:19

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.526		-11.6	50.0
Fluoranthene-d10	1.025	0.985		-3.9	50.0
2-Fluorophenol	0.932	0.835		-10.4	50.0
Phenol-d6	1.152	0.978		-15.1	50.0
Nitrobenzene-d5	0.435	0.364		-16.3	50.0
2-Fluorobiphenyl	2.327	1.915		-17.7	50.0
2,4,6-Tribromophenol	0.182	0.169		-7.1	50.0
Terphenyl-d14	0.958	0.875		-8.7	50.0
1,4-Dioxane	0.444	0.456		2.7	50.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: BNA_N Calibration Date/Time: 03/31/2025 10:39

Lab File ID: BN036809.D Init. Calib. Date(s): 03/10/2025 03/10/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:42 15:19

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.537		-9.7	20.0
Fluoranthene-d10	1.025	0.990		-3.4	20.0
2-Fluorophenol	0.932	0.818		-12.2	20.0
Phenol-d6	1.152	0.953		-17.3	20.0
Nitrobenzene-d5	0.435	0.363		-16.6	20.0
2-Fluorobiphenyl	2.327	1.979		-15.0	20.0
2,4,6-Tribromophenol	0.182	0.161		-11.5	20.0
Terphenyl-d14	0.958	0.798		-16.7	20.0
1,4-Dioxane	0.444	0.454		2.3	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: BNA_N Calibration Date/Time: 03/31/2025 15:08

Lab File ID: BN036815.D Init. Calib. Date(s): 03/10/2025 03/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:42 15:19

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.595	0.543		-8.7	50.0
Fluoranthene-d10	1.025	0.994		-3.0	50.0
2-Fluorophenol	0.932	0.839		-10.0	50.0
Phenol-d6	1.152	0.976		-15.3	50.0
Nitrobenzene-d5	0.435	0.348		-20.0	50.0
2-Fluorobiphenyl	2.327	1.957		-15.9	50.0
2,4,6-Tribromophenol	0.182	0.159		-12.6	50.0
Terphenyl-d14	0.958	0.842		-12.1	50.0
1,4-Dioxane	0.444	0.447		0.7	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: RESOLUTION

ADDRESS:

CITY: STATE: ZIP:

ATTENTION: ELEANOR VIVAUDOU

PHONE: FAX:

PROJECT NAME: NWIRP BETHPAGE

PROJECT NO.: 10731872 LOCATION: BETHPAGE

PROJECT MANAGER: E. VIVAUDOU

e-mail: ELEANOR.VIVAUDOU@CARBON.COM

PHONE: FAX:

BILL TO: E. VIVAUDOU PO#:

ADDRESS: 27 ELLIS PL

CITY: OSSINING STATE: NY ZIP: 10563

ATTENTION: E. VIVAUDOU PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) _____ DAYS*

HARDCOPY (DATA PACKAGE): _____ DAYS*

EDD: STANDARD _____ DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B

+ Raw Data ☐ Other

☒ EDD FORMAT MSDC EDD CARB

1. 2. 3. 4. 5. 6. 7. 8. 9.

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	← Specify Preservatives												
			COMP	GRAB	DATE	TIME		A/E		E										
								1	2											
1.	TT174I1-20250320	6W		X	3-20-25	0920	3	2	1											
2.	TT162S1-20250320	6W		X		1135	3	2	1											
3.	TT188S1-20250320	6W		X		1120	9	6	3											MS/MSD
4.	RW09-MW01S-20250320	6W		X		1120	9	6	3											
5.	RW09-MW01D1-20250320	6W		X		1135	3	2	1											
6.	RW09-MW01D2-20250320	6W		X		1005	3	2	1											
7.	RW09-MW01D3-20250320	6W		X		1003	3	2	1											
8.	TT150S1-20250320	6W		X		1315	3	2	1											
9.	FB03-20250321	DW		X	3-21-25		3	2	1											
10.	TB	DW		X	3-14-25	-	2	2												

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP _____ °C
1. <u>[Signature]</u>	3-21-25	<u>[Signature]</u> 3-21-25	Comments:
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2. <u>[Signature]</u>		2. <u>[Signature]</u>	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. <u>[Signature]</u>	3-21-25	3. <u>[Signature]</u>	

Page 1 of 1

CLIENT: ☐ Hand Delivered ☐ Other

Shipment Complete

☐ YES ☐ NO

From: Wright, Stephen <Stephen.E.Wright@aecom.com>
Sent: Monday, March 24, 2025 1:31 PM
To: yazmeen@chemtech.net
Cc: Vivaudou, Eleanor
Subject: Re: Week 3 Bethpage, NY GW Sampling

Yes please...

From: Yazmeen Gomez <yazmeen@chemtech.net>
Sent: Monday, March 24, 2025 11:02 AM
To: Wright, Stephen <Stephen.E.Wright@aecom.com>
Cc: Vivaudou, Eleanor <Eleanor.Vivaudou@aecom.com>
Subject: RE: Week 3 Bethpage, NY GW Sampling

Good morning Stephen,

I wanted to confirm based off the quantity if you needed MS/MSD for both Sample 03 and 04?

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com    

From: Yazmeen Gomez <yazmeen@chemtech.net>
Sent: Monday, March 17, 2025 3:02 PM
To: 'Wright, Stephen' <Stephen.E.Wright@aecom.com>
Cc: 'Vivaudou, Eleanor' <Eleanor.Vivaudou@aecom.com>
Subject: RE: Week 3 Bethpage, NY GW Sampling

Hi Stephen,

Bottle order and pick up for Wednesday – and Pick up for Friday are confirmed as requested.

Best Regards,



Yazmeen Gomez
Sr. Project Manager
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3147
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com     

From: Wright, Stephen <Stephen.E.Wright@aecom.com>
Sent: Monday, March 17, 2025 2:57 PM
To: yazmeen@chemtech.net
Cc: Vivaudou, Eleanor <Eleanor.Vivaudou@aecom.com>
Subject: Week 3 Bethpage, NY GW Sampling

Hi Yazmeen,

We will need courier pickups at the Bethpage site on Wednesday (3/19) at 2PM and Friday (3/21) at 1 PM. I did a bottle count today and we will need one more case of HCL VOAs and one box of amber liters on Wednesday (as well as labels). Hopefully, all sampling for the event will be completed by Friday's pickup.

Let me know if you have any questions...

Steve

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1629	AECO15	Order Date : 3/21/2025 1:55:00 PM	Project Mgr : YAZMEEN
Client Name : AECOM Technical Services		Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only <i>NYS Cat B</i>
Client Contact : Eleanor Vivadou		Receive Date/Time : 3/24/2025 4:15:00 PM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services		Purchase Order : 3/21 17:15	Hard Copy Date :
Invoice Contact : Eleanor Vivadou			Date Signoff : 3/24/2025 10:50:10 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1629-01	TT17411-20250320	Water	03/20/2025	09:20					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-02	TT162S1-20250320	Water	03/20/2025	11:35					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-03	TT188S1-20250320	Water	03/20/2025	11:20					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-04	Q1629-03MS	Water	03/20/2025	11:20					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-05	Q1629-03MSD	Water	03/20/2025	11:20					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-07	RW09-MW01D1-20250320	Water	03/20/2025	11:35					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-08	RW09-MW01D2-20250320	Water	03/20/2025	10:05					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-09	RW09-MW01D3-20250320	Water	03/20/2025	10:03					

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1629 AECO15	Order Date : 3/21/2025 1:55:00 PM	Project Mgr : YAZMEEN
Client Name : AECOM Technical Services	Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only NYS ASP B
Client Contact : Eleanor Vivadou	Receive DateTime : 3/24/2025 4:15:00 PM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services	Purchase Order : 3121 17:15	Hard Copy Date :
Invoice Contact : Eleanor Vivadou		Date Signoff : 3/24/2025 10:50:10 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1629-10	TT150S1-20250320	Water	03/20/2025	13:15	VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-11	FB03-20250321	Water	03/21/2025	00:00	VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-12	TB	Water	03/21/2025	00:00	VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-14	RW09-MW01S-20250320	Water	03/20/2025	11:20	VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-15	Q1629-14MS	Water	03/20/2025	11:20	VOCMS Group1		8260-Low	10 Bus. Days	
Q1629-16	Q1629-14MSD	Water	03/20/2025	11:20	VOCMS Group1		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1629	AECO15	Order Date : 3/21/2025 1:55:00 PM	Project Mgr : YAZMEEN
Client Name : AECOM Technical Services		Project Name : NAVFAC NWIRP Bethpage	Report Type : Results Only NYS ASP B
Client Contact : Eleanor Vivadou		Receive DateTime : 3/24/2025 4:15:00 PM	EDD Type : EQUIS
Invoice Name : AECOM Technical Services		Purchase Order : 3121 17:15	Hard Copy Date :
Invoice Contact : Eleanor Vivadou			Date Signoff : 3/24/2025 10:50:10 AM

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
--------	-----------	--------	----------------	----------------	------	------------	--------	----------	--------------

Relinquished By : cl

Date / Time : 3-24-25 1010

Received By : Samy

Date / Time : 03/24/25 10:10 eg 154

Storage Area : VOA Refridgerator Room