

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: bn032825

SequenceID : bn032825

|  | NA    | NO          | YES         |
|--|-------|-------------|-------------|
| 1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)   | _____ | _____       | _____✓_____ |
| 2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria<br>(NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)  | _____ | _____       | _____✓_____ |
| 3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series   | _____ | _____       | _____✓_____ |
| 4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series | _____ | _____       | _____✓_____ |
| 5. GC/MS Calibration Met:  | _____ | _____       | _____✓_____ |
| a. Initial calibration Meet Criteria<br>If not met, list those compounds and their recoveries which fall outside the acceptable range.   | _____ | _____       | _____✓_____ |
| b. Continuous Calibration(CCC) Meet Criteria<br>If not met, list those compounds and their recoveries which fall outside the acceptable range.   | _____ | _____       | _____✓_____ |
| 6. Blank Contamination - If yes, list compounds and concentrations in each blank:  | _____ | _____✓_____ | _____       |
| a. B/N Fraction  |       |             |             |
| d. Acid Fraction   |       |             |             |

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

a. B/N Fraction

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

a. B/N Fraction

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

Internal standard fail in PB67295BL and PB167269BS which is not associated for required compound.

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

krunal

Analyst

03/31/2025

Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: bn032825

SequenceID : bn032825 NA NO YES