

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM

SAS No.: Q1629

SDG NO.: Q1629

Lab File ID: BN036556.D

DFTPP Injection Date: 03/10/2025

Instrument ID: BNA_N

DFTPP Injection Time: 11:03

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	58.6
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	52.3
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	50.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.8
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	9.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.9 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036557.D	03/10/2025	11:42
SSTDICC0.2	SSTDICC0.2	BN036558.D	03/10/2025	12:18
SSTDICCC0.4	SSTDICCC0.4	BN036559.D	03/10/2025	12:54
SSTDICC0.8	SSTDICC0.8	BN036560.D	03/10/2025	13:31
SSTDICC1.6	SSTDICC1.6	BN036561.D	03/10/2025	14:07
SSTDICC3.2	SSTDICC3.2	BN036562.D	03/10/2025	14:43
SSTDICC5.0	SSTDICC5.0	BN036563.D	03/10/2025	15:19