

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Mahesh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133187,VP133189
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199
CCC	VP133188,VP133190
Internal Standard/PEM	
ICV/I.BLK	VP133200
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045067.D	28 Feb 2025 01:03		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX045068.D	28 Feb 2025 01:27		JC/MD	Ok,M
3	VSTDIC005	VSTDIC005	VX045069.D	28 Feb 2025 02:13		JC/MD	Ok,M
4	VSTDIC020	VSTDIC020	VX045070.D	28 Feb 2025 02:37		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX045071.D	28 Feb 2025 03:00		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX045072.D	28 Feb 2025 03:23		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX045073.D	28 Feb 2025 03:47		JC/MD	Ok,M
8	IBLK	IBLK	VX045074.D	28 Feb 2025 04:10		JC/MD	Ok
9	VSTDICV050	ICVVX022825	VX045075.D	28 Feb 2025 04:33		JC/MD	Ok,M
10	BFB	BFB	VX045076.D	28 Feb 2025 10:03		JC/MD	Ok
11	VSTDIC050	VSTDIC050	VX045077.D	28 Feb 2025 10:32	V12668	JC/MD	Ok,M
12	VX0228MBL01	VX0228MBL01	VX045078.D	28 Feb 2025 11:00		JC/MD	Ok
13	VX0228WBL01	VX0228WBL01	VX045079.D	28 Feb 2025 11:23		JC/MD	Ok
14	VX0228WBS01	VX0228WBS01	VX045080.D	28 Feb 2025 11:46		JC/MD	Ok,M
15	VX0228WBSD01	VX0228WBSD01	VX045081.D	28 Feb 2025 12:13		JC/MD	Ok,M
16	Q1401-03	BP-VPB-192-GW-840-8	VX045082.D	28 Feb 2025 12:37	vial B pH<2	JC/MD	Ok
17	Q1401-06	BP-VPB-192-GW-900-9	VX045083.D	28 Feb 2025 13:00	vial B pH<2	JC/MD	Ok
18	Q1423-01	BP-VPB-192-TB-20250	VX045084.D	28 Feb 2025 13:23	vial B pH<2 TB	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX022825

Review By	Maresh Dadoda	Review On	2/28/2025 11:09:50 AM
Supervise By	Semsettin Yesilyurt	Supervise On	2/28/2025 11:11:09 AM
SubDirectory	VX022825	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133187,VP133189		
Initial Calibration Stds	VP133194,VP133195,VP133196,VP133197,VP133198,VP133199		
CCC	VP133188,VP133190		
Internal Standard/PEM			
ICV/I.BLK	VP133200		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q1423-03	BP-VPB-192-EB-20250	VX045085.D	28 Feb 2025 13:47	vial B pH<2 EB	JC/MD	Ok
20	Q1403-01	Storage-Blank-SOIL-RE	VX045086.D	28 Feb 2025 14:10	vial B pH<2	JC/MD	Ok
21	Q1403-02	Storage-Blank-WATER-	VX045087.D	28 Feb 2025 14:33	vial B pH<2	JC/MD	Ok
22	Q1435-01	286107	VX045088.D	28 Feb 2025 14:57	vial A pH<2	JC/MD	Ok
23	Q1403-04	Storage-Blank-SAMPLE	VX045089.D	28 Feb 2025 15:20	vial B pH<2	JC/MD	Ok
24	Q1462-02	FB	VX045090.D	28 Feb 2025 15:44	vial A pH<2 FB	JC/MD	Ok
25	Q1469-01	Storage-Blank-SOIL-RE	VX045091.D	28 Feb 2025 16:07	vial A pH<2	JC/MD	Ok
26	Q1469-02	Storage-Blank-WATER-	VX045092.D	28 Feb 2025 16:31	vial A pH<2	JC/MD	Ok
27	Q1469-03	Storage-Blank-WATER-	VX045093.D	28 Feb 2025 16:54	vial A pH<2	JC/MD	Ok
28	Q1469-04	Storage-Blank-SAMPLE	VX045094.D	28 Feb 2025 17:17	vial A pH<2	JC/MD	Ok
29	Q1403-03	Storage-Blank-WATER-	VX045095.D	28 Feb 2025 17:41	vial B pH<2	JC/MD	Ok
30	Q1462-01	MW2	VX045096.D	28 Feb 2025 18:04	vial A pH<2	JC/MD	Ok,M
31	IBLK	IBLK	VX045097.D	28 Feb 2025 18:27		JC/MD	Ok
32	VSTDCCC050	VSTDCCC050EC	VX045098.D	28 Feb 2025 18:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032525

Review By	John Carlone	Review On	3/26/2025 10:08:08 AM
Supervise By	Semsettin Yesilyurt	Supervise On	3/26/2025 1:11:57 PM
SubDirectory	VX032525	HP Acquire Method	HP Processing Method 82X022825W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133484
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133485,VP133486

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045430.D	25 Mar 2025 09:37		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045431.D	25 Mar 2025 11:31	CCAL failed high for comp. #2	JC/MD	Ok,M
3	VX0325MBL01	VX0325MBL01	VX045432.D	25 Mar 2025 11:59		JC/MD	Ok
4	VX0325WBL01	VX0325WBL01	VX045433.D	25 Mar 2025 12:22		JC/MD	Ok
5	VX0325WBS01	VX0325WBS01	VX045434.D	25 Mar 2025 12:46	Surrogate Fail	JC/MD	Not Ok
6	Q1608-10DL	DUP08-20250318DL	VX045435.D	25 Mar 2025 13:12	vial B pH<2	JC/MD	Ok
7	VX0325WBS02	VX0325WBS02	VX045436.D	25 Mar 2025 13:35		JC/MD	Ok,M
8	Q1608-17DL	RE132D5-20250318DL	VX045437.D	25 Mar 2025 14:03	vial B pH<2	JC/MD	Ok
9	Q1608-19DL	RE132D7-20250318DL	VX045438.D	25 Mar 2025 14:26	vial B pH<2	JC/MD	Ok
10	Q1608-21DL	TT101D2-20250319DL	VX045439.D	25 Mar 2025 14:50	vial B pH<2	JC/MD	Ok
11	Q1608-18DL	RE132D6-20250318DL	VX045440.D	25 Mar 2025 15:13	vial B pH<2	JC/MD	Ok
12	Q1608-14RE	RE132D2-20250318RE	VX045441.D	25 Mar 2025 15:36	vial B pH<2,CCAL Fail for Comp. #2	JC/MD	Confirms
13	Q1608-16DL	RE132D4-20250318DL	VX045442.D	25 Mar 2025 15:59	vial B pH<2	JC/MD	Ok
14	Q1629-12	TB	VX045443.D	25 Mar 2025 16:22	vial A pH<2 TB	JC/MD	Ok
15	Q1629-11	FB03-20250321	VX045444.D	25 Mar 2025 16:46	vial A pH<2 FB	JC/MD	Ok
16	Q1629-01	TT17411-20250320	VX045445.D	25 Mar 2025 17:09	vial A pH<2	JC/MD	Ok
17	Q1629-02	TT162S1-20250320	VX045446.D	25 Mar 2025 17:32	vial A pH<2	JC/MD	Ok

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032525

Review By	John Carlone	Review On	3/26/2025 10:08:08 AM
Supervise By	Semsettin Yesilyurt	Supervise On	3/26/2025 1:11:57 PM
SubDirectory	VX032525	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133484		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133485,VP133486		

18	Q1629-07	RW09-MW01D1-20250	VX045447.D	25 Mar 2025 17:55	vial A pH<2	JC/MD	Ok
19	Q1629-08	RW09-MW01D2-20250	VX045448.D	25 Mar 2025 18:18	vial A pH<2	JC/MD	Ok
20	Q1629-09	RW09-MW01D3-20250	VX045449.D	25 Mar 2025 18:41	vial A pH<2	JC/MD	Ok
21	Q1629-10	TT150S1-20250320	VX045450.D	25 Mar 2025 19:05	vial A pH<2	JC/MD	Ok
22	VX0325WBSD02	VX0325WBSD02	VX045451.D	25 Mar 2025 19:28	Surrogate Fail	JC/MD	Not Ok
23	Q1608-11	RE132D1-20250318	VX045452.D	25 Mar 2025 19:51	vial A pH<2	JC/MD	Ok
24	Q1608-12MS	RE132D1-20250318MS	VX045453.D	25 Mar 2025 20:14	vial A pH<2	JC/MD	Ok,M
25	Q1608-13MSD	RE132D1-20250318MS	VX045454.D	25 Mar 2025 20:37	vial A pH<2	JC/MD	Ok,M
26	VSTDCCC050	VSTDCCC050EC	VX045455.D	25 Mar 2025 21:00		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX032625

Review By	John Carlone	Review On	3/28/2025 9:50:07 AM
Supervise By	Maresh Dadoda	Supervise On	3/28/2025 3:29:43 PM
SubDirectory	VX032625	HP Acquire Method	HP Processing Method 82X022825W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133490		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133491,VP133492		

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX045456.D	26 Mar 2025 09:33		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX045457.D	26 Mar 2025 10:03	pH#Lot#V12668	JC/MD	Ok,M
3	VX0326MBL01	VX0326MBL01	VX045458.D	26 Mar 2025 10:31		JC/MD	Ok
4	VX0326WBL01	VX0326WBL01	VX045459.D	26 Mar 2025 10:54		JC/MD	Ok
5	VX0326WBS01	VX0326WBS01	VX045460.D	26 Mar 2025 11:18		JC/MD	Ok,M
6	Q1629-14	RW09-MW01S-202503	VX045461.D	26 Mar 2025 11:46	vial A pH<2	JC/MD	Ok
7	Q1629-15MS	RW09-MW01S-202503	VX045462.D	26 Mar 2025 12:09	vial A pH<2	JC/MD	Ok,M
8	Q1629-16MSD	RW09-MW01S-202503	VX045463.D	26 Mar 2025 12:33	vial A pH<2	JC/MD	Ok,M
9	IBLK	IBLK	VX045464.D	26 Mar 2025 12:56		JC/MD	Ok
10	IBLK	IBLK	VX045465.D	26 Mar 2025 13:19		JC/MD	Ok
11	Q1629-03	TT188S1-20250320	VX045466.D	26 Mar 2025 13:43	vial A pH<2	JC/MD	Ok
12	Q1629-04MS	TT188S1-20250320MS	VX045467.D	26 Mar 2025 14:06	vial A pH<2	JC/MD	Ok,M
13	Q1629-05MSD	TT188S1-20250320MS	VX045468.D	26 Mar 2025 14:29	vial A pH<2	JC/MD	Ok,M
14	IBLK	IBLK	VX045469.D	26 Mar 2025 14:53		JC/MD	Ok
15	VSTDCCC050	VSTDCCC050EC	VX045470.D	26 Mar 2025 15:35		JC/MD	Ok,M

M : Manual Integration