

LAB CHRONICLE

OrderID:	Q1629	OrderDate:	3/21/2025 1:55:00 PM
Client:	AECOM Technical Services, Inc.	Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258
Contact:	Eleanor Vivadou	Location:	I41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1629-01	TT174I1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-02	TT162S1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-03	TT188S1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/26/25	03/21/25
Q1629-07	RW09-MW01D1-2025 0320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-08	RW09-MW01D2-2025 0320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-09	RW09-MW01D3-2025 0320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-10	TT150S1-20250320	Water	VOCMS Group1	8260-Low	03/20/25		03/25/25	03/21/25
Q1629-11	FB03-20250321	Water	VOCMS Group1	8260-Low	03/21/25		03/25/25	03/21/25
Q1629-12	TB	Water	VOCMS Group1	8260-Low	03/21/25		03/25/25	03/21/25
Q1629-14	RW09-MW01S-20250 320	Water	VOCMS Group1	8260-Low	03/20/25		03/26/25	03/21/25

Hit Summary Sheet
SW-846

SDG No.: Q1629
Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	TT174I1-20250320								
Q1629-01	TT174I1-20250320	Water	1,1-Dichloroethene	1.20		0.23	0.75	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	Acetone	2.50	J	1.50	3.80	5.00	ug/L
Q1629-01	TT174I1-20250320	Water	Methyl tert-butyl Ether	0.39	J	0.16	0.50	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	1,1-Dichloroethane	3.40		0.23	0.50	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	Trichloroethene	2.00		0.090	0.75	1.00	ug/L
Q1629-01	TT174I1-20250320	Water	Tetrachloroethene	0.42	J	0.23	0.50	1.00	ug/L
			Total Voc :	9.91					
			Total Concentration:	9.91					
Client ID:	TT162S1-20250320								
Q1629-02	TT162S1-20250320	Water	Acetone	1.80	J	1.50	3.80	5.00	ug/L
Q1629-02	TT162S1-20250320	Water	Trichloroethene	11.9		0.090	0.75	1.00	ug/L
Q1629-02	TT162S1-20250320	Water	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
			Total Voc :	14.8					
			Total Concentration:	14.8					
Client ID:	TT188S1-20250320								
Q1629-03	TT188S1-20250320	Water	Acetone	2.20	J	1.50	3.80	5.00	ug/L
Q1629-03	TT188S1-20250320	Water	Trichloroethene	5.30		0.090	0.75	1.00	ug/L
			Total Voc :	7.50					
			Total Concentration:	7.50					
Client ID:	RW09-MW01D1-20250320								
Q1629-07	RW09-MW01D1-20250320	Water	Acetone	1.80	J	1.50	3.80	5.00	ug/L
			Total Voc :	1.80					
			Total Concentration:	1.80					
Client ID:	RW09-MW01D2-20250320								
Q1629-08	RW09-MW01D2-20250320	Water	Acetone	2.00	J	1.50	3.80	5.00	ug/L
			Total Voc :	2.00					
			Total Concentration:	2.00					
Client ID:	TT150S1-20250320								
Q1629-10	TT150S1-20250320	Water	Acetone	2.50	J	1.50	3.80	5.00	ug/L
Q1629-10	TT150S1-20250320	Water	Trichloroethene	11.6		0.090	0.75	1.00	ug/L
Q1629-10	TT150S1-20250320	Water	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
			Total Voc :	15.2					
			Total Concentration:	15.2					
Client ID:	FB03-20250321								
Q1629-11	FB03-20250321	Water	Acetone	4.00	J	1.50	3.80	5.00	ug/L
			Total Voc :	4.00					
			Total Concentration:	4.00					
Client ID:	TB								
Q1629-12	TB	Water	Acetone	1.60	J	1.50	3.80	5.00	ug/L

Hit Summary Sheet
SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Total Voc :				1.60					
Total Concentration:				1.60					



SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT174I1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045445.D	1		03/25/25 17:09	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	1.20		0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.39	J	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	3.40		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	2.00		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT174I1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045445.D	1		03/25/25 17:09	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.42	J	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.0		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	48.8		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	47.2		89 - 112		94%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.5		85 - 114		97%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70100	5.55				
540-36-3	1,4-Difluorobenzene	138000	6.757				
3114-55-4	Chlorobenzene-d5	122000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	47200	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT174I1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045445.D	1		03/25/25 17:09	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT162S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045446.D	1		03/25/25 17:32	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	11.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT162S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045446.D	1		03/25/25 17:32	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.9		81 - 118		112%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		80 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	47.5		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.9		85 - 114		104%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	68000	5.544				
540-36-3	1,4-Difluorobenzene	134000	6.757				
3114-55-4	Chlorobenzene-d5	122000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	51000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT162S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045446.D	1		03/25/25 17:32	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045466.D	1		03/26/25 13:43	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.20	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQM	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	UM	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	5.30		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045466.D	1		03/26/25 13:43	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	UM	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	UM	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	UM	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.5		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	49.4		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	46.4		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.8		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	68100	5.544				
540-36-3	1,4-Difluorobenzene	134000	6.757				
3114-55-4	Chlorobenzene-d5	120000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	47700	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-03		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045466.D	1		03/26/25 13:43	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045447.D	1		03/25/25 17:55	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.80	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045447.D	1		03/25/25 17:55	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.3		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	47.8		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.7		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	74700	5.543				
540-36-3	1,4-Difluorobenzene	147000	6.757				
3114-55-4	Chlorobenzene-d5	136000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	52500	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-07		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045447.D	1		03/25/25 17:55	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D2-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045448.D	1		03/25/25 18:18	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.00	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D2-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045448.D	1		03/25/25 18:18	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.1		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	50.6		80 - 119		101%	SPK: 50
2037-26-5	Toluene-d8	48.7		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	76200	5.55				
540-36-3	1,4-Difluorobenzene	148000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	55200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D2-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-08		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045448.D	1		03/25/25 18:18	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D3-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045449.D	1		03/25/25 18:41	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D3-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045449.D	1		03/25/25 18:41	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.2		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	48.0		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.7		85 - 114		103%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71700	5.544				
540-36-3	1,4-Difluorobenzene	137000	6.757				
3114-55-4	Chlorobenzene-d5	129000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	54400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01D3-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045449.D	1		03/25/25 18:41	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT150S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045450.D	1		03/25/25 19:05	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	2.50	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	11.6		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT150S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045450.D	1		03/25/25 19:05	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	1.10		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.6		81 - 118		109%	SPK: 50
1868-53-7	Dibromofluoromethane	49.7		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	47.3		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.9		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	72000	5.55				
540-36-3	1,4-Difluorobenzene	142000	6.757				
3114-55-4	Chlorobenzene-d5	131000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	55200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT150S1-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-10		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045450.D	1		03/25/25 19:05	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	FB03-20250321		SDG No.:	Q1629	
Lab Sample ID:	Q1629-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045444.D	1		03/25/25 16:46	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	4.00	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	FB03-20250321		SDG No.:	Q1629	
Lab Sample ID:	Q1629-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045444.D	1		03/25/25 16:46	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.4		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	47.9		89 - 112		96%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.6		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71100	5.544				
540-36-3	1,4-Difluorobenzene	141000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	49200	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	FB03-20250321		SDG No.:	Q1629	
Lab Sample ID:	Q1629-11		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045444.D	1		03/25/25 16:46	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TB		SDG No.:	Q1629	
Lab Sample ID:	Q1629-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045443.D	1		03/25/25 16:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	1.60	J	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TB		SDG No.:	Q1629	
Lab Sample ID:	Q1629-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045443.D	1		03/25/25 16:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.0		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.1		80 - 119		98%	SPK: 50
2037-26-5	Toluene-d8	47.3		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.6		85 - 114		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	70900	5.549				
540-36-3	1,4-Difluorobenzene	138000	6.756				
3114-55-4	Chlorobenzene-d5	128000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	51700	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/21/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TB		SDG No.:	Q1629	
Lab Sample ID:	Q1629-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045443.D	1		03/25/25 16:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045461.D	1		03/26/25 11:46	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	UQ	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045461.D	1		03/26/25 11:46	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.9		81 - 118		102%	SPK: 50
1868-53-7	Dibromofluoromethane	48.4		80 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	47.1		89 - 112		94%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.3		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	72700	5.544				
540-36-3	1,4-Difluorobenzene	140000	6.757				
3114-55-4	Chlorobenzene-d5	130000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	52400	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320		SDG No.:	Q1629	
Lab Sample ID:	Q1629-14		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045461.D	1		03/26/25 11:46	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q1629**

Client: **AECOM Technical Services, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1629-01	TT174I1-20250320	1,2-Dichloroethane-d4	50	54.0	108	81	118
		Dibromofluoromethane	50	48.8	98	80	119
		Toluene-d8	50	47.2	94	89	112
		4-Bromofluorobenzene	50	48.5	97	85	114
Q1629-02	TT162S1-20250320	1,2-Dichloroethane-d4	50	55.9	112	81	118
		Dibromofluoromethane	50	50.1	100	80	119
		Toluene-d8	50	47.5	95	89	112
		4-Bromofluorobenzene	50	51.9	104	85	114
Q1629-03	TT188S1-20250320	1,2-Dichloroethane-d4	50	53.5	107	81	118
		Dibromofluoromethane	50	49.4	99	80	119
		Toluene-d8	50	46.4	93	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
Q1629-04MS	TT188S1-20250320MS	1,2-Dichloroethane-d4	50	53.7	107	81	118
		Dibromofluoromethane	50	49.3	99	80	119
		Toluene-d8	50	46.5	93	89	112
		4-Bromofluorobenzene	50	51.1	102	85	114
Q1629-05MSD	TT188S1-20250320MSD	1,2-Dichloroethane-d4	50	56.3	113	81	118
		Dibromofluoromethane	50	53.0	106	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	57.5	115 *	85	114
Q1629-07	RW09-MW01D1-20250320	1,2-Dichloroethane-d4	50	55.3	111	81	118
		Dibromofluoromethane	50	49.7	99	80	119
		Toluene-d8	50	47.9	96	89	112
		4-Bromofluorobenzene	50	50.7	101	85	114
Q1629-08	RW09-MW01D2-20250320	1,2-Dichloroethane-d4	50	55.1	110	81	118
		Dibromofluoromethane	50	50.6	101	80	119
		Toluene-d8	50	48.7	97	89	112
		4-Bromofluorobenzene	50	50.5	101	85	114
Q1629-09	RW09-MW01D3-20250320	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	48.0	96	89	112
		4-Bromofluorobenzene	50	51.7	103	85	114
Q1629-10	TT150S1-20250320	1,2-Dichloroethane-d4	50	54.6	109	81	118
		Dibromofluoromethane	50	49.7	99	80	119
		Toluene-d8	50	47.3	95	89	112
		4-Bromofluorobenzene	50	50.9	102	85	114
Q1629-11	FB03-20250321	1,2-Dichloroethane-d4	50	55.4	111	81	118
		Dibromofluoromethane	50	49.5	99	80	119
		Toluene-d8	50	47.9	96	89	112
		4-Bromofluorobenzene	50	49.6	99	85	114
Q1629-12	TB	1,2-Dichloroethane-d4	50	53.0	106	81	118
		Dibromofluoromethane	50	49.1	98	80	119
		Toluene-d8	50	47.3	95	89	112
		4-Bromofluorobenzene	50	50.6	101	85	114
Q1629-14	RW09-MW01S-20250320	1,2-Dichloroethane-d4	50	50.9	102	81	118
		Dibromofluoromethane	50	48.4	97	80	119
		Toluene-d8	50	47.1	94	89	112
		4-Bromofluorobenzene	50	49.3	99	85	114
Q1629-15MS	RW09-MW01S-20250320MS	1,2-Dichloroethane-d4	50	54.0	108	81	118
		Dibromofluoromethane	50	51.2	102	80	119

Surrogate Summary

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1629-15MS	RW09-MW01S-20250320MS	Toluene-d8	50	47.4	95	89	112
		4-Bromofluorobenzene	50	53.5	107	85	114
Q1629-16MSD	RW09-MW01S-20250320MSD	1,2-Dichloroethane-d4	50	41.9	84	81	118
		Dibromofluoromethane	50	40.1	80	80	119
		Toluene-d8	50	38.1	76 *	89	112
		4-Bromofluorobenzene	50	42.5	85	85	114
VX0325WBL01	VX0325WBL01	1,2-Dichloroethane-d4	50	58.3	117	81	118
		Dibromofluoromethane	50	54.2	108	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	54.8	110	85	114
VX0325WBS02	VX0325WBS02	1,2-Dichloroethane-d4	50	56.6	113	81	118
		Dibromofluoromethane	50	53.4	107	80	119
		Toluene-d8	50	50.2	100	89	112
		4-Bromofluorobenzene	50	54.3	109	85	114
VX0326WBL01	VX0326WBL01	1,2-Dichloroethane-d4	50	53.2	106	81	118
		Dibromofluoromethane	50	48.1	96	80	119
		Toluene-d8	50	46.4	93	89	112
		4-Bromofluorobenzene	50	47.1	94	85	114
VX0326WBS01	VX0326WBS01	1,2-Dichloroethane-d4	50	53.8	108	81	118
		Dibromofluoromethane	50	51.0	102	80	119
		Toluene-d8	50	48.3	97	89	112
		4-Bromofluorobenzene	50	52.8	106	85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

		Sample				Rec	RPD		Limits		
Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID :	Q1629-15MS	Client Sample ID :	RW09-MW01S-20250320MS					Datafile :	VX045462.D		
Dichlorodifluoromethane	50	0	61.2	ug/L	122				32	152	
Chloromethane	50	0	50.8	ug/L	102				50	139	
Vinyl chloride	50	0	48.8	ug/L	98				58	137	
Bromomethane	50	0	54.3	ug/L	109				53	141	
Chloroethane	50	0	54.0	ug/L	108				60	138	
Trichlorofluoromethane	50	0	53.8	ug/L	108				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	59.0	ug/L	118				70	136	
1,1-Dichloroethene	50	0	52.8	ug/L	106				71	131	
Acetone	250	0	260	ug/L	104				39	160	
Carbon disulfide	50	0	48.6	ug/L	97				64	133	
Methyl tert-butyl Ether	50	0	51.9	ug/L	104				71	124	
Methyl Acetate	50	0	48.3	ug/L	97				56	136	
Methylene Chloride	50	0	50.4	ug/L	101				74	124	
trans-1,2-Dichloroethene	50	0	52.9	ug/L	106				75	124	
1,1-Dichloroethane	50	0	52.3	ug/L	105				77	125	
Cyclohexane	50	0	53.6	ug/L	107				71	130	
2-Butanone	250	0	270	ug/L	108				56	143	
Carbon Tetrachloride	50	0	54.9	ug/L	110				72	136	
cis-1,2-Dichloroethene	50	0	51.1	ug/L	102				78	123	
Bromochloromethane	50	0	49.7	ug/L	99				78	123	
Chloroform	50	0	53.9	ug/L	108				79	124	
1,1,1-Trichloroethane	50	0	56.0	ug/L	112				74	131	
Methylcyclohexane	50	0	57.5	ug/L	115				72	132	
Benzene	50	0	51.3	ug/L	103				79	120	
1,2-Dichloroethane	50	0	57.6	ug/L	115				73	128	
Trichloroethene	50	0	49.6	ug/L	99				79	123	
1,2-Dichloropropane	50	0	50.6	ug/L	101				78	122	
Bromodichloromethane	50	0	53.9	ug/L	108				79	125	
4-Methyl-2-Pentanone	250	0	280	ug/L	112				67	130	
Toluene	50	0	52.9	ug/L	106				80	121	
t-1,3-Dichloropropene	50	0	57.5	ug/L	115				73	127	
cis-1,3-Dichloropropene	50	0	55.7	ug/L	111				75	124	
1,1,2-Trichloroethane	50	0	51.9	ug/L	104				80	119	
2-Hexanone	250	0	280	ug/L	112				57	139	
Dibromochloromethane	50	0	54.4	ug/L	109				74	126	
1,2-Dibromoethane	50	0	53.2	ug/L	106				77	121	
Tetrachloroethene	50	0	53.2	ug/L	106				74	129	
Chlorobenzene	50	0	52.9	ug/L	106				82	118	
Ethyl Benzene	50	0	54.9	ug/L	110				79	121	
m/p-Xylenes	100	0	110	ug/L	110				80	121	
o-Xylene	50	0	54.1	ug/L	108				78	122	
Styrene	50	0	56.1	ug/L	112				78	123	
Bromoform	50	0	54.6	ug/L	109				66	130	
Isopropylbenzene	50	0	55.1	ug/L	110				72	131	
1,1,2,2-Tetrachloroethane	50	0	50.8	ug/L	102				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec			RPD	Limits		
					Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	52.0	ug/L	104				80	119	
1,4-Dichlorobenzene	50	0	51.2	ug/L	102				79	118	
1,2-Dichlorobenzene	50	0	52.2	ug/L	104				80	119	
1,2-Dibromo-3-Chloropropane	50	0	58.9	ug/L	118				62	128	
1,2,4-Trichlorobenzene	50	0	51.7	ug/L	103				69	130	
1,2,3-Trichlorobenzene	50	0	51.3	ug/L	103				69	129	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec		RPD		Limits		RPD
		Result			Qual	RPD	Qual	Low	High		
Lab Sample ID :	Q1629-16MSD	Client Sample ID :	RW09-MW01S-20250320MSD					Datafile :	VX045463.D		
Dichlorodifluoromethane	50	0	61.4	ug/L	123		0		32	152	20
Chloromethane	50	0	48.8	ug/L	98		4		50	139	20
Vinyl chloride	50	0	48.1	ug/L	96		1		58	137	20
Bromomethane	50	0	55.4	ug/L	111		2		53	141	20
Chloroethane	50	0	56.6	ug/L	113		5		60	138	20
Trichlorofluoromethane	50	0	53.1	ug/L	106		1		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	58.3	ug/L	117		1		70	136	20
1,1-Dichloroethene	50	0	52.9	ug/L	106		0		71	131	20
Acetone	250	0	260	ug/L	104		0		39	160	20
Carbon disulfide	50	0	50.0	ug/L	100		3		64	133	20
Methyl tert-butyl Ether	50	0	53.7	ug/L	107		3		71	124	20
Methyl Acetate	50	0	49.5	ug/L	99		2		56	136	20
Methylene Chloride	50	0	50.6	ug/L	101		0		74	124	20
trans-1,2-Dichloroethene	50	0	53.1	ug/L	106		0		75	124	20
1,1-Dichloroethane	50	0	52.7	ug/L	105		1		77	125	20
Cyclohexane	50	0	54.9	ug/L	110		2		71	130	20
2-Butanone	250	0	270	ug/L	108		0		56	143	20
Carbon Tetrachloride	50	0	58.7	ug/L	117		7		72	136	20
cis-1,2-Dichloroethene	50	0	52.9	ug/L	106		3		78	123	20
Bromochloromethane	50	0	48.4	ug/L	97		3		78	123	20
Chloroform	50	0	55.1	ug/L	110		2		79	124	20
1,1,1-Trichloroethane	50	0	57.6	ug/L	115		3		74	131	20
Methylcyclohexane	50	0	60.0	ug/L	120		4		72	132	20
Benzene	50	0	53.4	ug/L	107		4		79	120	20
1,2-Dichloroethane	50	0	60.7	ug/L	121		5		73	128	20
Trichloroethene	50	0	54.0	ug/L	108		8		79	123	20
1,2-Dichloropropane	50	0	54.1	ug/L	108		7		78	122	20
Bromodichloromethane	50	0	57.6	ug/L	115		7		79	125	20
4-Methyl-2-Pentanone	250	0	290	ug/L	116		4		67	130	20
Toluene	50	0	55.7	ug/L	111		5		80	121	20
t-1,3-Dichloropropene	50	0	61.9	ug/L	124		7		73	127	20
cis-1,3-Dichloropropene	50	0	59.3	ug/L	119		6		75	124	20
1,1,2-Trichloroethane	50	0	54.5	ug/L	109		5		80	119	20
2-Hexanone	250	0	300	ug/L	120		7		57	139	20
Dibromochloromethane	50	0	58.5	ug/L	117		7		74	126	20
1,2-Dibromoethane	50	0	56.7	ug/L	113		6		77	121	20
Tetrachloroethene	50	0	53.5	ug/L	107		1		74	129	20
Chlorobenzene	50	0	54.3	ug/L	109		3		82	118	20
Ethyl Benzene	50	0	56.4	ug/L	113		3		79	121	20
m/p-Xylenes	100	0	110	ug/L	110		0		80	121	20
o-Xylene	50	0	54.6	ug/L	109		1		78	122	20
Styrene	50	0	57.2	ug/L	114		2		78	123	20
Bromoform	50	0	56.6	ug/L	113		4		66	130	20
Isopropylbenzene	50	0	56.7	ug/L	113		3		72	131	20
1,1,2,2-Tetrachloroethane	50	0	53.3	ug/L	107		5		71	121	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec			RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High		
1,3-Dichlorobenzene	50	0	53.6	ug/L	107		3		80	119	20	
1,4-Dichlorobenzene	50	0	53.5	ug/L	107		4		79	118	20	
1,2-Dichlorobenzene	50	0	53.2	ug/L	106		2		80	119	20	
1,2-Dibromo-3-Chloropropane	50	0	60.7	ug/L	121		3		62	128	20	
1,2,4-Trichlorobenzene	50	0	53.9	ug/L	108		4		69	130	20	
1,2,3-Trichlorobenzene	50	0	52.6	ug/L	105		3		69	129	20	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
Lab Sample ID :	Q1629-04MS	Client Sample ID :	TT188S1-20250320MS						Datafile :	VX045467.D	
Dichlorodifluoromethane	50	0	46.4	ug/L	93				32	152	
Chloromethane	50	0	40.7	ug/L	81				50	139	
Vinyl chloride	50	0	40.6	ug/L	81				58	137	
Bromomethane	50	0	46.0	ug/L	92				53	141	
Chloroethane	50	0	43.2	ug/L	86				60	138	
Trichlorofluoromethane	50	0	46.8	ug/L	94				65	141	
1,1,2-Trichlorotrifluoroethane	50	0	54.8	ug/L	110				70	136	
1,1-Dichloroethene	50	0	47.9	ug/L	96				71	131	
Acetone	250	2.20	280	ug/L	111				39	160	
Carbon disulfide	50	0	34.3	ug/L	69				64	133	
Methyl tert-butyl Ether	50	0	54.1	ug/L	108				71	124	
Methyl Acetate	50	0	71.5	ug/L	143	*			56	136	
Methylene Chloride	50	0	49.1	ug/L	98				74	124	
trans-1,2-Dichloroethene	50	0	47.9	ug/L	96				75	124	
1,1-Dichloroethane	50	0	52.7	ug/L	105				77	125	
Cyclohexane	50	0	46.2	ug/L	92				71	130	
2-Butanone	250	0	290	ug/L	116				56	143	
Carbon Tetrachloride	50	0	53.1	ug/L	106				72	136	
cis-1,2-Dichloroethene	50	0	50.5	ug/L	101				78	123	
Bromochloromethane	50	0	53.5	ug/L	107				78	123	
Chloroform	50	0	55.2	ug/L	110				79	124	
1,1,1-Trichloroethane	50	0	55.3	ug/L	111				74	131	
Methylcyclohexane	50	0	48.9	ug/L	98				72	132	
Benzene	50	0	49.7	ug/L	99				79	120	
1,2-Dichloroethane	50	0	58.4	ug/L	117				73	128	
Trichloroethene	50	5.30	53.5	ug/L	96				79	123	
1,2-Dichloropropane	50	0	51.4	ug/L	103				78	122	
Bromodichloromethane	50	0	56.3	ug/L	113				79	125	
4-Methyl-2-Pentanone	250	0	290	ug/L	116				67	130	
Toluene	50	0	51.6	ug/L	103				80	121	
t-1,3-Dichloropropene	50	0	56.5	ug/L	113				73	127	
cis-1,3-Dichloropropene	50	0	55.2	ug/L	110				75	124	
1,1,2-Trichloroethane	50	0	53.5	ug/L	107				80	119	
2-Hexanone	250	0	300	ug/L	120				57	139	
Dibromochloromethane	50	0	56.2	ug/L	112				74	126	
1,2-Dibromoethane	50	0	53.5	ug/L	107				77	121	
Tetrachloroethene	50	0	49.3	ug/L	99				74	129	
Chlorobenzene	50	0	51.8	ug/L	104				82	118	
Ethyl Benzene	50	0	53.2	ug/L	106				79	121	
m/p-Xylenes	100	0	100	ug/L	100				80	121	
o-Xylene	50	0	51.7	ug/L	103				78	122	
Styrene	50	0	54.0	ug/L	108				78	123	
Bromoform	50	0	54.9	ug/L	110				66	130	
Isopropylbenzene	50	0	51.8	ug/L	104				72	131	
1,1,2,2-Tetrachloroethane	50	0	52.0	ug/L	104				71	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Rec	Qual	RPD	Qual	Low	High	RPD
1,3-Dichlorobenzene	50	0	50.4	ug/L	101				80	119	
1,4-Dichlorobenzene	50	0	50.0	ug/L	100				79	118	
1,2-Dichlorobenzene	50	0	52.3	ug/L	105				80	119	
1,2-Dibromo-3-Chloropropane	50	0	60.7	ug/L	121				62	128	
1,2,4-Trichlorobenzene	50	0	51.8	ug/L	104				69	130	
1,2,3-Trichlorobenzene	50	0	51.1	ug/L	102				69	129	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample Result	Result	Units	Rec		RPD		Limits		RPD
					Rec	Qual	RPD	Qual	Low	High	
Lab Sample ID :	Q1629-05MSD	Client Sample ID :	TT188S1-20250320MSD					Datafile :	VX045468.D		
Dichlorodifluoromethane	50	0	65.0	ug/L	130		33	*	32	152	20
Chloromethane	50	0	55.3	ug/L	111		30	*	50	139	20
Vinyl chloride	50	0	52.1	ug/L	104		25	*	58	137	20
Bromomethane	50	0	59.9	ug/L	120		26	*	53	141	20
Chloroethane	50	0	54.7	ug/L	109		23	*	60	138	20
Trichlorofluoromethane	50	0	57.1	ug/L	114		20		65	141	20
1,1,2-Trichlorotrifluoroethane	50	0	63.4	ug/L	127		15		70	136	20
1,1-Dichloroethene	50	0	57.1	ug/L	114		18		71	131	20
Acetone	250	2.20	300	ug/L	119		7		39	160	20
Carbon disulfide	50	0	52.6	ug/L	105		42	*	64	133	20
Methyl tert-butyl Ether	50	0	58.3	ug/L	117		7		71	124	20
Methyl Acetate	50	0	51.2	ug/L	102		33	*	56	136	20
Methylene Chloride	50	0	55.0	ug/L	110		11		74	124	20
trans-1,2-Dichloroethene	50	0	57.5	ug/L	115		18		75	124	20
1,1-Dichloroethane	50	0	57.7	ug/L	115		9		77	125	20
Cyclohexane	50	0	58.0	ug/L	116		23	*	71	130	20
2-Butanone	250	0	310	ug/L	124		7		56	143	20
Carbon Tetrachloride	50	0	61.2	ug/L	122		14		72	136	20
cis-1,2-Dichloroethene	50	0	56.9	ug/L	114		12		78	123	20
Bromochloromethane	50	0	49.6	ug/L	99		8		78	123	20
Chloroform	50	0	59.6	ug/L	119		8		79	124	20
1,1,1-Trichloroethane	50	0	62.0	ug/L	124		11		74	131	20
Methylcyclohexane	50	0	61.0	ug/L	122		22	*	72	132	20
Benzene	50	0	56.5	ug/L	113		13		79	120	20
1,2-Dichloroethane	50	0	64.8	ug/L	130	*	10		73	128	20
Trichloroethene	50	5.30	60.4	ug/L	110		14		79	123	20
1,2-Dichloropropane	50	0	56.8	ug/L	114		10		78	122	20
Bromodichloromethane	50	0	61.0	ug/L	122		8		79	125	20
4-Methyl-2-Pentanone	250	0	320	ug/L	128		10		67	130	20
Toluene	50	0	59.1	ug/L	118		14		80	121	20
t-1,3-Dichloropropene	50	0	64.7	ug/L	129	*	14		73	127	20
cis-1,3-Dichloropropene	50	0	63.1	ug/L	126	*	13		75	124	20
1,1,2-Trichloroethane	50	0	57.8	ug/L	116		8		80	119	20
2-Hexanone	250	0	330	ug/L	132		10		57	139	20
Dibromochloromethane	50	0	61.7	ug/L	123		9		74	126	20
1,2-Dibromoethane	50	0	59.5	ug/L	119		11		77	121	20
Tetrachloroethene	50	0	56.7	ug/L	113		14		74	129	20
Chlorobenzene	50	0	58.3	ug/L	117		12		82	118	20
Ethyl Benzene	50	0	59.8	ug/L	120		12		79	121	20
m/p-Xylenes	100	0	120	ug/L	120		18		80	121	20
o-Xylene	50	0	58.7	ug/L	117		13		78	122	20
Styrene	50	0	60.9	ug/L	122		12		78	123	20
Bromoform	50	0	60.5	ug/L	121		10		66	130	20
Isopropylbenzene	50	0	58.0	ug/L	116		11		72	131	20
1,1,2,2-Tetrachloroethane	50	0	55.2	ug/L	110		6		71	121	20

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Parameter	Spike	Sample	Result	Units	Rec			RPD	Limits		
		Result			Qual	RPD	Qual	Low	High	RPD	
1,3-Dichlorobenzene	50	0	56.0	ug/L	112		11		80	119	20
1,4-Dichlorobenzene	50	0	54.8	ug/L	110		9		79	118	20
1,2-Dichlorobenzene	50	0	56.8	ug/L	114		8		80	119	20
1,2-Dibromo-3-Chloropropane	50	0	64.5	ug/L	129	*	6		62	128	20
1,2,4-Trichlorobenzene	50	0	55.8	ug/L	112		7		69	130	20
1,2,3-Trichlorobenzene	50	0	54.6	ug/L	109		7		69	129	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045436.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0325WBS02	Dichlorodifluoromethane	20	19.5	ug/L	98			32	152	
	Chloromethane	20	16.0	ug/L	80			50	139	
	Vinyl chloride	20	16.0	ug/L	80			58	137	
	Bromomethane	20	19.6	ug/L	98			53	141	
	Chloroethane	20	19.7	ug/L	99			60	138	
	Trichlorofluoromethane	20	20.0	ug/L	100			65	141	
	1,1,2-Trichlorotrifluoroethane	20	22.6	ug/L	113			70	136	
	1,1-Dichloroethene	20	19.6	ug/L	98			71	131	
	Acetone	100	110	ug/L	110			39	160	
	Carbon disulfide	20	14.0	ug/L	70			64	133	
	Methyl tert-butyl Ether	20	21.5	ug/L	108			71	124	
	Methyl Acetate	20	29.4	ug/L	147		*	56	136	
	Methylene Chloride	20	20.2	ug/L	101			74	124	
	trans-1,2-Dichloroethene	20	20.1	ug/L	101			75	124	
	1,1-Dichloroethane	20	21.5	ug/L	108			77	125	
	Cyclohexane	20	19.7	ug/L	99			71	130	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	21.8	ug/L	109			72	136	
	cis-1,2-Dichloroethene	20	20.8	ug/L	104			78	123	
	Bromochloromethane	20	24.3	ug/L	121			78	123	
	Chloroform	20	22.3	ug/L	112			79	124	
	1,1,1-Trichloroethane	20	22.3	ug/L	112			74	131	
	Methylcyclohexane	20	20.5	ug/L	103			72	132	
	Benzene	20	20.5	ug/L	103			79	120	
	1,2-Dichloroethane	20	23.9	ug/L	119			73	128	
	Trichloroethene	20	20.5	ug/L	103			79	123	
	1,2-Dichloropropane	20	21.1	ug/L	106			78	122	
	Bromodichloromethane	20	22.7	ug/L	114			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	21.4	ug/L	107			80	121	
	t-1,3-Dichloropropene	20	22.1	ug/L	111			73	127	
	cis-1,3-Dichloropropene	20	22.4	ug/L	112			75	124	
	1,1,2-Trichloroethane	20	22.3	ug/L	112			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	22.0	ug/L	110			74	126	
	1,2-Dibromoethane	20	21.7	ug/L	109			77	121	
	Tetrachloroethene	20	21.6	ug/L	108			74	129	
	Chlorobenzene	20	21.4	ug/L	107			82	118	
	Ethyl Benzene	20	21.6	ug/L	108			79	121	
	m/p-Xylenes	40	43.1	ug/L	108			80	121	
	o-Xylene	20	21.4	ug/L	107			78	122	
	Styrene	20	21.7	ug/L	109			78	123	
	Bromoform	20	21.4	ug/L	107			66	130	
	Isopropylbenzene	20	22.5	ug/L	113			72	131	
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107			71	121	
	1,3-Dichlorobenzene	20	21.0	ug/L	105			80	119	
	1,4-Dichlorobenzene	20	20.5	ug/L	103			79	118	
	1,2-Dichlorobenzene	20	21.1	ug/L	106			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045436.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0325WBS02	1,2-Dibromo-3-Chloropropane	20	22.1	ug/L	111			62	128	
	1,2,4-Trichlorobenzene	20	20.1	ug/L	101			69	130	
	1,2,3-Trichlorobenzene	20	20.3	ug/L	102			69	129	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045460.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0326WBS01	Dichlorodifluoromethane	20	18.0	ug/L	90			32	152	
	Chloromethane	20	15.2	ug/L	76			50	139	
	Vinyl chloride	20	15.3	ug/L	77			58	137	
	Bromomethane	20	18.6	ug/L	93			53	141	
	Chloroethane	20	20.0	ug/L	100			60	138	
	Trichlorofluoromethane	20	18.3	ug/L	92			65	141	
	1,1,2-Trichlorotrifluoroethane	20	21.6	ug/L	108			70	136	
	1,1-Dichloroethene	20	17.8	ug/L	89			71	131	
	Acetone	100	100	ug/L	100			39	160	
	Carbon disulfide	20	12.8	ug/L	64			64	133	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			71	124	
	Methyl Acetate	20	27.9	ug/L	140		*	56	136	
	Methylene Chloride	20	19.0	ug/L	95			74	124	
	trans-1,2-Dichloroethene	20	18.7	ug/L	94			75	124	
	1,1-Dichloroethane	20	20.1	ug/L	101			77	125	
	Cyclohexane	20	18.2	ug/L	91			71	130	
	2-Butanone	100	100	ug/L	100			56	143	
	Carbon Tetrachloride	20	20.3	ug/L	102			72	136	
	cis-1,2-Dichloroethene	20	19.4	ug/L	97			78	123	
	Bromochloromethane	20	21.0	ug/L	105			78	123	
	Chloroform	20	21.1	ug/L	106			79	124	
	1,1,1-Trichloroethane	20	21.0	ug/L	105			74	131	
	Methylcyclohexane	20	18.5	ug/L	93			72	132	
	Benzene	20	19.1	ug/L	96			79	120	
	1,2-Dichloroethane	20	22.3	ug/L	112			73	128	
	Trichloroethene	20	18.9	ug/L	95			79	123	
	1,2-Dichloropropane	20	19.8	ug/L	99			78	122	
	Bromodichloromethane	20	21.5	ug/L	108			79	125	
	4-Methyl-2-Pentanone	100	110	ug/L	110			67	130	
	Toluene	20	19.7	ug/L	99			80	121	
	t-1,3-Dichloropropene	20	20.4	ug/L	102			73	127	
	cis-1,3-Dichloropropene	20	20.3	ug/L	102			75	124	
	1,1,2-Trichloroethane	20	20.7	ug/L	104			80	119	
	2-Hexanone	100	110	ug/L	110			57	139	
	Dibromochloromethane	20	21.1	ug/L	106			74	126	
	1,2-Dibromoethane	20	20.4	ug/L	102			77	121	
	Tetrachloroethene	20	20.4	ug/L	102			74	129	
	Chlorobenzene	20	19.8	ug/L	99			82	118	
	Ethyl Benzene	20	20.4	ug/L	102			79	121	
	m/p-Xylenes	40	40.8	ug/L	102			80	121	
	o-Xylene	20	19.8	ug/L	99			78	122	
	Styrene	20	20.5	ug/L	103			78	123	
	Bromoform	20	20.6	ug/L	103			66	130	
	Isopropylbenzene	20	20.9	ug/L	104			72	131	
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102			71	121	
	1,3-Dichlorobenzene	20	19.4	ug/L	97			80	119	
	1,4-Dichlorobenzene	20	19.2	ug/L	96			79	118	
	1,2-Dichlorobenzene	20	19.9	ug/L	100			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1629

Client: AECOM Technical Services, Inc.

Analytical Method: SW8260-Low

Datafile : VX045460.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0326WBS01	1,2-Dibromo-3-Chloropropane	20	21.2	ug/L	106			62	128	
	1,2,4-Trichlorobenzene	20	18.9	ug/L	95			69	130	
	1,2,3-Trichlorobenzene	20	18.7	ug/L	94			69	129	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0325WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1629

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: VX045433.D

Lab Sample ID: VX0325WBL01

Date Analyzed: 03/25/2025

Time Analyzed: 12:22

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0325WBS02	VX0325WBS02	VX045436.D	03/25/2025
TB	Q1629-12	VX045443.D	03/25/2025
FB03-20250321	Q1629-11	VX045444.D	03/25/2025
TT174I1-20250320	Q1629-01	VX045445.D	03/25/2025
TT162S1-20250320	Q1629-02	VX045446.D	03/25/2025
RW09-MW01D1-20250320	Q1629-07	VX045447.D	03/25/2025
RW09-MW01D2-20250320	Q1629-08	VX045448.D	03/25/2025
RW09-MW01D3-20250320	Q1629-09	VX045449.D	03/25/2025
TT150S1-20250320	Q1629-10	VX045450.D	03/25/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0326WBL01

Lab Name: CHEMTECH

Contract: AECO15

Lab Code: CHEM Case No.: Q1629

SAS No.: Q1629 SDG NO.: Q1629

Lab File ID: VX045459.D

Lab Sample ID: VX0326WBL01

Date Analyzed: 03/26/2025

Time Analyzed: 10:54

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0326WBS01	VX0326WBS01	VX045460.D	03/26/2025
RW09-MW01S-20250320	Q1629-14	VX045461.D	03/26/2025
RW09-MW01S-20250320MS	Q1629-15MS	VX045462.D	03/26/2025
RW09-MW01S-20250320MSD	Q1629-16MSD	VX045463.D	03/26/2025
TT188S1-20250320	Q1629-03	VX045466.D	03/26/2025
TT188S1-20250320MS	Q1629-04MS	VX045467.D	03/26/2025
TT188S1-20250320MSD	Q1629-05MSD	VX045468.D	03/26/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045067.D BFB Injection Date: 02/28/2025
Instrument ID: MSVOA_X BFB Injection Time: 01:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.7
75	30.0 - 60.0% of mass 95	53.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1) 1
174	50.0 - 100.0% of mass 95	73.8
175	5.0 - 9.0% of mass 174	5.8 (7.9) 1
176	95.0 - 101.0% of mass 174	70.6 (95.6) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX045068.D	02/28/2025	01:27
VSTDIC005	VSTDIC005	VX045069.D	02/28/2025	02:13
VSTDIC020	VSTDIC020	VX045070.D	02/28/2025	02:37
VSTDIC050	VSTDIC050	VX045071.D	02/28/2025	03:00
VSTDIC100	VSTDIC100	VX045072.D	02/28/2025	03:23
VSTDIC150	VSTDIC150	VX045073.D	02/28/2025	03:47

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045430.D BFB Injection Date: 03/25/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.2
75	30.0 - 60.0% of mass 95	55.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	72.4
175	5.0 - 9.0% of mass 174	5.4 (7.5) 1
176	95.0 - 101.0% of mass 174	69.5 (96.1) 1
177	5.0 - 9.0% of mass 176	4.3 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045431.D	03/25/2025	11:31
VX0325WBL01	VX0325WBL01	VX045433.D	03/25/2025	12:22
VX0325WBS02	VX0325WBS02	VX045436.D	03/25/2025	13:35
TB	Q1629-12	VX045443.D	03/25/2025	16:22
FB03-20250321	Q1629-11	VX045444.D	03/25/2025	16:46
TT174I1-20250320	Q1629-01	VX045445.D	03/25/2025	17:09
TT162S1-20250320	Q1629-02	VX045446.D	03/25/2025	17:32
RW09-MW01D1-20250320	Q1629-07	VX045447.D	03/25/2025	17:55
RW09-MW01D2-20250320	Q1629-08	VX045448.D	03/25/2025	18:18
RW09-MW01D3-20250320	Q1629-09	VX045449.D	03/25/2025	18:41
TT150S1-20250320	Q1629-10	VX045450.D	03/25/2025	19:05
VSTDCCC050EC	VSTDCCC050	VX045455.D	03/25/2025	21:00

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045456.D BFB Injection Date: 03/26/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:33
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	23.8
75	30.0 - 60.0% of mass 95	57.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.2
173	Less than 2.0% of mass 174	0.8 (1.1) 1
174	50.0 - 100.0% of mass 95	70.8
175	5.0 - 9.0% of mass 174	5.4 (7.6) 1
176	95.0 - 101.0% of mass 174	68.7 (97.1) 1
177	5.0 - 9.0% of mass 176	4.9 (7.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX045457.D	03/26/2025	10:03
VX0326WBL01	VX0326WBL01	VX045459.D	03/26/2025	10:54
VX0326WBS01	VX0326WBS01	VX045460.D	03/26/2025	11:18
RW09-MW01S-20250320	Q1629-14	VX045461.D	03/26/2025	11:46
RW09-MW01S-20250320MS	Q1629-15MS	VX045462.D	03/26/2025	12:09
RW09-MW01S-20250320MSD	Q1629-16MSD	VX045463.D	03/26/2025	12:33
TT188S1-20250320	Q1629-03	VX045466.D	03/26/2025	13:43
TT188S1-20250320MS	Q1629-04MS	VX045467.D	03/26/2025	14:06
TT188S1-20250320MSD	Q1629-05MSD	VX045468.D	03/26/2025	14:29
VSTDCCC050EC	VSTDCCC050	VX045470.D	03/26/2025	15:35

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045431.D Date Analyzed: 03/25/2025
Instrument ID: MSVOA_X Time Analyzed: 11:31
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	83260	5.54	144945	6.76	129576	10.05
UPPER LIMIT	166520	6.044	289890	7.257	259152	10.549
LOWER LIMIT	41630	5.044	72472.5	6.257	64788	9.549
EPA SAMPLE NO.						
TT174I1-20250320	70075	5.55	137552	6.76	122364	10.05
TT162S1-20250320	68003	5.54	133718	6.76	122480	10.05
RW09-MW01D1-20250320	74720	5.54	146822	6.76	135722	10.05
RW09-MW01D2-20250320	76218	5.55	148282	6.76	136799	10.05
RW09-MW01D3-20250320	71670	5.54	137481	6.76	129073	10.05
TT150S1-20250320	71974	5.55	142479	6.76	130670	10.05
FB03-20250321	71073	5.54	141090	6.76	129742	10.05
TB	70921	5.55	138131	6.76	128366	10.06
VX0325WBL01	71237	5.55	137150	6.76	123936	10.05
VX0325WBS02	88260	5.54	159282	6.76	137098	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045431.D Date Analyzed: 03/25/2025
Instrument ID: MSVOA_X Time Analyzed: 11:31
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63016	12.018				
UPPER LIMIT	126032	12.518				
LOWER LIMIT	31508	11.518				
EPA SAMPLE NO.						
TT174I1-20250320	47236	12.02				
TT162S1-20250320	50952	12.02				
RW09-MW01D1-20250320	52479	12.02				
RW09-MW01D2-20250320	55234	12.02				
RW09-MW01D3-20250320	54362	12.02				
TT150S1-20250320	55180	12.02				
FB03-20250321	49151	12.02				
TB	51682	12.02				
VX0325WBL01	50683	12.02				
VX0325WBS02	60910	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045457.D Date Analyzed: 03/26/2025
Instrument ID: MSVOA_X Time Analyzed: 10:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91964	5.54	157025	6.75	138010	10.05
UPPER LIMIT	183928	6.038	314050	7.251	276020	10.549
LOWER LIMIT	45982	5.038	78512.5	6.251	69005	9.549
EPA SAMPLE NO.						
TT188S1-20250320	68093	5.54	134396	6.76	120469	10.06
TT188S1-20250320MS	79755	5.54	146350	6.76	128351	10.06
TT188S1-20250320MSD	78228	5.55	141777	6.76	126420	10.05
RW09-MW01S-20250320	72703	5.54	139868	6.76	129727	10.05
RW09-MW01S-20250320MS	81846	5.54	149200	6.76	128995	10.05
RW09-MW01S-20250320MSD	81226	5.55	143287	6.76	127464	10.06
VX0326WBL01	65792	5.54	128574	6.76	116243	10.06
VX0326WBS01	86104	5.55	155857	6.76	136930	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
Lab File ID: VX045457.D Date Analyzed: 03/26/2025
Instrument ID: MSVOA_X Time Analyzed: 10:03
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63053	12.018				
UPPER LIMIT	126106	12.518				
LOWER LIMIT	31526.5	11.518				
EPA SAMPLE NO.						
TT188S1-20250320	47673	12.02				
TT188S1-20250320MS	58769	12.02				
TT188S1-20250320MSD	58644	12.02				
RW09-MW01S-20250320	52360	12.02				
RW09-MW01S-20250320MS	58223	12.02				
RW09-MW01S-20250320MSD	56894	12.02				
VX0326WBL01	46967	12.02				
VX0326WBS01	61262	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0325WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045433.D	1		03/25/25 12:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0325WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045433.D	1		03/25/25 12:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	58.3		81 - 118		117%	SPK: 50
1868-53-7	Dibromofluoromethane	54.2		80 - 119		108%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.8		85 - 114		110%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	71200	5.55				
540-36-3	1,4-Difluorobenzene	137000	6.757				
3114-55-4	Chlorobenzene-d5	124000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	50700	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0325WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045433.D	1		03/25/25 12:22	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045459.D	1		03/26/25 10:54	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.23	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.22	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.16	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.68	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045459.D	1		03/26/25 10:54	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.12	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.15	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.2		81 - 118		106%	SPK: 50
1868-53-7	Dibromofluoromethane	48.1		80 - 119		96%	SPK: 50
2037-26-5	Toluene-d8	46.4		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.2		85 - 114		94%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	65800	5.544				
540-36-3	1,4-Difluorobenzene	129000	6.757				
3114-55-4	Chlorobenzene-d5	116000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	47000	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBL01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045459.D	1		03/26/25 10:54	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBS02		SDG No.:	Q1629
Lab Sample ID:	VX0325WBS02		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045436.D	1		03/25/25 13:35	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	19.5		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	16.0		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.0		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	19.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	19.7		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	20.0		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	22.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	19.6		0.23	0.75	1.00	ug/L
67-64-1	Acetone	110		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	14.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	21.5		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	29.4		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	20.2		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	20.1		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	21.5		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	19.7		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	110		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	21.8		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	20.8		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	24.3		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	22.3		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	22.3		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	20.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	20.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	23.9		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	20.5		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	21.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	22.7		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	21.4		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBS02		SDG No.:	Q1629
Lab Sample ID:	VX0325WBS02		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045436.D	1		03/25/25 13:35	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	22.1		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	22.4		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	22.3		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	22.0		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	21.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	21.6		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	21.4		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	21.6		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	43.1		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	21.4		0.12	0.50	1.00	ug/L
100-42-5	Styrene	21.7		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	21.4		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	22.5		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	21.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	21.0		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	20.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	21.1		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	22.1		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	20.1		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	20.3		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.6		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	53.4		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	54.3		85 - 114		109%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	88300	5.544				
540-36-3	1,4-Difluorobenzene	159000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	60900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0325WBS02		SDG No.:	Q1629
Lab Sample ID:	VX0325WBS02		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045436.D	1		03/25/25 13:35	VX032525

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBS01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045460.D	1		03/26/25 11:18	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	18.0		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	15.2		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	15.3		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	18.6		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	20.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.3		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	21.6		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	100		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	12.8		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	20.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	27.9		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	19.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.7		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	20.1		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	18.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	100		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	20.3		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.4		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	21.0		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	21.1		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	21.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	19.1		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	22.3		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	18.9		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.8		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	21.5		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	110		0.68	2.50	5.00	ug/L
108-88-3	Toluene	19.7		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBS01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045460.D	1		03/26/25 11:18	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	20.4		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	20.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.7		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	21.1		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.4		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	20.4		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	19.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	20.4		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	40.8		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	19.8		0.12	0.50	1.00	ug/L
100-42-5	Styrene	20.5		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	20.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	20.9		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	20.4		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	19.4		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	19.2		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.9		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	21.2		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.9		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.7		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.8		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.0		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	48.3		89 - 112		97%	SPK: 50
460-00-4	4-Bromofluorobenzene	52.8		85 - 114		106%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	86100	5.55				
540-36-3	1,4-Difluorobenzene	156000	6.757				
3114-55-4	Chlorobenzene-d5	137000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	61300	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	
Client Sample ID:	VX0326WBS01		SDG No.:	Q1629
Lab Sample ID:	VX0326WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045460.D	1		03/26/25 11:18	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045467.D	1		03/26/25 14:06	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	46.4		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	40.7		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	40.6		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	46.0		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	43.2		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	46.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	54.8		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	47.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	280		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	34.3		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	54.1		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	71.5		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	49.1		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	47.9		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	52.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	46.2		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	290		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	53.1		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	50.5		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	53.5		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	55.2		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	55.3		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	48.9		0.16	0.50	1.00	ug/L
71-43-2	Benzene	49.7		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	58.4		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	53.5		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	51.4		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	56.3		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	290		0.68	2.50	5.00	ug/L
108-88-3	Toluene	51.6		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045467.D	1		03/26/25 14:06	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	56.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	55.2		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	53.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	300		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	56.2		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	53.5		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	49.3		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	51.8		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	53.2		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	100		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	51.7		0.12	0.50	1.00	ug/L
100-42-5	Styrene	54.0		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	54.9		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	51.8		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	52.0		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	50.4		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	50.0		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	52.3		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	60.7		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	51.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	51.1		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.7		81 - 118		107%	SPK: 50
1868-53-7	Dibromofluoromethane	49.3		80 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	46.5		89 - 112		93%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.1		85 - 114		102%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	79800	5.544				
540-36-3	1,4-Difluorobenzene	146000	6.757				
3114-55-4	Chlorobenzene-d5	128000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	58800	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-04MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045467.D	1		03/26/25 14:06	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-15MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045462.D	1		03/26/25 12:09	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	61.2		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	50.8		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	48.8		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	54.3		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	54.0		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	53.8		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	59.0		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	52.8		0.23	0.75	1.00	ug/L
67-64-1	Acetone	260		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	48.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	51.9		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	48.3		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	50.4		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	52.9		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	52.3		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	53.6		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	270		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	54.9		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	51.1		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	49.7		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	53.9		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	56.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	57.5		0.16	0.50	1.00	ug/L
71-43-2	Benzene	51.3		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	57.6		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	49.6		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	50.6		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	53.9		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	280		0.68	2.50	5.00	ug/L
108-88-3	Toluene	52.9		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-15MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045462.D	1		03/26/25 12:09	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	57.5		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	55.7		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	51.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	280		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	54.4		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	53.2		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	53.2		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	52.9		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	54.9		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	110		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	54.1		0.12	0.50	1.00	ug/L
100-42-5	Styrene	56.1		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	54.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	55.1		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	50.8		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	52.0		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	51.2		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	52.2		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	58.9		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	51.7		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	51.3		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.0		81 - 118		108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.2		80 - 119		102%	SPK: 50
2037-26-5	Toluene-d8	47.4		89 - 112		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.5		85 - 114		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	81800	5.544				
540-36-3	1,4-Difluorobenzene	149000	6.757				
3114-55-4	Chlorobenzene-d5	129000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	58200	12.024				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MS		SDG No.:	Q1629	
Lab Sample ID:	Q1629-15MS		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045462.D	1		03/26/25 12:09	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045468.D	1		03/26/25 14:29	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	65.0		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	55.3		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	52.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	59.9		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	54.7		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	57.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	63.4		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	57.1		0.23	0.75	1.00	ug/L
67-64-1	Acetone	300		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	52.6		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	58.3		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	51.2		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	55.0		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	57.5		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	57.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	58.0		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	310		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	61.2		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	56.9		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	49.6		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	59.6		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	62.0		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	61.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	56.5		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	64.8		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	60.4		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	56.8		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	61.0		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	320		0.68	2.50	5.00	ug/L
108-88-3	Toluene	59.1		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045468.D	1		03/26/25 14:29	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	64.7		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	63.1		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	57.8		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	330		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	61.7		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	59.5		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	56.7		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	58.3		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	59.8		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	120		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	58.7		0.12	0.50	1.00	ug/L
100-42-5	Styrene	60.9		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	60.5		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	58.0		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	55.2		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	56.0		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	54.8		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	56.8		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	64.5		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	55.8		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	54.6		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.3		81 - 118		113%	SPK: 50
1868-53-7	Dibromofluoromethane	53.0		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	50.2		89 - 112		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	57.5	*	85 - 114		115%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	78200	5.55				
540-36-3	1,4-Difluorobenzene	142000	6.757				
3114-55-4	Chlorobenzene-d5	126000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	58600	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	TT188S1-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-05MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045468.D	1		03/26/25 14:29	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-16MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045463.D	1		03/26/25 12:33	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	61.4		0.22	0.50	1.00	ug/L
74-87-3	Chloromethane	48.8		0.32	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	48.1		0.26	0.75	1.00	ug/L
74-83-9	Bromomethane	55.4		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	56.6		0.47	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	53.1		0.33	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	58.3		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	52.9		0.23	0.75	1.00	ug/L
67-64-1	Acetone	260		1.50	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	50.0		0.21	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	53.7		0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	49.5		0.27	0.75	1.00	ug/L
75-09-2	Methylene Chloride	50.6		0.28	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	53.1		0.23	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	52.7		0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	54.9		1.50	2.50	5.00	ug/L
78-93-3	2-Butanone	270		0.98	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	58.7		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	52.9		0.19	0.75	1.00	ug/L
74-97-5	Bromochloromethane	48.4		0.22	0.50	1.00	ug/L
67-66-3	Chloroform	55.1		0.25	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	57.6		0.20	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	60.0		0.16	0.50	1.00	ug/L
71-43-2	Benzene	53.4		0.15	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	60.7		0.22	0.50	1.00	ug/L
79-01-6	Trichloroethene	54.0		0.090	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	54.1		0.20	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	57.6		0.22	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	290		0.68	2.50	5.00	ug/L
108-88-3	Toluene	55.7		0.14	0.50	1.00	ug/L

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-16MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045463.D	1		03/26/25 12:33	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	61.9		0.17	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	59.3		0.16	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	54.5		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	300		0.89	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	58.5		0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	56.7		0.15	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	53.5		0.23	0.50	1.00	ug/L
108-90-7	Chlorobenzene	54.3		0.12	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	56.4		0.13	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	110		0.24	1.00	2.00	ug/L
95-47-6	o-Xylene	54.6		0.12	0.50	1.00	ug/L
100-42-5	Styrene	57.2		0.15	0.50	1.00	ug/L
75-25-2	Bromoform	56.6		0.19	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	56.7		0.12	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	53.3		0.26	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	53.6		0.16	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	53.5		0.19	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	53.2		0.16	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	60.7		0.53	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	53.9		0.20	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	52.6		0.20	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	41.9		81 - 118		84%	SPK: 50
1868-53-7	Dibromofluoromethane	40.1		80 - 119		80%	SPK: 50
2037-26-5	Toluene-d8	38.1	*	89 - 112		76%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.5		85 - 114		85%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	81200	5.55				
540-36-3	1,4-Difluorobenzene	143000	6.757				
3114-55-4	Chlorobenzene-d5	127000	10.055				
3855-82-1	1,4-Dichlorobenzene-d4	56900	12.018				

Report of Analysis

Client:	AECOM Technical Services, Inc.		Date Collected:	03/20/25	
Project:	NAVFAC NWIRP Bethpage, NY Site 1 OU-2 - 32258		Date Received:	03/21/25	
Client Sample ID:	RW09-MW01S-20250320MSD		SDG No.:	Q1629	
Lab Sample ID:	Q1629-16MSD		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX045463.D	1		03/26/25 12:33	VX032625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>AECO15</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1629</u>
Instrument ID:	<u>MSVOA_X</u>	SAS No.:	<u>Q1629</u>
Heated Purge: (Y/N)	<u>N</u>	SDG No.:	<u>Q1629</u>
GC Column:	<u>DB-624UI</u>	Calibration Date(s):	<u>02/28/2025</u>
ID:	<u>0.18 (mm)</u>	Calibration Time(s):	<u>01:27</u>
			<u>03:47</u>

LAB FILE ID:	RRF001 = VX045068.D			RRF005 = VX045069.D		RRF020 = VX045070.D		
	RRF050 = VX045071.D			RRF100 = VX045072.D		RRF150 = VX045073.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.624	0.646	0.646	0.627	0.626	0.607	0.629	2.4
Chloromethane	0.755	0.821	0.828	0.753	0.721	0.744	0.770	5.7
Vinyl Chloride	0.773	0.755	0.774	0.761	0.765	0.758	0.764	1
Bromomethane		0.337	0.298	0.292	0.284	0.291	0.300	7
Chloroethane	0.373	0.421	0.366	0.373	0.297	0.286	0.352	14.5
Trichlorofluoromethane	0.978	1.061	1.050	1.050	0.982	0.948	1.012	4.7
1,1,2-Trichlorotrifluoroethane	0.526	0.613	0.595	0.594	0.596	0.563	0.581	5.4
1,1-Dichloroethene	0.609	0.620	0.612	0.623	0.620	0.603	0.614	1.3
Acetone	0.414	0.363	0.384	0.351	0.345	0.356	0.369	7
Carbon Disulfide	1.584	1.582	1.587	1.660	1.708	1.698	1.636	3.6
Methyl tert-butyl Ether	1.955	1.913	2.127	2.083	2.132	2.158	2.061	5
Methyl Acetate	0.954	0.835	0.903	0.867	0.869	0.899	0.888	4.6
Methylene Chloride	0.806	0.730	0.752	0.698	0.694	0.706	0.731	5.8
trans-1,2-Dichloroethene	0.540	0.619	0.603	0.631	0.634	0.616	0.607	5.7
1,1-Dichloroethane	1.200	1.223	1.280	1.242	1.270	1.264	1.247	2.5
Cyclohexane		1.021	1.087	1.108	1.142	1.052	1.082	4.3
2-Butanone	0.476	0.545	0.610	0.579	0.553	0.570	0.555	8.1
Carbon Tetrachloride	0.463	0.463	0.447	0.468	0.489	0.463	0.465	3
cis-1,2-Dichloroethene	0.687	0.746	0.765	0.762	0.767	0.769	0.749	4.2
Bromochloromethane	0.557	0.633	0.613	0.603	0.563	0.596	0.594	5
Chloroform	1.206	1.247	1.278	1.246	1.230	1.225	1.239	2
1,1,1-Trichloroethane	0.908	0.992	1.009	1.024	1.044	1.025	1.000	4.8
Methylcyclohexane	0.464	0.530	0.560	0.585	0.607	0.550	0.549	9.1
Benzene	1.321	1.459	1.491	1.496	1.497	1.424	1.448	4.7
1,2-Dichloroethane	0.487	0.525	0.545	0.528	0.524	0.520	0.521	3.6
Trichloroethene	0.319	0.351	0.339	0.341	0.354	0.336	0.340	3.7
1,2-Dichloropropane	0.354	0.378	0.382	0.371	0.376	0.373	0.372	2.7
Bromodichloromethane	0.478	0.503	0.536	0.524	0.528	0.528	0.516	4.2
4-Methyl-2-Pentanone	0.535	0.570	0.647	0.610	0.579	0.579	0.587	6.5
Toluene	0.716	0.872	0.892	0.898	0.874	0.845	0.849	8

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Calibration Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX045068.D		RRF005 = VX045069.D		RRF020 = VX045070.D		
		RRF050 = VX045071.D		RRF100 = VX045072.D		RRF150 = VX045073.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
t-1,3-Dichloropropene	0.304	0.389	0.436	0.469	0.490	0.502	0.431	17.3
cis-1,3-Dichloropropene	0.404	0.463	0.509	0.535	0.555	0.553	0.503	11.8
1,1,2-Trichloroethane	0.346	0.348	0.371	0.356	0.341	0.336	0.350	3.6
2-Hexanone	0.349	0.412	0.476	0.448	0.431	0.436	0.425	10.1
Dibromochloromethane	0.305	0.349	0.390	0.384	0.385	0.380	0.366	9
1,2-Dibromoethane	0.311	0.352	0.371	0.356	0.355	0.350	0.349	5.7
Tetrachloroethene	0.315	0.326	0.319	0.324	0.329	0.309	0.320	2.3
Chlorobenzene	0.968	1.054	1.090	1.092	1.100	1.045	1.058	4.7
Ethyl Benzene	1.566	1.794	1.889	1.952	1.972	1.888	1.843	8.1
m/p-Xylenes	0.555	0.672	0.711	0.724	0.715	0.673	0.675	9.3
o-Xylene	0.609	0.689	0.702	0.706	0.707	0.670	0.681	5.5
Styrene	0.879	1.060	1.170	1.181	1.183	1.134	1.101	10.7
Bromoform	0.209	0.234	0.276	0.276	0.300	0.300	0.266	13.9
Isopropylbenzene	3.397	4.034	3.999	4.135	4.006	3.845	3.903	6.8
1,1,2,2-Tetrachloroethane	1.395	1.479	1.513	1.419	1.391	1.396	1.432	3.6
1,3-Dichlorobenzene	1.502	1.652	1.710	1.668	1.675	1.649	1.643	4.4
1,4-Dichlorobenzene	1.605	1.702	1.665	1.687	1.669	1.643	1.662	2.1
1,2-Dichlorobenzene	1.479	1.695	1.735	1.668	1.687	1.622	1.648	5.5
1,2-Dibromo-3-Chloropropane	0.237	0.243	0.285	0.289	0.300	0.320	0.279	11.6
1,2,4-Trichlorobenzene	0.668	0.821	0.978	1.009	1.074	1.080	0.938	17.3
1,2,3-Trichlorobenzene	0.747	0.897	1.032	1.059	1.107	1.096	0.990	14.2
1,2-Dichloroethane-d4		0.836	0.784	0.757	0.783	0.817	0.795	3.9
Dibromofluoromethane		0.329	0.335	0.329	0.340	0.338	0.334	1.5
Toluene-d8		1.237	1.191	1.210	1.219	1.203	1.212	1.4
4-Bromofluorobenzene		0.383	0.393	0.402	0.410	0.421	0.402	3.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 11:31

Lab File ID: VX045431.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.798		26.87	20
Chloromethane	0.770	0.804	0.1	4.42	20
Vinyl Chloride	0.764	0.720		-5.76	20
Bromomethane	0.300	0.346		15.33	20
Chloroethane	0.352	0.394		11.93	20
Trichlorofluoromethane	1.012	1.093		8	20
1,1,2-Trichlorotrifluoroethane	0.581	0.699		20.31	20
1,1-Dichloroethene	0.614	0.651		6.03	20
Acetone	0.369	0.394		6.78	20
Carbon Disulfide	1.636	1.643		0.43	20
Methyl tert-butyl Ether	2.061	2.180		5.77	20
Methyl Acetate	0.888	0.856		-3.6	20
Methylene Chloride	0.731	0.729		-0.27	20
trans-1,2-Dichloroethene	0.607	0.641		5.6	20
1,1-Dichloroethane	1.247	1.311	0.1	5.13	20
Cyclohexane	1.082	1.188		9.8	20
2-Butanone	0.555	0.596		7.39	20
Carbon Tetrachloride	0.465	0.558		20	20
cis-1,2-Dichloroethene	0.749	0.773		3.2	20
Bromochloromethane	0.594	0.584		-1.68	20
Chloroform	1.239	1.317		6.3	20
1,1,1-Trichloroethane	1.000	1.132		13.2	20
Methylcyclohexane	0.549	0.674		22.77	20
Benzene	1.448	1.547		6.84	20
1,2-Dichloroethane	0.521	0.624		19.77	20
Trichloroethene	0.340	0.365		7.35	20
1,2-Dichloropropane	0.372	0.390		4.84	20
Bromodichloromethane	0.516	0.591		14.53	20
4-Methyl-2-Pentanone	0.587	0.681		16.01	20
Toluene	0.849	0.944		11.19	20
t-1,3-Dichloropropene	0.431	0.547		26.91	20
cis-1,3-Dichloropropene	0.503	0.610		21.27	20
1,1,2-Trichloroethane	0.350	0.370		5.71	20
2-Hexanone	0.425	0.519		22.12	20
Dibromochloromethane	0.366	0.425		16.12	20
1,2-Dibromoethane	0.349	0.390		11.75	20
Tetrachloroethene	0.320	0.353		10.31	20
Chlorobenzene	1.058	1.130	0.3	6.8	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 11:31

Lab File ID: VX045431.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.068		12.21	20
m/p-Xylenes	0.675	0.760		12.59	20
o-Xylene	0.681	0.753		10.57	20
Styrene	1.101	1.269		15.26	20
Bromoform	0.266	0.306	0.1	15.04	20
Isopropylbenzene	3.903	4.076		4.43	20
1,1,2,2-Tetrachloroethane	1.432	1.422	0.3	-0.7	20
1,3-Dichlorobenzene	1.643	1.755		6.82	20
1,4-Dichlorobenzene	1.662	1.732		4.15	20
1,2-Dichlorobenzene	1.648	1.733		5.16	20
1,2-Dibromo-3-Chloropropane	0.279	0.323		15.77	20
1,2,4-Trichlorobenzene	0.938	1.047		11.62	20
1,2,3-Trichlorobenzene	0.990	1.048		5.86	20
1,2-Dichloroethane-d4	0.795	0.876		10.19	20
Dibromofluoromethane	0.334	0.368		10.18	20
Toluene-d8	1.212	1.260		3.96	20
4-Bromofluorobenzene	0.402	0.481		19.65	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 21:00

Lab File ID: VX045455.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.819		30.21	50
Chloromethane	0.770	0.884	0.1	14.81	50
Vinyl Chloride	0.764	0.774		1.31	50
Bromomethane	0.300	0.362		20.67	50
Chloroethane	0.352	0.433		23.01	50
Trichlorofluoromethane	1.012	1.154		14.03	50
1,1,2-Trichlorotrifluoroethane	0.581	0.697		19.97	50
1,1-Dichloroethene	0.614	0.684		11.4	50
Acetone	0.369	0.395		7.05	50
Carbon Disulfide	1.636	1.714		4.77	50
Methyl tert-butyl Ether	2.061	2.334		13.25	50
Methyl Acetate	0.888	0.929		4.62	50
Methylene Chloride	0.731	0.785		7.39	50
trans-1,2-Dichloroethene	0.607	0.696		14.66	50
1,1-Dichloroethane	1.247	1.404	0.1	12.59	50
Cyclohexane	1.082	1.254		15.9	50
2-Butanone	0.555	0.613		10.45	50
Carbon Tetrachloride	0.465	0.566		21.72	50
cis-1,2-Dichloroethene	0.749	0.825		10.15	50
Bromochloromethane	0.594	0.654		10.1	50
Chloroform	1.239	1.426		15.09	50
1,1,1-Trichloroethane	1.000	1.219		21.9	50
Methylcyclohexane	0.549	0.662		20.58	50
Benzene	1.448	1.606		10.91	50
1,2-Dichloroethane	0.521	0.655		25.72	50
Trichloroethene	0.340	0.376		10.59	50
1,2-Dichloropropane	0.372	0.410		10.22	50
Bromodichloromethane	0.516	0.612		18.6	50
4-Methyl-2-Pentanone	0.587	0.701		19.42	50
Toluene	0.849	0.978		15.19	50
t-1,3-Dichloropropene	0.431	0.530		22.97	50
cis-1,3-Dichloropropene	0.503	0.608		20.88	50
1,1,2-Trichloroethane	0.350	0.391		11.71	50
2-Hexanone	0.425	0.525		23.53	50
Dibromochloromethane	0.366	0.438		19.67	50
1,2-Dibromoethane	0.349	0.404		15.76	50
Tetrachloroethene	0.320	0.359		12.19	50
Chlorobenzene	1.058	1.173	0.3	10.87	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/25/2025 21:00

Lab File ID: VX045455.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.140		16.11	50
m/p-Xylenes	0.675	0.775		14.81	50
o-Xylene	0.681	0.774		13.66	50
Styrene	1.101	1.315		19.44	50
Bromoform	0.266	0.310	0.1	16.54	50
Isopropylbenzene	3.903	4.278		9.61	50
1,1,2,2-Tetrachloroethane	1.432	1.479	0.3	3.28	50
1,3-Dichlorobenzene	1.643	1.781		8.4	50
1,4-Dichlorobenzene	1.662	1.787		7.52	50
1,2-Dichlorobenzene	1.648	1.805		9.53	50
1,2-Dibromo-3-Chloropropane	0.279	0.328		17.56	50
1,2,4-Trichlorobenzene	0.938	1.039		10.77	50
1,2,3-Trichlorobenzene	0.990	1.055		6.57	50
1,2-Dichloroethane-d4	0.795	0.954		20	50
Dibromofluoromethane	0.334	0.375		12.27	50
Toluene-d8	1.212	1.279		5.53	50
4-Bromofluorobenzene	0.402	0.501		24.63	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 10:03

Lab File ID: VX045457.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.797		26.71	20
Chloromethane	0.770	0.767	0.1	-0.39	20
Vinyl Chloride	0.764	0.728		-4.71	20
Bromomethane	0.300	0.338		12.67	20
Chloroethane	0.352	0.394		11.93	20
Trichlorofluoromethane	1.012	1.079		6.62	20
1,1,2-Trichlorotrifluoroethane	0.581	0.681		17.21	20
1,1-Dichloroethene	0.614	0.637		3.75	20
Acetone	0.369	0.430		16.53	20
Carbon Disulfide	1.636	1.630		-0.37	20
Methyl tert-butyl Ether	2.061	2.084		1.12	20
Methyl Acetate	0.888	0.831		-6.42	20
Methylene Chloride	0.731	0.719		-1.64	20
trans-1,2-Dichloroethene	0.607	0.640		5.44	20
1,1-Dichloroethane	1.247	1.286	0.1	3.13	20
Cyclohexane	1.082	1.149		6.19	20
2-Butanone	0.555	0.589		6.13	20
Carbon Tetrachloride	0.465	0.551		18.5	20
cis-1,2-Dichloroethene	0.749	0.757		1.07	20
Bromochloromethane	0.594	0.619		4.21	20
Chloroform	1.239	1.300		4.92	20
1,1,1-Trichloroethane	1.000	1.097		9.7	20
Methylcyclohexane	0.549	0.665		21.13	20
Benzene	1.448	1.526		5.39	20
1,2-Dichloroethane	0.521	0.613		17.66	20
Trichloroethene	0.340	0.363		6.76	20
1,2-Dichloropropane	0.372	0.391		5.11	20
Bromodichloromethane	0.516	0.597		15.7	20
4-Methyl-2-Pentanone	0.587	0.648		10.39	20
Toluene	0.849	0.930		9.54	20
t-1,3-Dichloropropene	0.431	0.522		21.11	20
cis-1,3-Dichloropropene	0.503	0.600		19.28	20
1,1,2-Trichloroethane	0.350	0.371		6	20
2-Hexanone	0.425	0.477		12.23	20
Dibromochloromethane	0.366	0.419		14.48	20
1,2-Dibromoethane	0.349	0.386		10.6	20
Tetrachloroethene	0.320	0.355		10.94	20
Chlorobenzene	1.058	1.132	0.3	6.99	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 10:03

Lab File ID: VX045457.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.063		11.94	20
m/p-Xylenes	0.675	0.754		11.7	20
o-Xylene	0.681	0.732		7.49	20
Styrene	1.101	1.239		12.53	20
Bromoform	0.266	0.303	0.1	13.91	20
Isopropylbenzene	3.903	4.277		9.58	20
1,1,2,2-Tetrachloroethane	1.432	1.413	0.3	-1.33	20
1,3-Dichlorobenzene	1.643	1.715		4.38	20
1,4-Dichlorobenzene	1.662	1.707		2.71	20
1,2-Dichlorobenzene	1.648	1.684		2.18	20
1,2-Dibromo-3-Chloropropane	0.279	0.313		12.19	20
1,2,4-Trichlorobenzene	0.938	0.998		6.4	20
1,2,3-Trichlorobenzene	0.990	1.018		2.83	20
1,2-Dichloroethane-d4	0.795	0.847		6.54	20
Dibromofluoromethane	0.334	0.361		8.08	20
Toluene-d8	1.212	1.218		0.5	20
4-Bromofluorobenzene	0.402	0.444		10.45	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 15:35

Lab File ID: VX045470.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.629	0.846		34.5	50
Chloromethane	0.770	0.790	0.1	2.6	50
Vinyl Chloride	0.764	0.765		0.13	50
Bromomethane	0.300	0.359		19.67	50
Chloroethane	0.352	0.417		18.47	50
Trichlorofluoromethane	1.012	1.170		15.61	50
1,1,2-Trichlorotrifluoroethane	0.581	0.755		29.95	50
1,1-Dichloroethene	0.614	0.697		13.52	50
Acetone	0.369	0.438		18.7	50
Carbon Disulfide	1.636	1.741		6.42	50
Methyl tert-butyl Ether	2.061	2.308		11.98	50
Methyl Acetate	0.888	0.898		1.13	50
Methylene Chloride	0.731	0.780		6.7	50
trans-1,2-Dichloroethene	0.607	0.694		14.33	50
1,1-Dichloroethane	1.247	1.395	0.1	11.87	50
Cyclohexane	1.082	1.261		16.54	50
2-Butanone	0.555	0.621		11.89	50
Carbon Tetrachloride	0.465	0.601		29.25	50
cis-1,2-Dichloroethene	0.749	0.830		10.81	50
Bromochloromethane	0.594	0.601		1.18	50
Chloroform	1.239	1.418		14.45	50
1,1,1-Trichloroethane	1.000	1.213		21.3	50
Methylcyclohexane	0.549	0.736		34.06	50
Benzene	1.448	1.694		16.99	50
1,2-Dichloroethane	0.521	0.678		30.13	50
Trichloroethene	0.340	0.392		15.29	50
1,2-Dichloropropane	0.372	0.431		15.86	50
Bromodichloromethane	0.516	0.638		23.64	50
4-Methyl-2-Pentanone	0.587	0.694		18.23	50
Toluene	0.849	1.027		20.97	50
t-1,3-Dichloropropene	0.431	0.582		35.03	50
cis-1,3-Dichloropropene	0.503	0.665		32.21	50
1,1,2-Trichloroethane	0.350	0.399		14	50
2-Hexanone	0.425	0.510		20	50
Dibromochloromethane	0.366	0.453		23.77	50
1,2-Dibromoethane	0.349	0.413		18.34	50
Tetrachloroethene	0.320	0.397		24.06	50
Chlorobenzene	1.058	1.242	0.3	17.39	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: AECO15

Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG No.: Q1629

Instrument ID: MSVOA_X Calibration Date/Time: 03/26/2025 15:35

Lab File ID: VX045470.D Init. Calib. Date(s): 02/28/2025 02/28/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 01:27 03:47

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.843	2.270		23.17	50
m/p-Xylenes	0.675	0.834		23.56	50
o-Xylene	0.681	0.807		18.5	50
Styrene	1.101	1.377		25.07	50
Bromoform	0.266	0.322	0.1	21.05	50
Isopropylbenzene	3.903	4.772		22.26	50
1,1,2,2-Tetrachloroethane	1.432	1.521	0.3	6.22	50
1,3-Dichlorobenzene	1.643	1.873		14	50
1,4-Dichlorobenzene	1.662	1.881		13.18	50
1,2-Dichlorobenzene	1.648	1.816		10.19	50
1,2-Dibromo-3-Chloropropane	0.279	0.329		17.92	50
1,2,4-Trichlorobenzene	0.938	1.092		16.42	50
1,2,3-Trichlorobenzene	0.990	1.095		10.61	50
1,2-Dichloroethane-d4	0.795	0.782		-1.63	50
Dibromofluoromethane	0.334	0.335		0.3	50
Toluene-d8	1.212	1.120		-7.59	50
4-Bromofluorobenzene	0.402	0.415		3.23	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.