

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
 Lab File ID: VX045457.D Date Analyzed: 03/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:03
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	91964	5.54	157025	6.75	138010	10.05
UPPER LIMIT	183928	6.038	314050	7.251	276020	10.549
LOWER LIMIT	45982	5.038	78512.5	6.251	69005	9.549
EPA SAMPLE NO.						
TT188S1-20250320	68093	5.54	134396	6.76	120469	10.06
TT188S1-20250320MS	79755	5.54	146350	6.76	128351	10.06
TT188S1-20250320MSD	78228	5.55	141777	6.76	126420	10.05
RW09-MW01S-20250320	72703	5.54	139868	6.76	129727	10.05
RW09-MW01S-20250320MS	81846	5.54	149200	6.76	128995	10.05
RW09-MW01S-20250320MSD	81226	5.55	143287	6.76	127464	10.06
VX0326WBL01	65792	5.54	128574	6.76	116243	10.06
VX0326WBS01	86104	5.55	155857	6.76	136930	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.