

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
 Lab File ID: VX045431.D Date Analyzed: 03/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 11:31
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63016	12.018				
UPPER LIMIT	126032	12.518				
LOWER LIMIT	31508	11.518				
EPA SAMPLE NO.						
TT174I1-20250320	47236	12.02				
TT162S1-20250320	50952	12.02				
RW09-MW01D1-20250320	52479	12.02				
RW09-MW01D2-20250320	55234	12.02				
RW09-MW01D3-20250320	54362	12.02				
TT150S1-20250320	55180	12.02				
FB03-20250321	49151	12.02				
TB	51682	12.02				
VX0325WBL01	50683	12.02				
VX0325WBS02	60910	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.