

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO15
 Lab Code: CHEM Case No.: Q1629 SAS No.: Q1629 SDG NO.: Q1629
 Lab File ID: VX045457.D Date Analyzed: 03/26/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:03
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	63053	12.018				
UPPER LIMIT	126106	12.518				
LOWER LIMIT	31526.5	11.518				
EPA SAMPLE NO.						
TT188S1-20250320	47673	12.02				
TT188S1-20250320MS	58769	12.02				
TT188S1-20250320MSD	58644	12.02				
RW09-MW01S-20250320	52360	12.02				
RW09-MW01S-20250320MS	58223	12.02				
RW09-MW01S-20250320MSD	56894	12.02				
VX0326WBL01	46967	12.02				
VX0326WBS01	61262	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.