



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: BAPS North Bergen **SDG No.:** Q1640
Contract: BAPS03 **Lab Code:** CHEM **Case No.:** Q1640 **SAS No.:** Q1640

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Beryllium	6.00	+/-6.00	U	6.00	P	03/27/2025	10:02	LB135215
	Cadmium	6.00	+/-6.00	U	6.00	P	03/27/2025	10:02	LB135215
	Lead	12.0	+/-12.0	U	12.0	P	03/27/2025	10:02	LB135215

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>BAPS North Bergen</u>		SDG No.: <u>Q1640</u>							
Contract: <u>BAPS03</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1640</u>	SAS No.: <u>Q1640</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Beryllium	6.00	+/-6.00	U	6.00	P	03/27/2025	10:32	LB135215
	Cadmium	6.00	+/-6.00	U	6.00	P	03/27/2025	10:32	LB135215
	Lead	12.0	+/-12.0	U	12.0	P	03/27/2025	10:32	LB135215
CCB02	Beryllium	6.00	+/-6.00	U	6.00	P	03/27/2025	11:22	LB135215
	Cadmium	6.00	+/-6.00	U	6.00	P	03/27/2025	11:22	LB135215
	Lead	12.0	+/-12.0	U	12.0	P	03/27/2025	11:22	LB135215
CCB03	Beryllium	6.00	+/-6.00	U	6.00	P	03/27/2025	12:14	LB135215
	Cadmium	6.00	+/-6.00	U	6.00	P	03/27/2025	12:14	LB135215
	Lead	12.0	+/-12.0	U	12.0	P	03/27/2025	12:14	LB135215
CCB04	Beryllium	6.00	+/-6.00	U	6.00	P	03/27/2025	13:19	LB135215
	Cadmium	6.00	+/-6.00	U	6.00	P	03/27/2025	13:19	LB135215
	Lead	12.0	+/-12.0	U	12.0	P	03/27/2025	13:19	LB135215
CCB05	Beryllium	6.00	+/-6.00	U	6.00	P	03/27/2025	13:57	LB135215
	Cadmium	6.00	+/-6.00	U	6.00	P	03/27/2025	13:57	LB135215
	Lead	12.0	+/-12.0	U	12.0	P	03/27/2025	13:57	LB135215