

**Metals**

- 4 -

**INTERFERENCE CHECK SAMPLE**

**Client:** BAPS North Bergen **SDG No.:** Q1640  
**Contract:** BAPS03 **Lab Code:** CHEM **Case No.:** Q1640 **SAS No.:** Q1640  
**ICS Source:** EPA **Instrument ID:** P4

| Sample ID      | Analyte   | Result<br>ug/L | True Value<br>ug/L | %<br>Recovery | Low<br>Limit<br>(ug/L) | High<br>Limit<br>(ug/L) | Analysis<br>Date | Analysis<br>Time | Run<br>Number |
|----------------|-----------|----------------|--------------------|---------------|------------------------|-------------------------|------------------|------------------|---------------|
| <b>ICSA01</b>  | Beryllium | 1.34           |                    |               | -6                     | 6                       | 03/27/2025       | 10:11            | LB135215      |
|                | Cadmium   | -3.65          | 1.0                | 365           | -5                     | 7                       | 03/27/2025       | 10:11            | LB135215      |
|                | Lead      | 4.40           |                    |               | -12                    | 12                      | 03/27/2025       | 10:11            | LB135215      |
| <b>ICSAB01</b> | Beryllium | 483            | 495                | 98            | 420                    | 570                     | 03/27/2025       | 10:15            | LB135215      |
|                | Cadmium   | 1000           | 972                | 103           | 826                    | 1120                    | 03/27/2025       | 10:15            | LB135215      |
|                | Lead      | 53.4           | 49.0               | 109           | 37                     | 61                      | 03/27/2025       | 10:15            | LB135215      |