

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: BAPS North Bergen

SDG No.: Q1640

Contract: BAPS03

Lab Code: CHEM

Case No.: Q1640

SAS No.: Q1640

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167335BS							
Beryllium	ug/L	100	96.3		96	80 - 120	P
Cadmium	ug/L	100	91.3		91	80 - 120	P
Lead	ug/L	500	456		91	80 - 120	P