

LB135209


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

38946



A= 0.0000e+000

B= 2.5567e-004

C= 6.7630e-003

Rho= 0.9999523

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/0.9)
0.0	0.000	-0.004	-0.004	-44	0.000	-44	0				
0.2	0.200	0.186	-0.014	702	0.0 %	702					
2.5	2.500	2.530	0.030	9868	0.0 %	9868					
5.0	5.000	4.961	-0.039	19379	0.0 %	19379					
7.5	7.500	7.563	0.063	29553	0.0 %	29553					
10.0	10.000	9.964	-0.036	38946	0.0 %	38946					

6/0.9)
17
11
0

LB135209 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Mar 2025 11:17:55

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-44	-	0.0000	7470A	PPB	27 Mar 2025 11:31:11	S	Std
0.2 - 1	50.2	702	-	0.2000	7470A	PPB	27 Mar 2025 11:33:29	S	Std
2.5 - 1	52.5	9868	-	2.5000	7470A	PPB	27 Mar 2025 11:35:47	S	Std
5.0 - 1	55	19379	-	5.0000	7470A	PPB	27 Mar 2025 11:38:04	S	Std
7.5 - 1	55	29553	-	7.5000	7470A	PPB	27 Mar 2025 11:40:21	S	Std
10.0 - 1	57.5	36651	-	10.0000	7470A	PPB	27 Mar 2025 11:42:39	S	Std
10.0 - 1	510	38946	-	10.0000	7470A	PPB	27 Mar 2025 11:45:26	S	Std
ICV63 - 1	ICV63	15976	4.0914	-	7470A	PPB	27 Mar 2025 12:09:50	U	SMPL
ICB63 - 1	ICB63	-313	-0.0733	-	7470A	PPB	27 Mar 2025 12:12:05	U	SMPL
CCV88 - 1	CCV88	18051	4.6219	-	7470A	PPB	27 Mar 2025 12:16:55	U	SMPL
CCB88 - 1	CCB88	-296	-0.0689	-	7470A	PPB	27 Mar 2025 12:19:12	U	SMPL
CRA - 1	CRA	846	0.2231	-	7470A	PPB	27 Mar 2025 12:21:29	U	SMPL
HighStd - 1	HighStd	37208	9.5199	-	7470A	PPB	27 Mar 2025 12:26:23	U	SMPL
ChkStd - 1	ChkStd	28638	7.3288	-	7470A	PPB	27 Mar 2025 12:28:40	U	SMPL
PB167356BL - 1	PBW	-39	-0.0032	-	7470A	PPB	27 Mar 2025 12:33:39	U	SMPL
PB167356BS - 1	LCSW	16921	4.3330	-	7470A	PPB	27 Mar 2025 12:35:54	U	SMPL
Q1168-09 - 1	MDL-WATER-03-QT1-2025	180	0.0528	-	7470A	PPB	27 Mar 2025 12:43:22	U	SMPL
Q1625-01 - 1	FRAC-TANK-A1291	6369	1.6352	-	7470A	PPB	27 Mar 2025 12:45:39	U	SMPL
Q1641-02 - 1	NP-WS-003	-422	-0.1011	-	7470A	PPB	27 Mar 2025 12:47:56	U	SMPL
Q1643-06 - 1	A4212	-8	0.0047	-	7470A	PPB	27 Mar 2025 12:50:16	U	SMPL
Q1644-03 - 1	031925	5	0.0080	-	7470A	PPB	27 Mar 2025 12:52:31	U	SMPL
CCV89 - 1	CCV89	20296	5.1959	-	7470A	PPB	27 Mar 2025 12:54:46	U	SMPL
CCB89 - 1	CCB89	-211	-0.0472	-	7470A	PPB	27 Mar 2025 12:57:02	U	SMPL
Q1655-01 - 1	TANK-COMP	54	0.0206	-	7470A	PPB	27 Mar 2025 12:59:20	U	SMPL
Q1659-01 - 1	3808	154	0.0461	-	7470A	PPB	27 Mar 2025 13:01:35	U	SMPL
Q1659-02 - 1	3809	114	0.0359	-	7470A	PPB	27 Mar 2025 13:03:50	U	SMPL
Q1659-02DUP - 1	3809DUP	137	0.0418	-	7470A	PPB	27 Mar 2025 13:06:07	U	SMPL
Q1659-02MS - 1	3809MS	9736	2.4960	-	7470A	PPB	27 Mar 2025 13:08:22	U	SMPL
Q1659-02MSD - 1	3809MSD	8946	2.2940	-	7470A	PPB	27 Mar 2025 13:10:38	U	SMPL
Q1659-02LX5 - 1		-276	-0.0638	-	7470A	PPB	27 Mar 2025 13:12:57	U	SMPL
Q1659-02A - 1		11605	2.9739	-	7470A	PPB	27 Mar 2025 13:15:17	U	SMPL
Q1168-07 - 1		190	0.0553	-	7470A	PPB	27 Mar 2025 13:28:05	U	SMPL
Q1168-07RE - 1		532	0.1428	-	7470A	PPB	27 Mar 2025 13:30:20	U	SMPL
CCV90 - 1	CCV90	21226	5.4337	-	7470A	PPB	27 Mar 2025 13:32:35	U	SMPL
CCB90 - 1	CCB90	-263	-0.0605	-	7470A	PPB	27 Mar 2025 13:34:51	U	SMPL