



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1641
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1641 **SAS No.:** Q1641
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	112	40 - 160	CV	03/27/2025	12:21	LB135209
CRI01	Aluminum	96.9	100	97	40 - 160	P	03/28/2025	09:07	LB135228
	Arsenic	20.0	20.0	100	40 - 160	P	03/28/2025	09:07	LB135228
	Barium	88.4	100	88	40 - 160	P	03/28/2025	09:07	LB135228
	Cadmium	5.79	6.0	96	40 - 160	P	03/28/2025	09:07	LB135228
	Chromium	9.77	10.0	98	40 - 160	P	03/28/2025	09:07	LB135228
	Cobalt	28.9	30.0	96	40 - 160	P	03/28/2025	09:07	LB135228
	Copper	22.0	20.0	110	40 - 160	P	03/28/2025	09:07	LB135228
	Iron	89.2	100	89	40 - 160	P	03/28/2025	09:07	LB135228
	Lead	11.6	12.0	97	40 - 160	P	03/28/2025	09:07	LB135228
	Nickel	38.9	40.0	97	40 - 160	P	03/28/2025	09:07	LB135228
	Selenium	21.9	20.0	109	40 - 160	P	03/28/2025	09:07	LB135228
	Silver	10.6	10.0	106	40 - 160	P	03/28/2025	09:07	LB135228
	Tin	34.8	40.0	87	40 - 160	P	03/28/2025	09:07	LB135228
	Zinc	40.3	40.0	101	40 - 160	P	03/28/2025	09:07	LB135228